



**SUPPLEMENTAL / BID BULLETIN**  
**UNIVERSITY OF THE PHILIPPINES MANILA**  
*The Health Sciences Center*  
**Bids and Awards Committee 1**  
Taft Avenue, Manila  
Trunk Line No. 8554-8400 Local 3014/3015



**BAC ADDENDUM NO. 2020-069**

**for the Supply, Delivery, Installation, Testing and Commissioning of  
Ophthalmic Surgical Microscope with Stereo Coaxial Illumination  
16 December 2020**

Pursuant to Section 22.5.1 of the 2016 Revised Implementing Rules and Regulations of Republic Act No. 9184, the Bids and Awards Committee 1 is issuing this bid bulletin to modify or amend the following items in the Bid Documents in response to / to address the request / clarification of the prospective bidder/s who attended the pre-bid conference held on 11 December 2020:

**1. The Project Title should be modified as:**

From	To
Ophthalmic Surgical Microscope with Stereo Coaxial Illumination	<b><u>Supply, Delivery, Installation, Testing and Commissioning of</u></b> Ophthalmic Surgical Microscope with Stereo Coaxial Illumination

**2. The following should be modified in Section 1 as:**

No.	From	To
<b>1</b>	The University of the Philippines Manila – Philippine General Hospital (UPM-PGH), through the General Appropriations Act CY 2021 intends to apply the sum of Twelve Million Three Hundred Thousand Pesos & 00/100 (Php12,300,000.00), inclusive of all taxes, such as, but not limited to, value added tax (VAT), income tax, local taxes, and other fiscal levies, being the ABC to payments under the contract for Supply, Delivery, and Testing of Ophthalmic Surgical Microscope with Stereo Coaxial Illumination under Project Reference No.: PUR20-11-1007. Bids received in excess of the ABC shall be automatically rejected at bid opening.	The University of the Philippines Manila – Philippine General Hospital (UPM-PGH), through the General Appropriations Act CY <del>2021</del> <b>2020</b> intends to apply the sum of Twelve Million Three Hundred Thousand Pesos & 00/100 (Php12,300,000.00), inclusive of all taxes, such as, but not limited to, value added tax (VAT), income tax, local taxes, and other fiscal levies, being the ABC to payments under the contract for Supply, Delivery, and Testing of Ophthalmic Surgical Microscope with Stereo Coaxial Illumination under Project Reference No.: PUR20-11-1007. Bids received in excess of the ABC shall be automatically rejected at bid opening.

3. The following Technical Specifications should be modified as:

No.	From	To
A.1	LED light source illuminated by a fiber-optic system	LED light source illuminated by a fiber-optic system <u>or direct light source</u>
A.6	Built-in filters: blue-blocking filter halogen mode, retina protection device	Built-in filters: blue-blocking filter halogen mode, retina protection device <u>or equivalent retinal protection feature</u>
B.3	Magnification range of 4.1x-21x or greater with 10x eyepiece, or 5.1x to 30.7x with 12.5 eyepiece	Magnification range of 4.1x-21x or greater with 10x eyepiece, <del>or 5.1x to 30.7x with 12.5 eyepiece</del>
B.6	2 sets (1 pair for surgeon and 1 pair for assist) of 10x wide-field eyepieces with eyecups and diopter adjustment range between -5 diopters to +5 diopters or greater range	<u>At least two (2)</u> sets (1 pair for surgeon and 1 pair for assist) of <u>eye pieces with magnification of 10x wide field for first two (2) sets and 8.3x or 12.5x for optional additional sets</u> , eyepieces with eyecups and diopter adjustment range between -5 diopters to +5 diopters or greater range
B.7	2 sets (1pair for surgeon and 1pair for assist) of 8.3x or 12.5x wide-field eyepieces with eyecups and diopter adjustment range between -5 diopters to +5 diopters or greater range	<del>2 sets (1pair for surgeon and 1pair for assist) of 8.3x or 12.5x wide-field eyepieces with eyecups and diopter adjustment range between -5 diopters to +5 diopters or greater range</del>
B.11	X-Y coupling travel range of 61-62mm	X-Y coupling travel range of 61-62mm <u>/ 61-62mm</u>
C	Fundus viewing system	Fundus viewing system <u>(proprietary or third party)</u>
C.5	At least 1 unit metal sterilization tray	At least <u>one (1)</u> unit metal sterilization tray <u>compatible with the fundus viewing system</u>
D.1	Built-in display panel with touch screen capability for control of illumination settings and for live viewing of the surgical procedure	Built-in display panel with touch screen capability for control of illumination settings <del>and for live viewing of the surgical procedure</del>
D.2	At least 1 unit wired foot control panel with at least 3cm cable	At least 1 unit wired foot control panel with at least <del>3cm</del> <u>3m</u> cable
E	Camera, Display, and Recording	Camera, Display, and Recording <u>(third party)</u>
H.1	Sterilizable tray	Sterilizable tray <u>compatible with autoclavable parts and accessories of the surgical microscope</u>

4. The following Terms and Conditions should be modified as:

No.	From	To
B.1	Delivery Period: Within thirty (30) calendar days after receipt of Notice to Proceed (NTP).	Delivery Period: Within <del>thirty (30)</del> <u>sixty (60)</u> calendar days after receipt of Notice to Proceed (NTP).

No.	From	To
<b>B.3</b>	Warranty Period / coverage of Warranty: One (1) years on parts and. Free bin-annual preventive maintenance for three (3) years on service. Free quarterly preventive maintenance during the warranty. Warranty period shall commence from the date of acceptance by the end-user after installation, testing and commissioning.	Warranty Period / coverage of Warranty: One (1) years on parts and. <del>Free bin-annual preventive maintenance for three (3) years on service.</del> Free quarterly preventive maintenance during the warranty <u>period</u> . Warranty period shall commence from the date of acceptance by the end-user after installation, testing and commissioning.
<b>C.7</b>	Guarantee letter from the manufacturer and local distributor to ensure availability of supplies, parts and accessories for at least ten (10) years after expiration of the warranty period.	Guarantee letter from the manufacturer <del>and local distributor</del> to ensure availability of supplies, parts and accessories for at least ten (10) years after expiration of the warranty period.

**5. The following Technical Specifications should be added:**

➤ Fundus Viewing System

***Additional requirements if fundus viewing system will come from a third-party provider:***

- a. Autoclavable fundus viewing system***
- b. Co-axial alignment with the operating microscope***
- c. Automatic inversion of the image when the viewing system is swung in and out of the observation beam***
- d. Electronic foot pedal focusing***
- e. Lens material with amorphous carbon coating or quartz glass***
- f. Third-party provider must be in the business of manufacturing the vitreoretinal fundus viewing systems for at least 25 years***
- g. Third-party manufacturer must have presence in USA and Western Europe***
- h. The model of the fundus viewing system was released 2016 or later***

**6. Please be reminded of the following:**

- **The color of the folder to be used is RED:**
- **Dropping of bids is not later than 9:00AM, 8 January 2020; and**
- **Opening of bids is on 9:30 AM onwards, 8 January 2020.**

This shall form an integral part of the Bid Documents.

For the information and guidance of all concerned.

**Dean LEONARDO R. ESTACIO, Jr., PhD**  
Chairperson, Bids and Awards Committee 1

Received by the Bidder:

\_\_\_\_\_  
Signature over Printed Name

\_\_\_\_\_  
Name of Company

\_\_\_\_\_  
Date