

Proj. Ref. No.: **PUR23-06-0502**

Original Opening of Bids: **01September2023**

Re-Bid Opening: **29 September 2023**

Total ABC: **Php2,100,000.00**

End-User: **DEPARTMENT OF RADIOLOGY**

Project: **SUPPLY, DELIVERY AND INSTALLATION OF VALVE
REGULATED LEAD-ACID BATTERY COMPATIBLE
TO EXISTING FUJI DIGITAL MOBILE X-RAY MACHINE
(20 PCS./SET) - (REBID)**

Contract: **SINGLE BID**

Item No.	Qty.	UOM	Item Description	Approved Budget for the Contract per Unit (PHP)	Quotations (all taxes included)	
					in figures	in figures
1	7	Set	Valve Regulated Lead-Acid Battery Compatible to Existing FUJI Digital Mobile X-Ray Machine, 20 pcs./set	300,000.00		
	1		Specifications: Nominal Voltage: 12V Nominal Capacity (3C>7min): 9Ah Length: 151mm Width: 65mm Height: 99mm Weight: 2.5kg Internal Resistance: 23mΩ Operating Temperature Range: Charging: - 0 to 40 °C Discharging: -20 to 55°C Storage: - 15 to 50°C Float Charge Voltage: 13.7 to 13.8 V Cycle Use Charging Voltage: 14.6 to 14.8 V Float Charge Temperature Coefficient: -30mV/°C Maximum Charging Current: 40 A Month Self Discharge: 2% (20°C) Terminal: T2 (250) Container Material: ABS (UL94-HB) Non-spillable Flame Retardant			
TOTAL APPROVED BUDGET FOR THE CONTRACT:				Php2,100,000.00		

Approved by:

Dean CHARLOTTE M. CHIONG, MD., PhD.
Chairperson

(Signature over Printed Name of President / Gen. Manager)

(Name & Address of Company)

Proj. Ref. No.: **PUR23-06-0502**

End-User: **DEPARTMENT OF RADIOLOGY**
Project: **SUPPLY, DELIVERY AND INSTALLATION OF VALVE**
REGULATED LEAD-ACID BATTERY COMPATIBLE
TO EXISTING FUJI DIGITAL MOBILE X-RAY MACHINE
(20 PCS./SET) - (REBID)

Contract: **SINGLE BID**

Original Opening of Bids: **01September2023**

Re-Bid Opening: **29 September 2023**

Total ABC: **Php2,100,000.00**

TERMS AND CONDITIONS:

1. Delivery period is within fourteen (14) working days after the receipt of P.O. by the supplier
2. Brand new items
3. One (1) year warranty with certificate attached
4. The supplier shall make the deliveries in accordance with schedule, quality and specifications of the award and purchase order (PO). Failure to comply with the same shall be ground for cancellation of the award and P.O. issued.
5. The goods shall be delivered at the Property and Supply Division, Philippine General Hospital, during office hours.
6. Deliveries should be free from defects. Any defective item(s) shall be replaced within seven (7) working days.

Approved by:

Dean CHARLOTTE M. CHIONG, MD., PhD.
Chairperson

(Signature over Printed Name of President / Gen. Manager)

(Name & Address of Company)