

Project Ref. N : **PUR16-10-1483**
End-User : **DEPARTMENT OF OUT-PATIENT SERVICES (DOPS), PGH**
Project : **SUPPLY, DELIVERY AND INSTALLATION OF MECHANICAL OPERATING ROOM TABLE**

Opening of Bid: **10 November 2016**

ITEM NO.	QTY	UNIT	ITEM DESCRIPTION	APPROVED BUDGET FOR THE CONTRACT PER UNIT (Php)	QUOTATIONS (all taxes included)	
					(in figures)	(in words)
1	3	lot	<p align="center">SUPPLY, DELIVERY AND INSTALLATION OF MECHANICAL OPERATING ROOM TABLE Brand new, not reconditioned including its Parts & Accessories</p> <p>Specifications/Features:</p> <ol style="list-style-type: none"> 1. All functions are manually adjusted with Pneumatic spring support 2. Gas spring assisted on back section adjustment 3. Design: with ergonomic contours in all position and Y-shape design base guarantees maximum steadiness and freedom of movement of the surgical team 4. Table top length: 1800-2100mm 5. Table top sections: 6 sections 6. Table top width across side rails: 550 mm-600mm 7. Patient weight capacity: at least 130 kg. 8. Pads thickness : 50-70mm <p>Adjustment Ranges:</p> <ol style="list-style-type: none"> 1. Height <ol style="list-style-type: none"> a. Lowest: must not be higher than 740mm from floor to table top without foam b. Highest: must not be lower than 1040mm from floor to table top without foam 2. Lateral tilt <ol style="list-style-type: none"> a. Right tilt at least 20 degrees from zero position b. Left tilt at least 20 degrees from zero position 3. Trendelenburg: Must not be less than 25 degrees 4. Reverse Trendelenburg: Must not be less than 25 deg 5. Back section <ol style="list-style-type: none"> a. Tilt up: Must not be less than 70 degrees b. Tilt down: must not be less than 45 degrees 6. Leg section <ol style="list-style-type: none"> a. Tilt up: must not be less than 20 degrees b. Tilt down: must not be less than 80 degrees 	1,500,000.00		

Approved by:

Dean MONET M. LOQUIAS, RPh, PhD
Chairperson

(Signature over Printed Name of President/Gen. Manager)

(Name & Address of Company)

Project Ref. N : **PUR16-10-1483**
 End-User : **DEPARTMENT OF OUT-PATIENT SERVICES (DOPS), PGH**
 Project : **SUPPLY, DELIVERY AND INSTALLATION OF MECHANICAL OPERATING ROOM TABLE**

Opening of Bid: **10 November 2016**

ITEM NO.	QTY	UNIT	ITEM DESCRIPTION	APPROVED BUDGET FOR THE CONTRACT PER UNIT (Php)	QUOTATIONS (all taxes included)	
					(in figures)	(in words)
			<p>Features:</p> <ol style="list-style-type: none"> 1. The base, top frame column cladding and running gear must be constructed with stainless steel material. It must be sturdy and should not warp or bend when step upon by an operator weighing at least 150 lbs 2. Must have 4 double-jointed large casters and pedal brake for locking mechanism 3. The table must provide adequate foot room for unobstructed work especially on transferring patients and during operations. 4. The pads must be CFC-free, integrally foamed, electrically conductive, and non- Velcro fixation 5. The table must have full x-ray capability. 6. The table must be modular, with at least 5 table-top sections. (head, back, seat and leg sections) 7. Must have exchangeable table top system, allows a flexible, time-saving and efficient work flow. 8. Must have kidney elevator 9. Each unit must be supplied with the following accessories: <ol style="list-style-type: none"> a. Head rest b. Back section c. Back extension d. Two part leg section e. Standard arm support with clamp (2 pcs) f. Body strap g. Extendable Anesthesia screen with clamp h. Infusion pole with at least 4 hooks (1pc/unit) i. Clamp radial setting (2pcs) j. Stirrups (2pcs) 			
TOTAL APPROVED BUDGET FOR THE CONTRACT				4,500,000.00		

Approved by:

Dean MONET M. LOQUIAS, RPh, PhD
 Chairperson

 (Signature over Printed Name of President/Gen. Manager)

 (Name & Address of Company)

Project Ref. N : PUR16-10-1483
End-User : DEPARTMENT OF OUT-PATIENT SERVICES (DOPS), PGH
Project : SUPPLY, DELIVERY AND INSTALLATION OF MECHANICAL OPERATING ROOM TABLE

Opening of Bid: 10 November 2016

TERMS & CONDITIONS:

1. Indicate brand, model and country of origin.
2. Notarized certificate that the product is brand new, not reconditioned including parts and accessories.
3. Delivery should be done within one hundred twenty (120) calendar days commencing on the 3rd working day of notification through confirmed fax that the approved Purchase Order / Contract is already available for pick-up.
4. The system must be delivered ready to use. Cables, connectors, adapters and accessories not specified but are essential for the unit to function must be provided free by the winning bidder.
5. The offered Operating Room Table must be the latest model of the series.
6. Certification from at least five (5) Tertiary Hospitals that they have been supplied the same equipment of the same brand being offered by the supplier
7. Undertaking to provide product orientation and operations training for end user and troubleshooting training for biomedical engineers.
8. Warranty: At least two (2) years on parts and service warranty. Warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.
9. Provide free regular quarterly preventive maintenance and calibration during the warranty period.
10. Service engineers are certified factory trained on service and repair. One service engineer must be available locally to provide quick on-site support. Certification must be provided.
11. Provision of a loaner unit should the machine need repair during the warranty period.
12. Must provide a guarantee letter from the manufacturer and local distributor to ensure availability of supplies, parts and accessories for ten (10) years after end-users acceptance.
13. Operations training for end user and troubleshooting training for biomedical engineers.
14. The machine offered must have been demonstrated and/or tried and evaluated by the end user. Certification by end-user (to include brand and model) must be submitted.
- 14. Notarized certificate that the bidder must be in the business of distributing operating room beds equipment for the last for the last five (5) years.**
- 15. Notarized certificate that the bidder and its principal must have been in partnership for the last five (5) years.**
- 16. Notarized certificates of good track records within five to ten (5-10) years from five (5) government and/or private hospitals within Metro Manila / Philippines.**
17. Local service engineers: Must submit a list of factory trained local engineers with certificate of
18. The bidder must provide a quotation for the preventive maintenance cost after the warranty period expires.
19. The machine offered must have been demonstrated and evaluated by the end users. Certificate of demonstration including the brand and model being offered must be submitted.
20. Include technical specifications (brochure / catalogue).
21. All components must be of the same brand from the same manufacturer unless other wise specified as "Third Party"
22. The winning bidder must provide original and hard copy of operators and service manuals upon delivery.
23. The bidder guarantees that the bid offer is the lowest price in the market.
- 23. All certifications (#14, #15 and # 16) required in the Terms & Conditions shall be placed in the Envelope 1 (Technical Component).**
- 24. The bidder is required to submit single bid for the item. Two or more bids offer is automatically disqualified.**
- 25. Compliance with the provisions of Republic Act No. 9184 and other applicable laws.**

Approved by:

Dean MONET M. LOQUIAS, RPh, PhD
Chairperson

(Signature over Printed Name of President/Gen. Manager)

(Name & Address of Company)