

Proj. Ref. No.: **PUR22-05-0476**
End-User: **DEPARTMENT OF PEDIATRICS, DIVISION OF NEONATOLOGY**
Project: **SUPPLY, DELIVERY AND TESTING OF TEN (10) INFANT WARMERS FOR THE NEWLY RENOVATED NEONATAL ICU**
Contract: **Single Bid**

Opening of Bids: **01 JULY, 2021**
ABC: **Php10,000,000.00**

Item No.	Qty.	UOM	Item Description	Unit Price	Quotations (all taxes included)	
					in figures	in words
1	10	Unit	INFANT WARMERS FOR THE NEWLY RENOVATED NEONATAL ICU	1,000,000.00		
			Technical Specifications:			
			1. The unit must have microprocessor based control systems with self-test function performed at power up and continuously during operation.			
			2. Must have the following modes: a. Pre-warm b. Manual c. Servo Baby Mode			
			3. Heater output: from 0 to 100% in 5% to 10% increments			
			4. Heater must be able to swivel +/- 90 to 180 degrees to facilitate x-ray and other procedures.			
			5. Must be height adjustable			
			6. Patient bed must be able tilt and position patient in Trendelenburg and reverse Trendelenburg of at least 12 to 15 degrees			
			7. Patient bed must be able to rotate 360 degrees.			
			8. With heat conductive Gel mattress			
			9. Must have a built-in weighing scale.			
			10. Must have a built-in x-ray cassette.			
			11. Baby control (servo) mode: 34 to 38 degrees centigrade (or wider range) in increments of 0.1 degrees centigrade.			
			12. Must have reusable skin temperature probe with accuracy of +/- 0.1 °C			
			13. Skin Temperature Display: 18 to 42 °C or 25 to 40 °C or 26 to 45°C or 26 to 45°C			
			14. Temperature display resolution: +/- 0.1°C			

Approved by:

~Original Signed~

Dean BIENVENIDO S. BALOTRO, RPh, DBA, MS
Chairperson

(Signature over Printed Name of President / Gen. Manager)

(Name & Address of Company)

Proj. Ref. No.: **PUR22-05-0476**
End-User: **DEPARTMENT OF PEDIATRICS, DIVISION OF NEONATOLOGY**
Project: **SUPPLY, DELIVERY AND TESTING OF TEN (10) INFANT WARMERS FOR THE NEWLY RENOVATED NEONATAL ICU**
Contract: **Single Bid**

Opening of Bids: **01 JULY, 2021**
ABC: **Php10,000,000.00**

Item No.	Qty.	UOM	Item Description	Unit Price	Quotations (all taxes included)	
					in figures	in words
			15. Audio-visual alarm for overheating beyond set range			
			16. Must have an integrated APGAR score timer			
			17. Bright light of at least 500 lux for procedure and observation with an estimated lamp life of not less than 4,000 hours.			
			18. Power input: 220 VAC, 50/60 Hz, 600 to 800 watts max			
			19. Built-in battery or UPS to supply for at least one (1) hour in case of power failure			
			20. Must have an IV Pole.			
			21. Must be provided with an AVR with surge protector.			
Approved Budget for the Contract (ABC) :				PHP10,000,000.00		

TERMS & CONDITIONS:

A. Requirement/s if declared as Lowest/Single Calculated Bids:

1. Presentation of Technical data sheet and/or presentation of a prototype equipment within seven (7) calendar days after receipt of Notice of Lowest/ Single Calculated Bid.

B. Requirement/s if awarded the contract:

1. Delivery Period: Within Ninety (90) calendar days after receipt of Notice to Proceed (NTP)
2. Delivery Place: Equipment Section, Property & Supply Division, Philippine General Hospital, Taft Avenue, Manila
3. Warranty Period / Coverage of Warranty: One (1) year on parts and two (2) years on service. Free quarterly preventive maintenance during the warranty period. Warranty Period shall commence from the date of acceptance by the end user after installation, testing and commissioning.
4. Signed service level agreement with the Philippine General Hospital.
5. Original hard copy (not photocopy) or soft copy of operators and service manuals in English Language.
6. Training: product orientation for end users and troubleshooting training for at least two (2) biomedical engineers for one (1) day.
7. Quotation of the Annual Preventive Maintenance Cost after the warranty period expires
8. Acceptance Procedures and Parameters: Visual and functional testing.

Approved by:

~Original Signed~

Dean BIENVENIDO S. BALOTRO, RPh, DBA, MS
Chairperson

(Signature over Printed Name of President / Gen. Manager)

(Name & Address of Company)

Proj. Ref. No.: **PUR22-05-0476**
End-User: **DEPARTMENT OF PEDIATRICS, DIVISION OF NEONATOLOGY**
Project: **SUPPLY, DELIVERY AND TESTING OF TEN (10) INFANT
WARMERS FOR THE NEWLY RENOVATED NEONATAL ICU**
Contract: **Single Bid**

Opening of Bids: **01 JULY, 2021**
ABC: **Php10,000,000.00**

C. Documents Required of the Bidder to be submitted during Post-Qualification:

1. Brochures/ Technical data sheet
2. SEC registration to prove that the supplier is in the business of importing and supplying medical equipment
3. Certified true copy of the Certificate of Distributorship for the last three (3) years. The principal and the local distributor must have been in business partnership for at least three (3) years.
4. Certificate that the Brand must have been in the local market for at least five (5) years. Proof required: Invoices
5. The Brand must have been installed in at least five (5) government and/or private hospitals. A list of the hospital and contact no must be submitted.
6. Certification by the supplier that at least one service engineer is available locally to provide quick on-site support
7. Certificate of Performance Evaluation from the Single Largest Contract.
8. List of local Service Center/s
9. License to Operate (LTO) from the Philippine FDA

D. Documents Required of the Principal to be Submitted During Post-Qualification

1. Certification that the manufacturer has been in the business of manufacturing Hospital Equipment for at least Ten (10) years.
2. Guarantee Letter from the manufacturer to ensure availability of supplies, parts and accessories for at least five (5) years after expiration of the warranty period
3. Certification by the principal that service engineers are factory trained on service and repair.
4. ISO compliance Certificate of the manufacturer.
5. List of the manufacturer's office and contact details in the following territories: Western Europe, US/Canada and Japan.

Approved by:

~Original Signed~

Dean BIENVENIDO S. BALOTRO, RPh, DBA, MS
Chairperson

(Signature over Printed Name of President / Gen. Manager)

(Name & Address of Company)