

Proj. Ref. No.: **PUR22-06-0522**

Opening of Bids: **21 October 2022**

End-User: **DEPARTMENT OF SURGERY/
DIVISION OF UROLOGY**

ABC: PHP**771,200.00**

Project: **SUPPLY, DELIVERY AND TESTING OF ONE (1) LOT
TRANSURETHRAL RESECTOSCOPE SET COMPATIBLE
WITH THE EXISTING EQUIPMENT (REBID)**

Contract: **SINGLE BID**

Item No.	Qty.	UOM	Item Description	Unit Cost	Quotations (all taxes included)	
					in figures	in words
1	1	lot	TRANSURETHRAL RESECTOSCOPE SET COMPATIBLE WITH EXISTING EQUIPMENT	771,200.00		
			Components:			
			A. Outer sheath B. Inner sheath C. Outer sheath obturator D. Working element E. High tension cord F. Telescope lens G. Autoclavable container H. Consumables 1. Cutting loops 2. Pointed coagulating electrode 3. Ball-shaped coagulating electrode			
			Specifications			
			A. Outer sheath (1 @ Php63,300.00) Resectoscope outer sheath 1. Must be Fr 24 or Fr 26 in diameter, 2. Must have an oblique ceramic beak 3. Must have two working channels- one for inflow and one for outflow			
			B. Resectoscope Inner sheath (1 @ Php43,200.00) 1. Must be compatible with a Fr 24 or Fr 26 outer sheath 2. Must be fully rotating (360 degrees both directions) 3. Must have 2 working channels, one for inflow and one for outflow			

Approved by:

Dean CHARLOTTE M. CHIONG, MD, PhD
Chairperson

(Signature over Printed Name of President / Gen. Manager)

(Name & Address of Company)

Proj. Ref. No.: **PUR22-06-0522**

Opening of Bids: **21 October 2022**

End-User: **DEPARTMENT OF SURGERY/
DIVISION OF UROLOGY**

ABC: PHP**771,200.00**

Project: **SUPPLY, DELIVERY AND TESTING OF ONE (1) LOT
TRANSURETHRAL RESECTOSCOPE SET COMPATIBLE
WITH THE EXISTING EQUIPMENT (REBID)**

Contract: **SINGLE BID**

Item No.	Qty.	UOM	Item Description	Unit Cost	Quotations (all taxes included)	
					in figures	in words
			C. Outer sheath obturator (1 @ Php9,900.00) 1. Must have a blunt tip or tip may be fixed or flexible 2. Must compatible with Fr 24 or 26 outer sheath			
			D. Working element to provide monopolar cutting coagulating currents (1 @ Php143,000.00) 1. Must be compatible with Fr 24 to 26 outer and inner sheaths 2. Must be compatible with angled cutting loops, pointed cutting/coagulating electrodes and ball-shaped coagulating electrodes 3. Must have a movable thumb support with motion by means of a spring 4. At rest position, the electrode must be fully inside the sheath 5. Must be compatible with existing Covidien Valleylab Force FX cautery machines			
			E. Unipolar High frequency cord (1 @ Php4,400.00) Designed to transfer current from the cautery machine to the working element 1. Must have a 4mm plug. System with provision for adaptors for a 4mm plug are also acceptable 2. Must be at least length 300 cm in length 3. Must be compatible with existing Covidien Valleylab Force FX cautery machines			

Approved by:

Dean CHARLOTTE M. CHIONG, MD, PhD
Chairperson

(Signature over Printed Name of President / Gen. Manager)

(Name & Address of Company)

Proj. Ref. No.: **PUR22-06-0522**

Opening of Bids: **21 October 2022**

End-User: **DEPARTMENT OF SURGERY/
DIVISION OF UROLOGY**

ABC: PHP**771,200.00**

Project: **SUPPLY, DELIVERY AND TESTING OF ONE (1) LOT
TRANSURETHRAL RESECTOSCOPE SET COMPATIBLE
WITH THE EXISTING EQUIPMENT (REBID)**

Contract: **SINGLE BID**

Item No.	Qty.	UOM	Item Description	Unit Cost	Quotations (all taxes included)	
					in figures	in words
			F. Telescope lens (1 @ Php331,000.00) 1. 12 degrees lens with fiberoptic light transmission or equivalent technology 2. Max diameter: 4mm 3. Must be compatible with a Fr 24 or 26 working element 4. Must be autoclavable 5. Must be at most 30cm in length 6. Must be compatible with existing video equipment			
			G. Autoclavable container (1 @ Php18,000.00) 1. Must be able to contain all components of the set 2. Must be autotclavable			
			H. Consumables a. Angled cutting loops (24 @ Php4,400.00) For resection and coagulation of tissue i. Must be compatible with a Fr 24 or 26 working element b. Pointed coagulating electrode (6 @ Php4,400.00) Pointed electrode for cutting through tissue and for pinpoint coagulation i. Must be compatible with a Fr 24 or 26 working element c. Ball-shaped coagulating electrode (6 @ Php4,400.00) Ball shaped electrode for vaporizing of tissue and for coagulation of surfaces i. Must be compatible with a Fr 24 or 26 working element			
Total Approved Budget for the Contract:				Php771,200.00		

Approved by:

Dean CHARLOTTE M. CHIONG, MD, PhD
Chairperson

(Signature over Printed Name of President / Gen. Manager)

(Name & Address of Company)

Proj. Ref. No.: **PUR22-06-0522**

Opening of Bids: **21 October 2022**

End-User: **DEPARTMENT OF SURGERY/
DIVISION OF UROLOGY**

ABC: PHP**771,200.00**

Project: **SUPPLY, DELIVERY AND TESTING OF ONE (1) LOT
TRANSURETHRAL RESECTOSCOPE SET COMPATIBLE
WITH THE EXISTING EQUIPMENT (REBID)**

Contract: **SINGLE BID**

TERMS AND CONDITIONS:

A. Requirement/s if declared as Lowest/Single Calculated Bids:

Presentation of Technical data sheet and/or presentation of a prototype equipment within seven (7) calendar days after receipt of Notice of Lowest / Single Calculated Bid.

B. Requirement/s if awarded the contract:

1. Delivery Period: Within ninety (90) Calendar days after receipt of Notice to Proceed (NTP).
2. Delivery Place: Equipment Section, Property & Supply Division, Philippine General Hospital, Taft Avenue, Manila.
3. Warranty Period / Coverage of Warranty: One (1) year of parts service. Free quarterly preventive maintenance during the warranty period. Warranty Period shall commence from the date of acceptance by the end user after installation and testing
4. Signed service level agreement with the Philippine General Hospital.
5. Original Hard copy or (not photocopy) soft copy of operators and service manuals in English Language.
6. Product orientation for end users and troubleshooting training for at least two (2) biomedical engineers.
7. Quotation of the Annual Preventive Maintenance Cost after the warranty period expires
8. Acceptance Procedures and Parameters: Visual and functional testing.
9. Consumables must be readily replaceable and should be always available in the market once initial stocks are consumed

C. Documents required of the bidder to be submitted during Post-Qualification:

1. Brochures/Technical data Sheet.
2. SEC registration to prove that the supplier is in the business of importing and supplying medical equipment.
3. Certified true copy of the Certificate of Distributorship for the last five (5) years. The principal and the local distributor must have been in business partnership for at least five (5) years.
4. Certified true copy of the Certificate of Distributorship for the last ten (10) years. The principal and the local distributor must have been in business partnership for at least ten (10) years.
5. The model of the equipment must have been introduced in 2010 or later.

Approved by:

Dean CHARLOTTE M. CHIONG, MD, PhD
Chairperson

(Signature over Printed Name of President / Gen. Manager)

(Name & Address of Company)

Proj. Ref. No.: **PUR22-06-0522**

Opening of Bids: **21 October 2022**

End-User: **DEPARTMENT OF SURGERY/
DIVISION OF UROLOGY**

ABC: PHP**771,200.00**

Project: **SUPPLY, DELIVERY AND TESTING OF ONE (1) LOT
TRANSURETHRAL RESECTOSCOPE SET COMPATIBLE
WITH THE EXISTING EQUIPMENT (REBID)**

Contract: **SINGLE BID**

6. Certificate that the Brand has been in the local market for at least five (5) years. Proof Required: Invoices or Purchase Orders.
7. The Brand must have been installed in at least ten (10) government and/or private hospitals. A list of the hospital and contact number must be submitted.
8. Certification by the Supplier that at least one service engineer is available locally to provide quick on-support
9. Certification that components of the system are the same brand except when otherwise stated as "third party" in the specification.
10. List of local Service Center/s
11. Certificate of Performance Evaluation from the Single Largest Contract.
12. License to Operate (LTO) from the Philippine FDA.

D. Documents Required of the Principal to be submitted during Post-Qualification:

1. Certification that the manufacturer has been in the business of manufacturing hospital equipment for at least 10 years.
2. Guarantee letter from the manufacturer to ensure availability of supplies, parts and accessories for at least five (5) years after expiration of the warranty period.
3. Certification by the principal that service engineers are factory trained on service and repair.
4. ISO compliance certificate of the manufacturer.
5. List of the manufacturer's office and contact details in the following territories: Western Europe, US/Canada and Japan.

Approved by:

Dean CHARLOTTE M. CHIONG, MD, PhD
Chairperson

(Signature over Printed Name of President / Gen. Manager)

(Name & Address of Company)