



PHILIPPINE GENERAL HOSPITAL
The National University Hospital
University of the Philippines Manila
PURCHASING OFFICE
Taft Avenue, Manila

"PHIC-Accredited Health Care Provider"
ISO 9001 Certified

REQUEST FOR QUOTATION / PROPOSAL

Date: 14 December 2023

Purchase Request No.: **PUR23-11-1079**

Mode of Procurement: NEGOTIATED PROCUREMENT - SMALL VALUE

Please quote at your government price inclusive of VAT for the item/s listed below. It will be appreciated if you can submit your quotation duly signed by your representative at the **PURCHASING OFFICE 2nd flr. Righ Service WingNear Ward 8 PHILIPPINE GENERAL HOSPITAL, Taft Avenue, Manila** not later than 3:00pm on **21 December 2023**

General Conditions:

- All entries must be typewritten or in print.
- Delivery Period: **FIFTEEN (15) calendar days**
- Warranty shall be for a minimum period of six (6) months for the supplies and materials, one (1) year for equipment from the date of acceptance by the procuring entity.
- Bidders must indicate the BRAND and MODEL NUMBER offered when appropriate.
- Price validity shall be for a period of 90 calendar days from the date of opening of quotation/proposal.
- Documentary requirements should be attached upon submission of the quotation/proposal
 - ☒ Valid Business/Mayor's Permit
 - ☒ PhilGEPS Registration Number/Certificate
 - ☒ Latest Business/Income Tax Return
 - ☒ Notarized Omnibus Sworn Statement
- Supplier is required to submit single bid/offer only for each item. Two or more bids/offer is automatically disqualified.
- O T H E R S

For further information, you may call at (02) 8554-8400 local No. (3023) and look for **Mrs. THELMA C. BERNARDO.**

JULIET E. MADIZ
Acting Head

END-USER: Department of Medicine- Division of Cardiovascular Medicine, Central ECG Station

Item No.	Item Description	Qty	UOM	Unit Price (PHP)	Quoted Unit Price	Statement of Compliance		Remarks (Brand and Specifications)
						Yes	No	
1	SUPPLY, DELIVERY, AND TESTING AND COMMISSIONING OF SIX (6) UNITS ELECTROCARDIOGRAPH MACHINE FOR THE CENTRAL ECG STATION, DIVISION OF CARDIOVASCULAR MEDICINE OF THE DEPARTMENT OF MEDICINE , PHILIPPINE GENERAL HOSPITAL.	6	unit	150,000.00				
	Please see Annex A							
Approved Budget for the Contract: 900,000.00								

I/We, the undersigned Supplier, hereby OFFER to supply/deliver/perform the above described items.

Name of Company: _____
Name of Representative: _____
Position / Designation: _____
Signature _____

Item No.	Item Description	Qty	UOM	Unit Price (PHP)	Quoted Unit Price	Statement of Compliance		Remarks (Brand & Specifications)
						Yes	No	
	SUPPLY, DELIVERY, AND TESTING AND COMMISSIONING OF SIX (6) UNITS ELECTROCARDIOGRAPH MACHINE FOR THE CENTRAL ECG STATION, DIVISION OF CARDIOVASCULAR MEDICINE OF THE DEPARTMENT OF MEDICINE,							
1	PHILIPPINE GENERAL HOSPITAL.	6	unit	150,000.00				
	Technical Specifications:							
	A. General Features							
	1. Records and prints 12 lead resting ECG with at least 10-seconds duration as a standard feature							
	2. 10 lead wire, 12 lead simultaneous acquisition with programmable lead configuration with on-screen preview of acquired minimum range 10 second ECG waveform and optional measurement and interpretation.							
	3. With start-up time minimum range of less than 7 seconds							
	4. With 12-lead analysis computerized measurement for adults and pediatrics							
	5. With writer speed of 5, 12.5, 25, and 50 mm/s							
	6. With sensitivity/gain of 2.5, 5, 10, 20, 10/5 (split calibration) mm/mv							
	7. Provides visual indication of signal quality							
	8. With display type of atleast 4.3-inch (110mm) diagonal, TFT LCD with LED graphics backlit (color optional)							
	9. Display data on heart rate, patient ID, clock, battery power indicator, waveforms, lead labels, speed, gain and filter settings, warning messages, information messages, prompts, 12 leads standard display							
	10. With disconnected lead detection, excessive AC noise, baseline wander and muscle tremor messages							
	B. Power Supply							
	1. External AC/DC adaptor or battery operation							
	2. 100-240 VAC							
	3. Input frequency -50 to 60 Hz							
	4. With replaceable and rechargeable battery type and battery charge time of at least 3 hours from total discharge (with display off)							
	5. Battery capacity: with at least 360 minutes of continuous operation without recording or at least 100 minutes continuous rhythm print at 25 mm/s and 10mm/mv							
	C. ECG Storage Format							
	1. With external storage of at least 200 ECG's in external memory (SD Card) with ECG storage in PDF format and XML format							
	D. Accessories							
	1. IEC/AHA leadwire and electrode adaptor sets							
	2. 10 lead patient cable							
	3. Reusable electrodes, six (6) chest and four (4) limb electrodes							
	4. Country-specific power cords							
	E. Consumables							
	1. Roll paper at least 10 rolls							
	2. Electrode cream 250 ml/tube at least 10 tubes							
	3. SD cards compatible with unit, at least 12 pcs							
	Brand:							
	Country of Origin:							
	Delivery Date:							

Approved Budget for the Contract: 900,000.00

Total Amount of Quotation (in Words & in Figures)

I/We, the undersigned Supplier, hereby OFFER to supply/ deliver/ perform the above described items.

Name of Company: _____
Name of Representative: _____
Position / Designation: _____
Signature: _____