

University of the Philippines Manila

The Health Sciences Center

SPECIAL BIDS AND AWARDS COMMITTEE V (SBAC V)

BID NO. : SBAC V - PUR15-09-1496
 End -User : DEPARTMENT OF OBSTETRICS AND GYNECOLOGY (OB-GYN)
 Project : SUPPLY & DELIVERY OF COLOR DOPPLER ULTRASOUND SYSTEM
 ABC=P 2,200,000.00

Opening of Bids: 03 November 2015

Item No.	Qty.	Unit	ITEM/DESCRIPTION	APPROVED BUDGET FOR THE CONTRACT PER UNIT (PhP)	QUOTATIONS (in figures / in words) (all taxes included)
1	1	Unit	Color Doppler Ultrasound System * High Definition LED Monitor, 17-23inches with articulating arm * Probes: 2-8MHz convex probe (with Needle Biopsy Guide) 4-9MHz Endocautery Probe (with Needle Biopsy Guide) 4-8MHz 3D volume convex probe *Imaging Modes: 2D Mode M Mode Color Doppler Pulse Wave Spectral Doppler 3D Imaging Mode Dual Mode Quad modes Zoom Mode * 3 Probe Connectors detachable transducer holders, cable hook * Image Storage: Line Memory (min 180secs) Loop Memory (min 180secs) Fully Digital Real Time Recovery * Auto - Calculation, Quick Measurement *Applications: Obstetrics, Fetal Echo, Gynercology, Abdomen Package *Full patient Database & Complete D.Com System *Auxillary: DVD Multi Drive Digital B/W & Color Video Printer USB Printer DVD Recorder UPS Installation Application Training onsite 2 weeks for Faculty & Fellows Warranty Parts & Services for 2 years	2,200,000.00	
				2,200,000.00	

Approved by:



Dir. ALVIN B. MARCELO
 Chair

 Signature over Printed Name of President/Gen. Manager

 (Name & Address of Company)

University of the Philippines Manila
The Health Sciences Center
SPECIAL BIDS AND AWARDS COMMITTEE V (SBAC V)

BID NO. : SBAC V – PUR15-09-1496
End User : DEPARTMENT OF OBSTETRICS AND GYNECOLOGY (OB-GYN)
Project : SUPPLY & DELIVERY OF COLOR DOPPLER ULTRASOUND SYSTEM

TERMS AND CONDITIONS

1. Delivery Date: 30 calendar days from receipt of purchase order
2. Installation Period: within 3 days after delivery
3. The supplies must have been in the business of supplying medical equipments for the past 10 years
4. One service engineer must be available locally to provide quick on-site support. Must submit a list of factory trained local engineers with certificate of training from the manufacturer
5. Warranty of 2-3 years for parts and service
6. Must provide a guarantee letter from the manufacturer and local distributor to ensure availability of supplies, parts and accessories for at least ten (10) years after date of purchase order
7. The machine offered must have been demonstrated and tried by the end user (Department of OB-GYN) within 1 week after installation. Certification by end user must be presented
8. On- site training for faculty and fellows of the Section of Ob-Gyn Ultrasound for 2 weeks after installation
9. Quarterly preventive maintenance by the supplier's service engineers (free of charge) within the warranty period

OTHER REQUIREMENTS

Must be supplied with:

1. At least 15 bottles of ultrasonic gel
2. 10 rolls of type V thermal paper
3. Five canisters of prove cleanser/sanitizer
4. Protective cover for the probes

Approved by:



Dir. Alvin B. Marcelo
Chair

Signature over Printed Name of President/Gen. Manager

Name & Address of Company