

BIDFORM
(Technical Specification)

“FOR ENVELOPE 1”

Bid no. : BAC3-PNEI-PR8880
Date : November 04, 2019

Unit : Philippine National Ear Institute

REQUIRED SPECIFICATIONS	OFFERED SPECIFICATIONS	REFERENCE S (include Supporting documents)	COMMENTS/REMARK S (clarify inclusion/exclusion)	EVALUATION
<p>Specifications :</p> <p>Supply and Delivery of Cochlear Implant Device</p> <p>-Cochlear Implant Device composed of (1) implantable biocompatible internal receiver connected with (2) implantable biocompatible electrode array, and (3) behind-the-ear external processor with ear hook, microphone and battery pack. The implants must meet the following minimum technical specifications :</p> <p>-Implants : Titanium body English language labels/packaging Unique implant identification Forward compatibility Thickness: does not exceed 6mm once implanted Weight: does not exceed 10g Impact Resistance: minimum of 1.0 as long as the end user will be benefited fully of said implant</p>				

Note: Pls. fill-up all applicable columns.

End-User:

End-user / Representative

Signature of Representative _____

Name of Representative _____

Position _____

Company _____

Address _____

Telephone/Fax/E-mail _____

**BIDFORM
(Technical Specification)**

“FOR ENVELOPE 1”

Bid no. : BAC3-PNEI-PR8880
Date : November 04, 2019

Unit : Philippine National Ear Institute

REQUIRED SPECIFICATIONS	OFFERED SPECIFICATIONS	REFERENCE S (include Supporting documents)	COMMENTS/REMARK S (clarify inclusion/exclusion)	EVALUATION
<p>Virtual Channels: at least 120 virtual channels Stimulations Rates: A minimum of 31,000 pulses/sec is an acceptable stimulation rates, so long as it will fully benefit the user of the implant device and will not impair his hearing capabilities, nor damages his/her ear further Current range: 0-1200 µA/pulse phase Warranty: at least 10 years -Electrode: Lateral or peri-modular type Minimum of 16 stimulation contacts Designed to achieve atraumatic insertion Customizable electrode array with: - Option for : normal patients, malformed cochlea, cochlear ossification, residual hearing - Apical end diameter: should not exceed 0.5mm -Audio Processor (Behind the Ear) Behind the Ear processor with multiple programs English language labels/packaging Automatic sound management</p>				

Note: Pls. fill-up all applicable columns.

End-User:

_____ End-user / Representative

Signature of Representative _____
Name of Representative _____
Position _____
Company _____
Address _____
Telephone/Fax/E-mail _____

BIDFORM (Technical Specification)

“FOR ENVELOPE 1”

Bid no. : BAC3-PNEI-PR8880
Date : November 04, 2019

Unit : Philippine National Ear Institute

REQUIRED SPECIFICATIONS	OFFERED SPECIFICATIONS	REFERENCE S (include Supporting documents)	COMMENTS/REMARK S (clarify inclusion/exclusion)	EVALUATION
FM compatibility Wireless Connectivity for patient controls Easy to use remote control Input dynamic range: wide range, at least 70 dB Battery life: at least 35 hours Warranty: at least 2 years -Software: English language Updates: Free for lifetime Telemetry capabilities: allow for both intra-operative and post operative telemetry Works on Windows OS -Diagnostic Tool Kit to include: Impedance Telemetry Nerve Response Telemetry Electrically Evoked Stapedial Reflex Threshold Electrically Evoked Brainstem Response				

Note: Pls. fill-up all applicable columns.

End-User:

End-user / Representative

Signature of Representative _____

Name of Representative _____

Position _____

Company _____

Address _____

Telephone/Fax/E-mail _____

BIDFORM (Technical Specification)

“FOR ENVELOPE 1”

Bid no. : BAC3-PNEI-PR8880
Date : November 04, 2019

Unit : Philippine National Ear Institute

REQUIRED SPECIFICATIONS	OFFERED SPECIFICATIONS	REFERENCE S (include Supporting documents)	COMMENTS/REMARK S (clarify inclusion/exclusion)	EVALUATION
<p>-MRI conditions: safe for use with up to 1.5 Tesla Magnetic Resonance Imaging devices</p> <p>-Additional Bid Documents to be included in the Technical Components (Envelope 1)</p> <ol style="list-style-type: none"> 1. Brochures: English language brochure for cochlear implant devices 2. List of service centers 3. Certificate of Distributorship for cochlear implant device 4. FDA registration for cochlear implant device 5. ISO or equivalent document of the manufacturer 6. List of offices of the manufacturer: must have office in the USA or Western Europe AND Southeast Asia, preferably a local office in the Philippines <p>-Requirement/s if declared Lowest/Single Calculated Bid:</p> <ol style="list-style-type: none"> 1. Presentation of prototype sample of Cochlear Implant Device 2. List of specifications of options for customizable electrode array 				

Note: Pls. fill-up all applicable columns.

End-User:

End-user / Representative

Signature of Representative _____

Name of Representative _____

Position _____

Company _____

Address _____

Telephone/Fax/E-mail _____

BIDFORM (Technical Specification)

“FOR ENVELOPE 1”

Bid no. : BAC3-PNEI-PR8880
Date : November 04, 2019

Unit : Philippine National Ear Institute

	REQUIRED SPECIFICATIONS	OFFERED SPECIFICATIONS	REFERENCE S (include Supporting documents)	COMMENTS/REMARK S (clarify inclusion/exclusion)	EVALUATION
	<p>-Requirements if awarded contract:</p> <ol style="list-style-type: none"> 1. Storage and Maintenance: responsibility of storage and maintenance with winning bidder until beneficiary is identified 2. Delivery Period: within 2 weeks of identification of beneficiary patient 3. Delivery Place: Philippine General Hospital and/or National Cochlear Implant Centers in Visayas or Mindanao, list to follow 4. Warranty Period: 10 years for implantable device, 2 years for processor 5. Support Services: hospitalization and surgical care coverage, 2 years free programming and fitting of system, 1 year support for speech therapy, spare parts and repair support. 6. Arrangement for Payment and Delivery is Staggered <p>VAT INCLUSIVE</p>				

Note: Pls. fill-up all applicable columns.

End-User:

End-user / Representative

Signature of Representative _____

Name of Representative _____

Position _____

Company _____

Address _____

Telephone/Fax/E-mail _____