

SUPPLEMENTAL / BID BULLETIN



University of the Philippines Manila
The Health Sciences Center
SPECIAL BIDS & AWARDS COMMITTEE
8/F Central Block Building, PGH
Taft Avenue, Manila
Tel. Nos.: 554-8400 loc. 3014 to 3015



11 December 2015

ADDENDUM No. SBAC2015-19

This Addendum No. **SBAC2015-19** is issued to modify or amend the following in connection with the **SUPPLY & INSTALLATION OF HEMATOLOGY ANALYZER FOR PEDIATRICS HEMATOLOGY-ONCOLOGY PATIENTS – SINGLE BID** under PR/Bid No. **PUR15-09-1642** intended for the Department of Pediatrics which is scheduled on **18 December 2015**. This shall form an integral part of the Bid Documents.

1. *The project name should be corrected as:*
Supply & Installation of Hematology Analyzer for Pediatrics Hematology-Oncology Laboratory

2. *The following should be corrected in the Specifications as:*

From	To
1. Brand new Coulter Counter with the capability of identifying both normal and abnormal cells.	1. Brand new <u>Hematology Analyzer with five (5) parts differential</u> with the capability of identifying both normal and abnormal cells.
2. Designed for low volume of blood requirements about 100µl, 30µl in CBC mode and 50µl in CBC with differential mode which is deal for pediatric oncology patients whom large volume of blood is difficult to obtain.	2. <u>Must be designed for low volume of blood requirement of about 100µl</u> , which is deal for pediatric oncology patients whom large volume of blood is difficult to obtain.
4. With excellent precesion and accuracy for identifying both normal and abnormal cells	4. With excellent <u>precision</u> and accuracy: > <u>Must have an efficiency rate of at least 80% or better means it is capable of identifying blasts cells</u> > <u>Has the capability of correcting NRBC for every CBC analysis or its equivalent</u> > <u>The flagged morphologically abnormal samples were confirmed as appropriately by manual differentials</u> > <u>Ensure the repeatability and uniformity of analysis for all</u>

From	To
	<u>samples which ensure reliable results (triplicate reading or its equivalent). Follow the Levy-Jenning principle of standard deviation</u> <u>> Histogram</u>
8.c. Fluorescent platelet count with correlation to thrombopoetic markers and immature platelet fractions.	8.c. Fluorescent platelet count with correlation to thrombopoetic markers and immature platelet fractions <u>or its equivalent.</u>
B. Complete with the following: <ul style="list-style-type: none"> - Isolation Transformer - AVR, 3kVa - UPS, 3kVa - Epson LQ 3300 + 11 Printer - Host Interface Computer for Data Management 	<u>The following accessories should be included:</u> <ul style="list-style-type: none"> - Printer - <u>Monitor</u> - <u>UPS</u> - <u>Transformer</u> - <u>A cooling device to increase the life span of the machine (at least a 0.5HP split type air-conditioning unit)</u> - <u>Start-up</u> reagents

3. The following should be deleted in the Specifications as:
3. Convenient reagents which is deal for use in this environment
- a. Start-up reagents included
- b. Also included: Printer and monitor
- 6.c. Specimens like synovial or serous fluids of CSF
9. Must follow the Levy-Jenning principle of standard deviation
10. Must be designed to magnify the separation of population, data transformations mathematically enhance the analytical measurement to maximize population differences and expose cellular subpopulation
12. Quality assurance friendly system
- user defined standard operating procedure
 - must have multiple security level
 - software enable 21 CFR part 11 compliance
 - powerful and flexible software allows the processing and presentation of data to fit all needs including industrial application, research and quality control
 - automated calibration and calibration verification to ensure reliable results for both size and counts
 - V-check validation package offers comprehensive solution for today's quality assurance management
 - certification program to ensure instrument performance
4. The following should be added in the Specifications as:
- The machine and its accessories must be fit in a 3x1 meter space
 - Power supply: 220-240V
 - Reagents and Consumables:
 - a. One (1) year supply of start-up reagents based on the monthly census must be FREE which includes quality control reagents and calibrator

- b. Quality control reagents used during trouble shooting and calibration of the machine should be replaced by the supplier as well as consumed reagents during repair and preventive maintenance
 - c. Expiration of the reagents and quality control reagents should not be less than one (1) year from delivery.
 - d. No price increase for reagents for five (5) years.
- 5. Signatories for Envelope 1 and 2 should be the same person – President or General Manager.
- 6. Eligible bidders are required to submit the Bid Security either in the form of:
 - a. Cash/Cashier's/Manager's Check payable to the University of the Philippines Manila/Philippine General Hospital at two percent (2%) of the total Approved Budget of the Contract (ABC). Said amount should be paid at the UPM-PGH Cash Division and a photocopy of the Official Receipt should be attached to Envelope 1.
 - b. A combination of Cash/Cashier's/Manager's Check at a minimum of P50,000.00 and the balance at 2% Bank draft/guarantee or Irrevocable Letter of Credit issued by a Universal or Commercial Bank; Provided, however, that it shall be confirmed or authenticated by a Universal or Commercial Bank, if issued by a foreign bank. The Php50,000.00 should be paid at the UPM-PGH Cash Division and a photocopy of the Official Receipt should be attached to Envelope 1.
NOTE: Amount of Bid Security (Equal to percentage of the ABC) – First Php50,000.00 and the rest 2% bank draft/guarantee or irrevocable letter of credit
 - c. A combination of Cash/Cashier's/Manager's Check at a minimum of Php50,000.00 and the balance at 5% Surety Bond callable upon demand, it shall be accompanied by a certification by the Insurance Commission that the surety or insurance company is authorized to issue such instruments. The Php50,000.00 should be paid at the UPM-PGH Cash Division and a photocopy of the Official Receipt and original copy of the Surety Bond should be attached to Envelope 1.
NOTE: Amount of Bid Security (Equal to Percentage of the ABC) – First Php50,000.00 and the rest 5% Surety Bond.
 - d. Bid Securing Declaration
- 7. Bid Bulletin, Financial bid form – which includes bid prices, bill of quantities and applicable price schedules, and any other document required in the BIDS will form part of Envelope No. 2 and will be submitted together with Envelope No. 1, **on or before 9:00AM, 18 December 2015 (Friday)** at the Office of the Internal Audit, UP Manila, 8th Floor, Central Block Building, PGH Complex, Taft Avenue, Manila.

Please refer to the Instruction to Bidders for other information

For the information and guidance of all concerned.

Very truly yours,

ORIGINAL SIGNED

CHARLOTTE M. CHIONG, M.D.
 Chairperson 