



SUPPLEMENTAL / BID BULLETIN
UNIVERSITY OF THE PHILIPPINES MANILA
The Health Sciences Center
Special Bids and Awards Committee (DOH Funded)
Taft Avenue, Manila
Tel. #554-84-00 loc. 3014 / 3015



8 August 2016

ADDENDUM NO. SBAC(DOH)-VOM2016-12

This Addendum No. **SBAC(DOH)-VOM2016-12** is issued to modify or amend the following in connection with the **Supply, Delivery, Testing and Commissioning of Complete Stand Alone Modular Multi-Parameter Patient Monitor – Single Bid** with Project Reference No. **PUR16-07-0958** intended for TCVS, Department of Surgery, which is scheduled on **18 August 2016**. This shall form an integral part of the Bid Documents.

1. The following should be corrected in the Technical Specifications as:

From	To
2.a.6. Operation modes: diagnostic, monitor, filtered (electr-surgical)	2.a.6. Operation modes: diagnostic, monitor, filtered (<i>electro-surgical</i>)
2.b.1. SpO ₂ Cable with reusable finger probes- probe end should made of rubber and water proof – supports Traditional SpO ₂ with perfusion index, adult (pedia/infant)	2.b.1. SpO ₂ Cable with reusable finger probes- probe end should made of rubber and water proof – supports Traditional SpO ₂ with perfusion index, <i>infant/neonate (5), child (5) and adult (5)</i>
2.c.1 NIBP cable with cuffs for all patient types (adult large adult, small adult, child, infant, neonate) and adult thigh cuff	2.c.1 NIBP cable with cuffs for all patient types (adult large adult, small adult, child, infant, neonate) – <i>2 each patient type reusable</i> ; and adult thigh cuff – <i>5 each disposable</i>
2.c.3. Operation modes: manual, auto, STAT, venous puncture mode – turns the cuff into a tourniquet to assist venous puncture; NIBP customizable protoc	2.c.3. Operation modes: manual, auto, <i>STAT</i> ; <i>NIBP</i> customizable <i>protocol</i>
2.c.6. Measurement range selectable: adult, pedia, neon	2.c.6. Measurement range selectable: adult, pedia, <i>neonate</i>
2.d.1. multi-purpose (oral / nasopharyngeal / rectal temperature probe, autoclavable	2.d.1. multi-purpose (oral / nasopharyngeal / rectal temperature probe, <i>2 each</i>
2.e.1. Impedance method between RA AND LL with apnea alarm; has a cardiac overlay compensation by setting respiratory detection level to either manual or automatic	2.e.1. <i>Impedance</i> method between RA AND LL with apnea alarm; has a cardiac overlay compensation by setting respiratory detection level to either manual or automatic
3.a. Multi-Gas Analyzer Module (4 units)	3.a. Multi-Gas Analyzer Module (<i>2 units</i>)
3.b. ETCO ₂ Module with mainstream capnograph (4 units)	3.b. ETCO ₂ Module with mainstream <i>or sidestream</i> capnograph (4 units)
4. Configuration a. All units must have 1. Basic module (ECG, NIBP, Temperature, IBP) 2. Transport module 3. ETCO ₂ Module 4. Multi-gas analyzer	4. Configuration 1. Basic module (ECG, NIBP, Temperature, IBP), <i>4 units</i> 2. Transport module, <i>4 units</i> 3. ETCO ₂ Module, <i>4 units</i> 4. Multi-gas analyzer, <i>2 units</i> 5. NMT Module, <i>4 units</i>

From	To
5. NMT Module 6. PICCO Module 7. Consciousness monitor 8. Built-in 3 trace thermal recorder for printing real time and historical data	6. PICCO Module, <i>4 units</i> 7. Consciousness monitor, <i>4 units</i> 8. <i>Thermal recorder</i> for printing real time and historical data, <i>2 units</i>

2. The following should be corrected in the Required Software on all units as:

From	To
3.b. Ventilation values include minute volume, compliance, dead space, dead space/tidal volume ration and Alveolar Concentration	3.b. Ventilation values include minute volume, compliance, dead space, dead space/tidal volume ration and Alveolar <i>Concentration</i>
7.a. Support clinical decision tools that helps screen patient from sepsis using a checklist with information coming from automatic acquisition of vital signs and manual input of information such as white blood cell count	7.a. Support clinical decision tools that helps screen patient from sepsis using a checklist with information coming from automatic acquisition of vital signs and manual input of information <i>such</i> as white blood cell count

3. The following should be corrected in the Terms and Conditions as:

From	To
3. Delivery and installation should be done within ninety (90) calendar days commencing on the 3 rd working day of notification through confirmed fax that the approved Purchase Order / Contract is already available for pick-up.	3. Delivery and installation should be done within <i>one hundred twenty (120) calendar days</i> commencing on the 3 rd working day of notification through confirmed fax that the approved Purchase Order / Contract is already available for pick-up.
14. Notarized certification accompanied with proof that the bidder is in the business of supplying the unit/item for the last five (5) years.	14. Notarized certification accompanied with proof that the bidder is in the business of supplying the unit/item <i>(brand)</i> for the last five (5) years.
20. Must guaranteed that diagnostic technician to be able to assess the unit within 24 hours of notification.	20. Must <i>guarantee</i> that diagnostic technician <i>will be able</i> to assess the unit within 24 hours of notification.

4. The following should be deleted in the Terms & Conditions:

10. Provide certification issued by the end-user after the item/unit has been demonstrated and/or site visited/inspected and evaluated by specific end-user heads and to include brand and model. The site visit/demo should be exact machine model to be offered and the end-user's evaluation must be done prior to the opening of bids.

5. The following should be added in the Terms & Conditions:

➤ **Free software update for the life of the machine.**


6. Signatories for Envelope 1 and 2 should be the same person – President or General Manager.

7. Eligible bidders are required to submit the Bid Security either in the form of:
 - a. Cash/Cashier's/Manager's Check payable to the University of the Philippines Manila/Philippine General Hospital at two percent (2%) of the total Approved Budget of the Contract (ABC). Said amount should be paid at the UPM-PGH Cash Services Division and a photocopy of the Official Receipt should be attached to Envelope 1.
 - b. A combination of Cash/Cashier's/Manager's Check at a minimum of P50,000.00 and the balance at 2% Bank draft/guarantee or Irrevocable Letter of Credit issued by Universal or Commercial Bank; Provided, however, that it shall be confirmed or authenticated by a Universal or Commercial Bank, if issued by a foreign bank. The Php50,000.00 should be paid at the UPM-PGH Cash Services Division and a photocopy of the Official Receipt should be attached to Envelope 1.
NOTE: Amount of Bid Security (Equal to percentage of the ABC) – First Php50,000.00 and the rest 2% bank draft/guarantee or irrevocable letter of credit.
 - c. A combination of Cash/Cashier's/Manager's Check at a minimum of Php50,000.00 and the balance at 5% Surety Bond callable upon demand, it shall be accompanied by a certification by the Insurance Commission that the surety or insurance company is authorized to issue such instruments. The Php50,000.00 should be paid at the UPM-PGH Cash Services Division and a photocopy of the Official Receipt and original copy of the Surety Bond should be attached to Envelope 1.
NOTE: Amount of Bid Security (Equal to Percentage of the ABC) – First Php50,000.00 and the rest 5% Surety Bond.
 - d. Bid Securing Declaration

8. Bid Bulletin, Financial bid form – which includes bid prices, bill of quantities and applicable price schedules, and any other document required in the BIDS will form part of Envelope No. 2 and will be submitted together with Envelope No. 1, **on or before 9:00AM, 18 August 2016 (Thursday)** at the BAC Office (near at PGH Chapel), PGH, Taft Avenue, Manila.

9. Please refer to the Instruction to Bidders for other information.

For the information and guidance of all concerned.



Dean VICENTE O. MEDINA III, DDM, PhD
Chairperson, SBAC (DOH Funded)

Received by the Bidder:

 Signature over Printed Name

 Name of Company

 Date