



SUPPLEMENTAL / BID BULLETIN
UNIVERSITY OF THE PHILIPPINES MANILA
The Health Sciences Center
Special Bids and Awards Committee (DOH Funded)
Taft Avenue, Manila
Tel. #554-84-00 loc. 3014 / 3015



8 August 2016

ADDENDUM NO. SBAC(DOH)-VOM2016-14

This Addendum No. **SBAC(DOH)-VOM2016-14** is issued to modify or amend the following in connection with the **Supply, Delivery, Testing and Commissioning of Invasive Cardiac Monitor Complete with Neonate and Pedia Cuff – Single Bid** with Project Reference No. **PUR16-06-0823** intended for Department of Pediatrics, which is scheduled on **18 August 2016**. This shall form an integral part of the Bid Documents.

1. The following should be corrected in the Technical Specifications as:

From	To
2.a.vi. Operation modes: diagnostic, monitor, filtered (electr-surgical)	2.a.vi. Operation modes: diagnostic, monitor, filtered (<u>electro-surgical</u>)
2.b.i. SpO ₂ cable with reusable finger probed-probe end should made of rubber and water-proof supports Traditional SpO ₂ with perfusion index, adult and (pedia/infant)	2.b.i. SpO ₂ Cable with reusable <u>finger probe. End should</u> made of rubber and water <u>proof. Supports</u> traditional SpO ₂ with perfusion index. <u>With infant reusable probes – 5 pieces</u>
2.c.i. NIBP cable with cuffs for all patient types (adult, large adult, small adult, child, infant, neonate) and adult thigh cuff	2.c.i. NIBP cable with cuffs with the following sizes: <u>10 pieces – size 1,</u> <u>10 pieces – size 2,</u> <u>20 pieces – size 3,</u> <u>20 pieces – size 4,</u> <u>20 pieces – size 5</u>
2.c.iii. Operation modes: manual, auto, STAT, venous puncture mode – turns the cuff into a tourniquet to assist venous puncture; NIBP customizable protoc sequence – a sequence of up to four (4) automatic NIBP measurements, for example: Q5 (minutes) for the first hours, Q15 (minutes) for the second hours, Q30 (minutes) for the next hour and onwards	2.c.3. Operation modes: <u>manual, auto, STAT;</u> NIBP customizable protocol sequence – a sequence of up to four (4) automatic NIBP measurements, for example: Q5 (minutes) for the first hours, Q15 (minutes) for the second hours, Q30 (minutes) for the next hour and onwards
2.d. Temperature 1 Multipurpose (oral / nasopharyngeal / rectal temperature probe, autoclavable	2.d. Temperature 1 Multipurpose (oral / nasopharyngeal / rectal temperature probe (<u>2 pieces</u>)
Performance Specs: 9. Display: At least 15” Widescreen Display: color TFT LCD/LCD; resolution at least 1200 x 768	Performance Specs: 9. Display: <u>At least 12”</u> Widescreen Display: color TFT LCD/LCD; resolution at least <u>800 x 600</u>
Clinical decision tools. Supports clinical decision tools that help screen patient for sepsis using a checklist with information coming from automatic acquisition of vital signs and manual	<u>Preferably with</u> Clinical decision tools. Supports clinical decision tools that help screen patient for sepsis using a checklist with information coming from automatic acquisition of vital signs and

From	To
input of information such as white blood cell count.	manual input of information such as white blood cell count.

2. The following should be deleted in the Technical Specifications:
Ventilation calculator must be included, Built-in Ventilation calculator. Ventilation values include Minute Volume, Compliance, Dead Space, Dead Space/tidal Volume Ratio and Alveolar Concentration

3. The following should be deleted in the Terms & Conditions:
9. Must be compatible with existing modular cardiac monitors at the TCVS Operating Room.
16. Provide separate quotation for maintenance, repair and replacement of parts after the warranty period.


4. The following should be added in the Terms & Conditions:
 ➤ **Free software update for the life of the machine.**

5. Signatories for Envelope 1 and 2 should be the same person – President or General Manager.

6. Eligible bidders are required to submit the Bid Security either in the form of:
 - a. Cash/Cashier's/Manager's Check payable to the University of the Philippines Manila/Philippine General Hospital at two percent (2%) of the total Approved Budget of the Contract (ABC). Said amount should be paid at the UPM-PGH Cash Services Division and a photocopy of the Official Receipt should be attached to Envelope 1.
 - b. A combination of Cash/Cashier's/Manager's Check at a minimum of P50,000.00 and the balance at 2% Bank draft/guarantee or Irrevocable Letter of Credit issued by Universal or Commercial Bank; Provided, however, that it shall be confirmed or authenticated by a Universal or Commercial Bank, if issued by a foreign bank. The Php50,000.00 should be paid at the UPM-PGH Cash Services Division and a photocopy of the Official Receipt should be attached to Envelope 1.
NOTE: Amount of Bid Security (Equal to percentage of the ABC) – First Php50,000.00 and the rest 2% bank draft/guarantee or irrevocable letter of credit.
 - c. A combination of Cash/Cashier's/Manager's Check at a minimum of Php50,000.00 and the balance at 5% Surety Bond callable upon demand, it shall be accompanied by a certification by the Insurance Commission that the surety or insurance company is authorized to issue such instruments. The Php50,000.00 should be paid at the UPM-PGH Cash Services Division and a photocopy of the Official Receipt and original copy of the Surety Bond should be attached to Envelope 1.
NOTE: Amount of Bid Security (Equal to Percentage of the ABC) – First Php50,000.00 and the rest 5% Surety Bond.
 - d. Bid Securing Declaration

7. Bid Bulletin, Financial bid form – which includes bid prices, bill of quantities and applicable price schedules, and any other document required in the BIDS will form part of Envelope No. 2 and will be submitted together with Envelope No. 1, **on or before 9:00AM, 18 August 2016 (Thursday)** at the BAC Office (near at PGH Chapel), PGH, Taft Avenue, Manila.
8. Please refer to the Instruction to Bidders for other information.

For the information and guidance of all concerned.



Dean VICENTE O. MEDINA III, DDM, PhD
Chairperson, SBAC (DOH Funded)

Received by the Bidder:

Signature over Printed Name

Name of Company

Date