



Republic of the Philippines
University of the Philippines Manila
The Health Sciences Center
SPECIAL BIDS AND AWARDS COMMITTEE
(DOH FUND)
BAC Office, PGH Compound
Taft Avenue, Manila
Tel. No. 554-8400 loc. 3014/3015/3016



NOTICE OF AWARD

19 August 2016

MEDILINES DISTRIBUTORS, INC.
#7 Pioneer St. corner Sheridan St.
Mandaluyong City
Tel. No. 519-1373; 512-2012

Attention: **Mr. DANIEL C. ZULUETA**
General Manager

Dear Sir/Madame:

We are happy to notify you that your bid dated 14 July 2016 but awarded on 18 August 2016, for the supply of SUPPLY, DELIVERY, TESTING AND COMMISSIONING OF SYRINGE PUMP WITH TCI & PCA ANALGESIA/SEDATION CAPABILITIES for the Contract Price of equivalent to FOUR HUNDRED EIGHTY NINE THOUSAND NINE HUNDRED FIFTY SIX PESOS AND 00/100 ONLY (Php489,956.00), as corrected and modified in accordance with the Instructions to Bidders is hereby accepted.

You are hereby required to provide within **ten (10) days** the performance security in the form of a) Cash/Certified Check, Cashier's Check, Manager's Check, Bank Draft or Irrevocable Letter of Credit confirmed by a reputable local bank at 5% of the total amount of award which amounts to TWENTY FOUR THOUSAND FOUR HUNDRED NINETY SEVEN PESOS AND 80/100 ONLY (Php24,497.80) b) 30% of the total amount of award in the form of Surety Bond callable upon demand issued by any reputable surety or insurance company in the amount of ONE HUNDRED FORTY SIX THOUSAND NINE HUNDRED EIGHTY SIX PESOS AND 80/100 ONLY (Php146,986.80) to be submitted to the BAC Office, UPM PGH. Failure to provide the performance security shall constitute sufficient ground for the cancellation of the award and forfeiture of the bid security.

PGH16-06-0730 – DEPARTMENT OF ANESTHESIOLOGY

ITEM NO.	QTY.	UNIT	ARTICLES/DESCRIPTION	UNIT COST(Php)	TOTAL COST(Php)
1	4	unit	SYRINGE PUMP WITH TCI AND PCA ANALGESIA/ SEDATION CAPABILITIES Brand new, not reconditioned/refurbished its parts and accessories Brand/Model: B. BRAUN PERFUSOR <i>FOR SPACE</i> Country of Origin; Germany Please see attached supporting document : 1. Technical Specification TOTAL AWARD Php489,956.00	122,489.00	489,956.00

- More on Page 2 -

TERMS AND CONDITIONS:


1. Delivery and Installation within sixty (60) calendar days commencing on the 3rd working day of notification through confirmed fax that the approved purchase order/contract is already available for pick-up.

[Signature]
O. MACALVE 09/07/16


[Signature]
[Signature]

2. The system must be delivered ready to use. Cables, connectors, adapters and accessories not specified but are essential for the unit to function must be provided free by the winning bidder.
3. Five (5) years warranty on service and two (2) years on parts.
4. Warranty period shall be commence from the date of acceptance by the end-user after installation, testing and commissioning.
5. Automatic provision of a functioning service unit (same or higher capacity) within forty-eight (48) hours in case in case of malfunction/repair/pull-out of the purchased unit during the warranty period.
6. Provide free regular quarterly preventive maintenance and calibration during the service warranty.
7. The bidder must guarantee that a diagnostic technician will be able to evaluate the syringe pump within 48 hours of notification of malfunction.
8. Provide technical specifications (brochure/catalogue).
9. Provide original and hard copy of operators and service manuals upon delivery.
10. Free programming of drug library for the life of the machine.
11. The bidder guarantees that the bid offer is the lowest price in the market.
12. It is understood that all taxes and other impositions are included in the bid offer. However, the bidder concerned shall be liable for additional or other taxes and duties which may be imposed by the government during the effectivity of the contract,
13. Compliance with Republic Act No. 9184 and other applicable laws.

Very truly yours.


GERARDO D. LEGASPI, M.D.
 Director

CONFORME


JAY-R O. MACALUMB
 (Signature over Printed Name)
Sales Executive

(Designation)

09/07/16

(Date)

Technical Specifications

Copy
✓
PC


Tender's Specifications	Bidder's Specifications
<p>Syringe Pump with TCI and PCA Analgesia/Sedation Capabilities</p> <ol style="list-style-type: none"> 1. With manually set infusion rate plus Target Controlled Infusion (TCI) capability and upgradable to Patient Controlled Analgesia capability (with PCA kit) 2. Automatic rate calculation by calculating volume over time 3. Power supply: Built-in AC adapter; 11-16 V 4. Standard mains cord 5. Primary: 100-120 V AC – 50/60 Hz, 200-240V AC – 5-/60Hz, 110v 0.6 A/220 V 0.3 A 6. Lifetime: Minimum 10 years under continuous duty conditions 7. Has the following features: 220 VAC 50/60 Hz <ol style="list-style-type: none"> a. Dose rate mode b. Dose calculation c. Automatic calculation of delivery rate d. Drug library – up to 720 drug names can be stored e. Standby mode, adjustable from 1 minute to 24 hours 	<p>B.Braun Perfusor Space with TCI and PCA Analgesia/Sedation Capabilities</p> <ol style="list-style-type: none"> 1. With manually set infusion rate plus Target Controlled Infusion (TCI) capability and upgradable to Patient Controlled Analgesia capability (with PCA kit) 2. Automatic rate calculation by calculating volume over time 3. Power supply: Built-in AC adapter; External Low Voltage of 11-16 V 4. With external power supply cord 5. Primary : 100-240 V AC – 50/60 Hz 6. Lifetime: Minimum 10 years under continuous duty conditions 7. Has the following features: 100-240 VAC 50/60 Hz <ol style="list-style-type: none"> a. Dose rate mode b. Dose calculation c. Automatic calculation of delivery rate d. Drug library – up to 1500 drug names can be stored e. Standby mode, adjustable from 1 minute to 24 hours

MEDILINES DISTRIBUTORS, INC.
NO. 7 PICHELER COR. SHERIDAN
SHERIDAN, SAMA LUTUYONG CITY
TEL. (09) 999 634-91-32
FAX (09) 999 634-91-32

[Signature]
DA/HEL C. ZULUETA
GENERAL MANAGER

gpp / pec
 hr
 J

f. Help prompts, has data lock and alarm volume settings g. Has an infusion rate 0.01 – 999.9ml/h h. Has accuracy of set delivery rate $\pm 2\%$ i. Bonus delivery 1 – 1800 ml/h j. Occlusion alarm has 9 to 10 occlusion pressure levels k. Volume pre-selection 0.1 – 999ml l. With TIC capability m. Compatible with all syringe brands; can accommodate 12ml-, 20-, 35-, 50-60ml syringes	f. Help prompts, has data lock and alarm volume settings g. Has an infusion rate 0.01 – 999.9 ml/h h. Has accuracy of set delivery rate $\pm 2\%$ i. Bonus delivery 1 – 1800 ml/h j. Occlusion alarm has 9 occlusion pressure levels k. Volume pre-selection 0.1 – 9999ml l. With TCI capability m. Compatible with all syringe brands; can accommodate 2/3ml, 10ml, 20ml, 35ml, 50/60ml syringes
--	--

MEDLINE DISTRIBUTORS, INC.
 NO. 2 ROBERT COR. SIO FIDEL
 SUPERIOR, CALIFORNIA 94585
 (415) 334-1100 FAX (415) 334-1102
 TELEX 150000 MEDLINE

 MANUEL C. ZULUETA
 GENERAL MANAGER