

PR No. : PUR15-09-1395
End-User : DEPARTMENT OF SURGERY, PGH
Project : SUPPLY, DELIVERY, TESTING & COMMISSIONING OF A
DESK BASED BLOOD ANALYSIS SYSTEM

Opening of Bid: 7 December 2015

ITEM NO.	QTY	UNIT	ITEM DESCRIPTION	APPROVED PER UNIT (PhP)	QUOTATIONS (in figures / in words) (all taxes included)
1	1	unit	<p align="center">DESK BASED BLOOD ANALYSIS SYSTEM Brand new, not reconditioned Including its Parts & Accessories</p> <p><i>The standard system includes:</i></p> <ol style="list-style-type: none"> 1. Point of care, desk based 2. Single-use cartridge 3. Must have an interactive touch screen user interface and built-in thermal printer 4. Room temperature cartridge storage, no need for refrigeration is preferred. Self life of cartridges between three (3) to four (4) months at room temperature. 5. Self calibrates before running test 6. On- board automatic electronic quality control 7. Prepackaged onboard calibrant. No external reagents or gases required. 8. Closed sample introduction system 9. Requires only 0.2ml of blood sample from a syringe. 10. Results available in 2 minutes or less. 11. Can analyse the sodium, potassium, Calcium, Chloride, pH, pCO₂, pO₂, Urea, Glucose, Hemoglobin, Hematocrit, TCO₂, HCO₃, and SO₂. 	750,000.00	
TOTAL APPROVED BUDGET FOR THE CONTRACT				Php750,000.00	

TERMS & CONDITIONS:

1. Indicate brand, model and country of origin.
2. Notarized certificate that the product is brand new, not reconditioned including parts and accessories.
3. Delivery should be done within one hundred twenty (120) calendar days commencing on the 3rd working day of notification through confirmed fax that the approved Purchase Order / Contract is already available for pick-up.

Approved by:

Dean VICENTE O. MEDINA III, DDM, PhD
Chairperson

(Signature over Printed Name of President/Gen. Manager)

(Name & Address of Company)
\\cmp11.11.2015

PR No. : PUR15-09-1395
End-User : DEPARTMENT OF SURGERY, PGH
Project : SUPPLY, DELIVERY, TESTING & COMMISSIONING OF A
DESK BASED BLOOD ANALYSIS SYSTEM

Opening of Bid: 7 December 2015

cont... TERMS & CONDITIONS:

4. The system must be delivered ready to use. Cables, connectors, adapters, and accessories not specified in the specifications but are essential for the unit to function must be provided free by the supplier.
5. The machine offered must be the latest model.
6. Certification from at least five (5) tertiary hospitals that they have been supplied a desk based blood analyzer of the same brand being offered.
7. Undertaking to provide product orientation and operations training for end-user and troubleshooting training for biomedical engineers.
8. Two (2) years warranty on service and on parts.
9. Warranty period shall commence from the date of acceptance by the end-user after installation, testing and commissioning.
10. Free quarterly preventive maintenance during the two (2) year warranty period.
11. Service engineers are certified factory trained on service and repair. One (1) service engineers must be available locally to provide quick on-site support. Certification must be provided.
12. Provision of a loaner unit should the machine need repair during the warranty period.
13. Must provide a guarantee letter from the manufacturer and local distributor to ensure availability of supplies, parts and accessories for at least ten (10) years after date of purchase order.
14. The bidder must provide a quotation of the preventive maintenance cost after the warranty period expires.
15. Operations training for end user and trouble shooting training for biomedical engineers.
16. The machine offered must have been demonstrated and/or tried by the end-user. Certification by end user (to include brand and model) must be presented.
17. Software updates will be installed free of charge.
- 18. Notarized certificate that the bidder must be in the business of distributing operating room beds equipment for the last for the last five (5) years.**
- 19. Notarized certificate that the bidder and its principal must have been in partnership for the last five (5) years.**
- 20. Notarized certificates of good track records within five to ten (5-10) years from five (5) government and/or private hospitals within Metro Manila / Philippines.**
21. Local service engineers: Must submit a list of factory trained local engineers with certificate of training from the manufacturer.
22. The bidder must provide a quotation for the preventive maintenance cost after the warranty period expires.
23. The machine offered must have been demonstrated and evaluated by the end users. Certificate of demonstration including the brand and model being offered must be submitted.
24. Include technical specifications (brochure / catalogue).
25. All components must be of the same brand from the same manufacturer unless other wise specified as "Third Party"
26. Must provide original and hard copy of operators and service manuals upon delivery.
27. The bidder guarantees that the bid offer is the lowest price in the market.
- 28. All certifications (#15, #16 and # 17) required in the Terms & Conditions shall be placed in the Envelope 1 (Technical Component).**
- 29. The bidder is required to submit single bid for the item. Two or more bids offer is automatically disqualified.**
- 30. Compliance with the provisions of Republic Act No. 9184 and other applicable laws.**

Approved by:

Dean VICENTE O. MEDINA III, DDM, PhD
Chairperson

(Signature over Printed Name of President/Gen. Manager)

(Name & Address of Company)