

The Health Science Center
BIDS AND AWARDS COMMITTEE III (BAC 3)
 Joaquin Gonzales Hall, UP Manila
 Padre Faura St., corner Ma. Orosa St., Ermita, Manila
 Telefax #525-4322, 526-2275


NOTICE OF AWARD


March 25, 2014

GOLDEN BAT (FAR EAST) INC.
 52 Sct. Alcaraz St. cor. D. Tuazon
 Quezon City
 Tel. # 731-9745 to 46/742-2368, 731-5217

ITEM NO.	QTY.	ARTICLES/DESCRIPTION	UNIT PRICE (Php)	TOTAL PRICE (Php)
5.	4 units	<p>Biobase Model FH1200 Fumehood</p> <p>With:</p> <ul style="list-style-type: none"> 1- Free refurbish of one (1) existing unit of fumehood 1- Installation of canopy exhaust system for One (1) AAS Unit <p>- Location: First floor IP-chemical storage, second floor-organic and student research laboratory, third floor Biopharmaceutics and Biochemistry laboratory</p> <p>- External Dimension (LXWXH): 1200mm(47.0 in.) x 800mm (31.0 in.) x 2200mm (59.0 in.)</p> <p>- Internal Dimension (LXWXH): 1020mm(40 in.) x 760(26.0 in.) x 750mm (29 in.)</p> <p>- Maximum Sash Opening: 450mm (17.0 in.)</p> <p>- Exhaust volume/Static pressure: 720 CFM exhaust at 0.25" water max. external static pressure</p> <p>- Exhaust Outlet External Diameter: 300mm (11")</p> <p>- Main body: Cold-rolled steel with corrosion-proof coating</p> <p>- Solid chemical resistant physiochemical board</p> <p>- High grade melamine board with good acid and alkali resistance Function</p> <p>- Power Supply: 110/220V, 60Hz</p> <p>- Power Consumption: 21W (flourescent light only)</p> <p>- Water tap; Gas tap; water sin and base stand</p> <p>- Ducting requirements: From first floor to fourth floor, material: PVC</p> <p>- Conformance: CE & ISO</p> <p>- Electrical Requirement: 220V, 60Hz</p> <p>- Free refurbish of one (1) existing unit of fumehood</p>	183,753.75	735,015.00

Very truly yours,


JOSE FLORENCIO F. LAPEÑA JR., MD
 Vice-Chancellor for Administration

Conforme: 
JUANA CHARIS A. RICAMARA
 (Name of Representative of Bidder)
Golden Bat (Far East) Inc.
 (Name of Company)

Date: April 8, 2014

07 APR 2014