



PHILIPPINE GENERAL HOSPITAL
 THE NATIONAL UNIVERSITY HOSPITAL
 University of the Philippines Manila
 Health Sciences Center
 Taft Avenue, Manila

PHIC – Accredited Health Care Provider
 ISO 9001-2008 Certified

BAC ADDENDUM NO. 2018-91

**FOR THE SUPPLY AND DELIVERY OF CHEMICALS & REAGENTS WITH
 USE OF MACHINE (HEMATOLOGY) FOR 2019**

12 November 2018

Pursuant to Section 22.5.1 of the 2016 Revised Implementing Rules and Regulations of Republic Act No. 9184, the Bids and Awards Committee I is issuing this bid bulletin to modify or amend the following items in the Bid Documents:

I. The following items should be corrected as:

Item No.	From	To
I. MACHINE-REAGENT TIE-UP FOR COMPLETE BLOOD COUNT (CBC)		
1	Urine Test Strips, 150 strips/bottle	Reagents for Complete Blood Count (CBC)
II. MACHINE-REAGENT TIE-UP FOR COAGULATION TESTS		
2	Reagent for PT (Prothrombin Time) Quantity – 89,700	Reagent for PT (Prothrombin Time) Quantity – 89,700 test
3	Reagent for PTT (Partial Thromboplastin Time) Quantity – 80,200	Reagent for PTT (Partial Thromboplastin Time) Quantity – 80,200 test
4	Reagent for Factor VIII, At least 6-10 reagent bottles or sets/kit Quantity – 15	Reagent for Factor VIII, At least 6-10 reagent bottles or sets/kit Quantity – 15 kit
5	Reagent for Factor IX At least 2-6 reagent bottles or sets/kit Quantity – 15	Reagent for Factor IX At least 2-6 reagent bottles or sets/kit Quantity – 15 kit
6	Reagent for Protein C At least 2-4 reagent bottles or sets/kit Quantity – 12	Reagent for Protein C At least 2-4 reagent bottles or sets/kit Quantity – 12 kit
7	Reagent for Protein S At least 3-6 reagent bottles or sets/kit Quantity – 12	Reagent for Protein S At least 3-6 reagent bottles or sets/kit Quantity – 12 kit
8	Reagent for D Dimer At least 6-10 reagent bottles or sets/kit	Reagent for D Dimer At least 6-10 reagent bottles or sets/kit



PHILIPPINE GENERAL HOSPITAL
THE NATIONAL UNIVERSITY HOSPITAL
 University of the Philippines Manila
 Health Sciences Center
 Taft Avenue, Manila

PHIC – Accredited Health Care Provider
ISO 9001-2008 Certified

Item No.	From	To
	Quantity – 25	Quantity – 25 kit
9	Reagent for Lupus Anticoagulant Panel <ul style="list-style-type: none"> • DRVV Screen At least 10-12 reagent bottles or sets/kit Quantity – 3 • DRVV Confirm At least 10-12 reagent bottles or sets/kit Quantity – 3 • Silica Clotting Time At least 3-6 reagent bottles or sets/kit Quantity – 3 	Reagent for Lupus Anticoagulant Panel <ul style="list-style-type: none"> • DRVV Screen At least 10-12 reagent bottles or sets/kit Quantity – 3 kit • DRVV Confirm At least 10-12 reagent bottles or sets/kit Quantity – 3 • Silica Clotting Time At least 3-6 reagent bottles or sets/kit Quantity – 3

II. The following terms and conditions should be corrected as:

Item No.	From	To
	I. MACHINE-REAGENT TIE-UP FOR COMPLETE BLOOD COUNT (CBC)	
	Equipment	
5	Machine should be the latest model, and brand new.	<i>To be deleted</i>
8	Machine analyzer should be between 30-35 inches wide with built-in cabinets for reagents. It should have user-friendly input.	8. Machine analyzer should be not more than 36 inches wide with cabinets for reagents to fit in available space. It should have user-friendly input.
	Services and Maintenance for the Machines	
5	Definitions of severity level: Severity Level 1: acknowledge within 30 minutes from the time that the call was logged with the vendor and shall remedy defects and / or provide a workaround within 1 hour of notification of the problem, with a permanent solution within an agreed time frame.	Definitions of severity level: Severity Level 1: Complete loss of all services of the product and the situation is an emergency. The vendor will acknowledge within 30 minutes from the time that the call was logged with the vendor and shall remedy defects and / or provide a workaround within 1 hour of notification of the problem, with a



PHILIPPINE GENERAL HOSPITAL
 THE NATIONAL UNIVERSITY HOSPITAL
 University of the Philippines Manila
 Health Sciences Center
 Taft Avenue, Manila

PHIC – Accredited Health Care Provider
 ISO 9001-2008 Certified

Item No.	From	To
		permanent solution within an agreed time frame
II. MACHINE-REAGENT TIE-UP FOR COAGULATION TESTS		
	Equipment	
3	Machine should be the latest model, and brand new	<i>To be deleted</i>
	Services and Maintenance for the Machines	
5	Definitions of severity level: Severity Level 1: acknowledge within 30 minutes from the time that the call was logged with the vendor and shall remedy defects and / or provide a workaround within 1 hour of notification of the problem, with a permanent solution within an agreed time frame.	Definitions of severity level: Severity Level 1: Complete loss of all services of the product and the situation is an emergency. The vendor will acknowledge within 30 minutes from the time that the call was logged with the vendor and shall remedy defects and / or provide a workaround within 1 hour of notification of the problem, with a permanent solution within an agreed time frame

This shall form an integral part of the Bid Documents.

For guidance and information of all concerned.

Dean *M. M. Loquias*
MONET M. LOQUIAS, PhD
 Chairperson, Bids and Awards Committee I

M