



SUPPLEMENTAL / BID BULLETIN  
UNIVERSITY OF THE PHILIPPINES MANILA  
The Health Sciences Center  
Bids and Awards Committee 1  
Taft Avenue, Manila  
Trunk Line No. 8554-8400 Local 3014/3015



**BID BULLETIN NO. 2021-68**

**for the Supply and Delivery of Ten (10) Units Ambulatory Programmable Infusion Pump  
2 July 2021**

Pursuant to Section 22.5.1 of the 2016 Revised Implementing Rules and Regulations of Republic Act No. 9184, the Bids and Awards Committee 1 is issuing this bid bulletin to modify or amend the following items in the Bid Documents in response to / to address the request / clarification of the prospective bidder/s who attended the pre-bid conference held on 18 June 2021:

1. *The following Technical Specifications should be modified as:*

No.	From	To
5	Accuracy: +/-6% (nominal)	Accuracy: +/- <b>5 to</b> 6% (nominal)
27	Power Sources: AA (IEC LR6) alkaline batteries; AC adapter; rechargeable battery pack	Power Sources: AA (IEC LR6) alkaline <b>or lithium</b> batteries; AC adapter; rechargeable battery pack
34	Alarm Volume: High, medium, low 55db (+/-8dB) as measured at high volume	Alarm Volume: High, medium, low 55db (+/-8dB) as measured at high volume. <b><u>Units with &gt;=50dB for lowest level and &lt;=80dB for highest are also acceptable</u></b>

2. *Please be reminded of the following:*

- **The color of the folder to be used is RED;**
- **Dropping bids is not later than 9:00AM, 9 July 2021; and Opening of bids is on 9:30AM onwards, 9 July 2021.**

This shall form an integral part of the Bid Documents.

For the information and guidance of all concerned.

=Sgd.=

**Dean LEONARDO R. ESTACIO, Jr., PhD**  
Chairperson, Bids and Awards Committee 1

Received by the Bidder:

\_\_\_\_\_  
Signature over Printed Name

\_\_\_\_\_  
Name of Company

\_\_\_\_\_  
Date