



SUPPLEMENTAL / BID BULLETIN
UNIVERSITY OF THE PHILIPPINES MANILA
The Health Sciences Center
Bids and Awards Committee 1
Taft Avenue, Manila
Trunk Line No. 8554-8400 Local 3014/3015



BID BULLETIN NO. 2021-70

**for the Supply and Delivery of Medical Supplies for Cath Lab
(Framework Agreement)
23 July 2021**

Pursuant to Section 22.5.1 of the 2016 Revised Implementing Rules and Regulations of Republic Act No. 9184, the Bids and Awards Committee 1 is issuing this bid bulletin to modify or amend the following items in the Bid Documents in response to / to address the request / clarification of the prospective bidder/s who attended the pre-bid conference held on 16 July 2021:

1. *The following Technical Specifications should be modified as:*

Item No.	From	To
8	150mL fast turn syringe and quick fill tube (comptabile with existing Medrad Mark V Pro Vis Injector Syringe)	150mL fast turn syringe and quick fill tube (comptabile compatible with existing Medrad Mark V Pro Vis Injector Syringe)
9	150mL fast turn syringe and quick fill tube (comptabile with existing Medrad Mark 7 Injector Syringe)	150mL fast turn syringe and quick fill tube (comptabile compatible with existing Medrad Mark 7 Injector Syringe)
20	Cardiac Catheter Pack Fr 6 (Content: JL4, JR4 Fr, Pigtail Fr 5, Double Ended 0.035" Guidewire or Angled Hydrophilic Guidewire 0.035", 150cm & Introducer Sheath Fr 5)	Cardiac Catheter Pack Fr 6 Fr 5 (Content: JL4, JR4 Fr, Pigtail Fr 5, Double Ended 0.035" Guidewire or Angled Hydrophilic Guidewire 0.035", 150cm & Introducer Sheath Fr 5)
30	Angled Taper Fr 5 Glidecatheter, 100cm 100psi	Angled Taper Fr 5 Glidecatheter, 100cm 100psi 1000psi
36	Radial / Brachial Catheter 4.0 Fr 5 100-110cm	Radial / Brachial Catheter Fr 4.0 Fr 5.0 , 100-110cm
37	Y-adaptor kit with bleedback control valves (can be push click button or manual) inclusive of 6in kink resistant connecting tube	Y-adaptor kit with bleedback control valves (can be push click button or manual) inclusive of 6inches (15-20cm) kink resistant connecting tube
39	Cobra Middle Catheter Fr 5, 65cm, 1000psi	Cobra Middle Catheter Fr 5, 65cm-80cm , 1000psi
41	Simmons 1 Catheter Fr 5, 1000cm, 1000psi	Simmons 1 Catheter Fr 5, 1000cm 100cm , 1000psi
42	Simmons 2 Catheter Fr 5, 1000cm, 1000psi	Simmons 1 Catheter Fr 5, 1000cm 100cm , 1000psi
43	Straight Pigtail Catheter Fr 4, 110cm, 900psi	Straight Pigtail Catheter Fr 4, 110cm, 750psi-900psi
47	Yashiro Glidecatheter Fr 5, 70cm, 750psi	Yashiro Glidecatheter Fr 5, 70cm, 750psi-1000psi
51	0.025", J-Tip 150cm PTFE Coated Guidewire	0.025", J-Tip 145-150cm PTFE Coated Guidewire

Item No.	From	To
53	0.035", 150cm Angled Tip Hydrophilic Guidewire	0.035", 150cm Angled Tip or <u>J-Tip</u> Hydrophilic Guidewire
59	Introducer Sheath Fr 5, 7cm	Introducer Sheath Fr 5, <u>7cm-7.5cm</u>
62	Introducer Sheath Fr 5, 10cm with Introducer Needle	Introducer Sheath Fr 5, <u>10cm-11cm</u> with Introducer Needle
63	Introducer Sheath Fr 6, 10cm with Introducer Needle	Introducer Sheath Fr 6, <u>10cm-11cm</u> with Introducer Needle
64	Introducer Sheath Fr 7, 10cm with Introducer Needle	Introducer Sheath Fr 7, <u>10cm-11cm</u> with Introducer Needle
65	Introducer Sheath Fr 8, 10cm with Introducer Needle	Introducer Sheath Fr 8, <u>10cm-11cm</u> with Introducer Needle
72	Transradial Introducer Kit Fr 6 with puncture neede and hydrophilic guidewire	Transradial Introducer Kit Fr 6 with puncture <u>needle</u> and hydrophilic guidewire
73	Transradial Introducer Kit Fr 5 with puncture neede and hydrophilic guidewire	Transradial Introducer Kit Fr 5 with puncture <u>needle</u> and hydrophilic guidewire
83	Fractional Flow Reserve (FFR) Catheter	Fractional Flow Reserve (FFR) Catheter, <u>can be wired or wireless</u>
85	Thermodilution Catheter, Triple Lumen Fr 7 with thermistor 3.5cm from catheter tip	Thermodilution Catheter, <u>Triple Lumen or Four Lumen</u> Fr 7 with thermistor <u>3.5cm-4.0cm</u> from catheter tip
92	JL4 Fr 5 guiding Catheter, 100cm, 1000psi	JL4 Fr 5 guiding Catheter, 100cm, <u>700psi-1000psi</u>
93	JR4 Fr 5 guiding Catheter, 100cm	JR4 Fr 5 guiding Catheter, 100cm, <u>700psi-1000psi</u>
99	0.014"x180cm PTCA guidewire for tortuous vessels and highly stenosed lesions	0.014"x180cm PTCA guidewire for tortuous vessels and highly stenosed lesions, <u>tip load of 0.8gf</u>
100	0.014"x190cm PTCA Guidewire for chronic occlusion, pre shaped micro cone tip	0.014"x190cm PTCA Guidewire for chronic occlusion, pre shaped micro cone tip, <u>tip load of 3.5gf</u>
101	0.014"x180cm Floppy Straight PTCA guidewire, nitinol, irect-jointed 2 piece wire, hydrophilic coat	0.014"x <u>180cm-190cm</u> Floppy Straight PTCA guidewire, nitinol, <u>direct-jointed</u> 2 piece wire, hydrophilic coat
102	0.014"x180cm PTCA guidewire with coated polymer jacket	0.014" x <u>180cm- 190cm</u> PTCA guidewire with coated polymer jacket, <u>tip load of 0.8gf</u>
103	0.014"x180cm PTCA guidewire with composite core	0.014"x180cm PTCA guidewire with composite core, <u>tip load of 0.5gf</u>
104	0.014" x 18cm high torque balanced PTCA guidewire with moderate support	0.014" x 18cm <u>180cm-190cm</u> high torque balanced PTCA guidewire with moderate support
106	Coronary Microguide Catheter Fr 6, 150cm	Coronary Microguide Catheter Fr 6 <u>Fr 1.8 (tip size)</u> , 150cm
115	Biliary Drain Catheter Fr 10 x 40cm (Complete Set)	Biliary Drain / <u>Nephrostomy</u> Catheter Fr 10 x 40cm (Complete Set)
116	Biliary Drain Catheter Fr 12 x 40cm (Complete Set)	Biliary Drain / <u>Nephrostomy</u> Catheter Fr 12 x 40cm (Complete Set)
124	PGH Sterile Disposable Angiography Drapes (Regular Pack) - 2 pcs Image Intensifier (I.I.) or flat panel Sheet cover 24 x 58cm (dome size) - 1pc Image Intensifier (I.I.) or flat panel sheet cover 30 x 70cm (dome size) - 2pcs Shield Cover 100cm x 100cm	PGH Sterile Disposable Angiography Drapes (Regular Pack) - 2 pcs Image Intensifier (I.I.) or flat panel Sheet cover 24 x 58cm (dome size) - 1pc Image Intensifier (I.I.) or flat panel sheet cover 30 x 70cm (dome size) - 2pcs Shield Cover 100cm x 100cm

Item No.	From	To
	<ul style="list-style-type: none"> - 1pc Control Cover Plastic 140cm x 92 (flat form) - 1pc Monitor cover garterized 50cm x 50cm - 1pc Absorbable Angiography drape 4 holes - 1pc Back table cover zone reinforced 144cm x 92cm - 3pcs Surgical gown level 3 size: XL - 3pcs blue cloth towel 44cm x 66cm - 1pc Guidewire bowl wire tabs 2500mL - 2pcs small bowl 400mL - 1pc medicine cups 2oz 60cc - 1pc scalpel blade no 11 inches - 1pc mosquito forcep 5 inches 	<ul style="list-style-type: none"> - 1pc Control Cover Plastic 140cm x 92 (flat form) - 1pc Monitor cover garterized 50cm x 50cm - 1pc Absorbable Angiography drape 4 holes - 1pc Back table cover zone reinforced 144cm x 92cm 230cm - 3pcs Surgical gown level 3 size: XL - 3pcs blue cloth towel 44cm x 66cm - 1pc Guidewire bowl wire tabs 2500mL - 2pcs small bowl 400mL - 1pc medicine cups 2oz 60cc - 1pc scalpel blade no 11 inches - 1pc mosquito forcep 5 inches

2. *The following Terms and Conditions should be modified as:*

Item No.	From	To
10	Suppliers are required to submit sample/s in original packaging and provide at least 3 pcs for regular items and at least 1 pc for special (expensive) items. This will be subject for evaluation and as basis for inspection. Submission of sample/s shall be one (1) week before opening of bids - 25 June 2021.	Suppliers are required to submit sample/s in original packaging and provide at least 3 pcs for regular items and at least 1 pc for special (expensive) items. This will be subject for evaluation and as basis for inspection. Submission of sample/s shall be one (1) week before opening of bids - 25 June 2021 23 July 2021 (for new items).

3. *The following should be modified in the Bid Data Sheet:*

Item No.	From	To
14.1.a	The amount of not less than ____ [indicate the amount equivalent to two percent (2%) of ABC], if bid security is in cash, cashier's / manager's check, bank draft / guarantee or irrevocable letter of credit; or	The amount of not less than _____ [indicate the amount equivalent to two percent (2%) of ABC], if bid security is in cash, cashier's / manager's check, bank draft / guarantee or irrevocable letter of credit; or
14.1.b	The amount of not less than ____ [indicate the amount equivalent to five percent (5%) of ABC] if bid security is in Surety Bond.	The amount of not less than _____ [indicate the amount equivalent to five percent (5%) of ABC] if bid security is in Surety Bond.

4. *Please be reminded of the following:*

- The color of the folder to be used is BLUE;
- Dropping of bids is not later than 9:00AM, 30 July 2021; and
- Opening of bids is on 9:30AM onwards, 30 July 2021.

This shall form an integral part of the Bid Documents.

For the information and guidance of all concerned.

=Sgd.=

Dean LEONARDO R. ESTACIO, Jr., PhD
Chairperson, Bids and Awards Committee 1

Received by the Bidder:

Signature over Printed Name

Name of Company

Date