



SUPPLEMENTAL / BID BULLETIN
UNIVERSITY OF THE PHILIPPINES MANILA
The Health Sciences Center
Bids and Awards Committee 1
Taft Avenue, Manila
Trunk Line No. 8554-8400 Local 3014/3015



BID BULLETIN NO. 2022-96
26 August 2022

for the Supply and Delivery of Medical Supplies for
Anesthesiology
BAC1-2022-06-0041

Pursuant to Section 22.5.1 of the 2016 Revised Implementing Rules and Regulations of Republic Act No. 9184, the Bids and Awards Committee 1 is issuing this bid bulletin to modify or amend the following items in the Bid Documents in response to and address the request / clarification of the prospective bidder/s who attended the pre-bid conference held on 19 August 2022

1. *The following data should be modified in the Bid Data Sheet, Framework Agreement List and Technical Specifications as:*

Item No.	From	To
8	Airway, oral size 1, 70mm., sterile, oropharyngeal airway (guedel type)	Airway, oral size 1, 60mm -70mm, sterile, oropharyngeal airway (guedel type)
9	Airway, oral size 2, 80mm., sterile, oropharyngeal airway (guedel type)	Airway, oral size 2, 70mm -80mm, sterile, oropharyngeal airway (guedel type)
10	Airway, oral size 3, 90mm., sterile, oropharyngeal airway (guedel type)	Airway, oral size 3, 80mm -90mm, sterile, oropharyngeal airway (guedel type)
11	Airway, oral size 4, 100mm., sterile, oropharyngeal airway (guedel type)	Airway, oral size 4, 100mm- 110mm , sterile, oropharyngeal airway (guedel type)
12	Airway, oral, size 000** 30mm., sterile, oropharyngeal airway (guedel type)	Airway, oral, size 000** 30mm- 40mm , sterile, oropharyngeal airway (guedel type)
13	Airway, oral, size 00** 40mm., sterile, oropharyngeal airway (guedel type)	Airway, oral, size 00** 40mm- 50mm , sterile, oropharyngeal airway (guedel type)
14	Airway, oral, size 0** 50mm., sterile, oropharyngeal airway (guedel type)	Airway, oral, size 0** 50mm- 60mm , sterile, oropharyngeal airway (guedel type)

2. *Clarification on Terms and Conditions:*

Item No.	Query	Response
4.b	Valid Certificate of Product Registration (CPR) issued by the Food and Drug Administration (FDA). The name of the respective distributor should appear in the submitted CPR of the medical supplies. <i>Epoc Test Card or Blood Analysis Test Card not registered based on FDA Guidelines.</i>	TWG will have no objections provided that the submitted documents will clearly reflect this, and that they also include documents supporting their relationship with the principal.

Item No.	Query	Response
	<i>Supplier (3) has supporting documents but in principal's name.</i>	
4.e	<p>Certificate of Acceptance from at least three (3) major hospital issued within the year and should be supported with Sales Invoice (for new item/brand offered only).</p> <p><i>As existing supplier of Department of Anesthesia, is supplier (3) still required to submit this?</i></p>	<p>Yes, they are still required to submit this. All bidders are expected to comply with all the requirements as listed during each bidding process, regardless of any previous transactions which may have been completed.</p>
6	<p>Deliveries shall have at least one (1) year expiration date</p> <p><i>Blood analysis test cards usually have 3-4 months expiration. Supplier (3) is requesting to lower the adjust the said requirement accordingly.</i></p>	<p>If this is standard practice, the TWG would be agreeable to adjusting the expiration date expected for this item, provided that the winning bidder will be able to adjust the delivery process to ensure a continue supply to the hospital of the item.</p>

This shall form an integral part of the Bid Documents.

For the information and guidance of all concerned.

Original Signed

Dean CHARLOTTE M. CHIONG, MD, PhD
Chairperson, Bids and Awards Committee 1

Received by the Bidder:

Signature over Printed Name

Name of Company

Date