



**SUPPLEMENTAL / BID BULLETIN**  
**UNIVERSITY OF THE PHILIPPINES MANILA**  
*The Health Sciences Center*  
**Bids and Awards Committee 1**  
Taft Avenue, Manila  
Trunk Line No. 8554-8400 Local 3014/3015



**BID BULLETIN NO. 2022-97**  
**26 August 2022**

**for the Supply and Delivery of Medical Supplies for TCVS**  
**BAC1-2022-06-0044**

Pursuant to Section 22.5.1 of the 2016 Revised Implementing Rules and Regulations of Republic Act No. 9184, the Bids and Awards Committee 1 is issuing this bid bulletin to modify or amend the following items in the Bid Documents in response to and address the request / clarification of the prospective bidder/s who attended the pre-bid conference held on 19 August 2022

1. *The following data should be modified in the Bid Data Sheet, Framework Agreement List and Technical Specifications as:*

Item No.	From	To
<b>2</b>	<b>Chest drain</b> <ul style="list-style-type: none"> <li>- Model- Bedside</li> <li>- For Adult with 3-way Scheffler Valve (No Back flow)</li> <li>- 1.2-2300 liters container high capacity</li> <li>- can be used with or without water</li> <li>- Needle-less syringe port for sample specimen extraction:</li> <li>- built in hooks for bedside hanging</li> <li>- has drainage Tap:</li> <li>- Air Vent:</li> <li>- Airleak Indicator System for pneumothorax cases:</li> <li>- Can be attached to low suction pressure machine:</li> <li>- No clamping when draining of fluid</li> <li>- Packed in sterile packaging ready for use.</li> </ul>	<b>Chest drain</b> <ul style="list-style-type: none"> <li>- Model- Bedside</li> <li>- For Adult</li> <li>- with 3-way <del>Scheffler Valve</del> <b>valve</b> (No Back flow)</li> <li>- <del>1.2-2300 liters</del> <b>2300-2500 milliliters</b> container <del>high</del> capacity</li> <li>- can be used with or without water</li> <li>- Needle-less syringe port for sample specimen extraction:</li> <li>- built in hooks for bedside hanging</li> <li>- has drainage Tap:</li> <li>- Air Vent:</li> <li>- Airleak Indicator System for pneumothorax cases:</li> <li>- Can be attached to low suction pressure machine:</li> <li>- No clamping when draining of fluid</li> <li>- Packed in sterile packaging ready for use.</li> </ul>

2. *Clarification on Terms and Conditions:*

Item No.	Query	Response
<b>4.e</b>	<p>Certificate of Acceptance from at least three (3) major hospital issued within the year and should be supported with Sales Invoice (for new item/brand offered only).</p> <p><i>As existing supplier of Department of Anesthesia, is supplier (3) still required to submit this?</i></p>	<p>Yes, they are still required to submit this. All bidders are expected to comply with all the requirements as listed during each bidding process, regardless of any previous transactions which may have been completed.</p>

Item No.	Query	Response
<b>6</b>	<p>Deliveries shall have at least one (1) year expiration date</p> <p><i>Blood analysis test cards usually have 3-4 months expiration. Supplier (3) is requesting to lower the adjust the said requirement accordingly.</i></p>	<p>If this is standard practice, the TWG would be agreeable to adjusting the expiration date expected for this item, provided that the winning bidder will be able to adjust the delivery process to ensure a continue supply to the hospital of the item.</p>

This shall form an integral part of the Bid Documents.

For the information and guidance of all concerned.

*Original Signed*

**Dean CHARLOTTE M. CHIONG, MD, PhD**

*Chairperson, Bids and Awards Committee 1*

Received by the Bidder:

\_\_\_\_\_  
*Signature over Printed Name*

\_\_\_\_\_  
*Name of Company*

\_\_\_\_\_  
*Date*