



SUPPLEMENTAL / BID BULLETIN
UNIVERSITY OF THE PHILIPPINES MANILA
The Health Sciences Center
Bids and Awards Committee 1
Taft Avenue, Manila
Trunk Line No. 8554-8400 Local 3014/3015



BID BULLETIN NO. 2023-30
5 April 2023
for the Supply, Delivery, Installation, Testing and
Commissioning of One (1) Unit Robot Driven Gait Training
Machine and with Associated of Civil, Electrical and
Airconditioning Works
PUR23-03-0217

Pursuant to Section 22.5.1 of the 2016 Revised Implementing Rules and Regulations of Republic Act No. 9184, the Bids and Awards Committee 1 is issuing this bid bulletin to modify or amend the following items in the Bid Documents in response to and address the request / clarification of the prospective bidder/s who attended the pre-bid conference held on 31 March 2023:

1. *The following data should be modified in the Technical Specifications as:*

Item No.	From	To
B.1	Speed range 0 to 10km/hour	Speed range 0 to 10km/hour <u>(with Orthoses limited to 3.2km/hour)</u>
C.7	Maximum patient height: at least 200 cm (78.74 inches)	Maximum patient height: at least 200 cm (78.74 inches) <u>193cm (75.98 inches)</u>
E.2	Automatic Voltage Regulator appropriate for the equipment (Third Party)	Automatic Voltage Regulator <u>Uninterruptible Power Supply (UPS)</u> appropriate for the equipment (Third Party)
G.1	Temperature: 10°C to 30°C	Temperature: 10°C to 30°C <u>10 degrees Celsius to 35 degrees Celsius</u>
G.2	Humidity: 30% to 75% relative air humidity	Humidity: 30% <u>20%</u> to 75% relative air humidity
G.3	Air pressure: 700 hPa to 1060 hPa	Air pressure: 700 hPa <u>780 hPa</u> to 1060 hPa

This shall form an integral part of the Bid Documents.

For the information and guidance of all concerned.

=Original Signed=

Dean CHARLOTTE M. CHIONG, MD, PhD
Chairperson, Bids and Awards Committee 1

Received by the Bidder:

Signature over Printed Name

Name of Company

Date