



**SUPPLEMENTAL / BID BULLETIN**  
**UNIVERSITY OF THE PHILIPPINES MANILA**  
*The Health Sciences Center*  
**Bids and Awards Committee 1**  
Taft Avenue, Manila  
Trunk Line No. 8554-8400 Local 3014/3015



**BID BULLETIN NO. 2023-68**  
**06 October 2023**

**for the Supply, Delivery and Testing of One (1) Lot Endotracheal /  
Endoanal Ultrasound System**  
**PUR23-08-0771**

Pursuant to Section 22.5.1 of the 2016 Revised Implementing Rules and Regulations of Republic Act No. 9184, the Bids and Awards Committee 1 is issuing this bid bulletin to modify or amend the following items in the Bid Documents in response to and address the request / clarification of the prospective bidder/s who attended the pre-bid conference held on 29 September 2023:

1. *The following should be modified Section VII (Technical Specifications) as:*

Item No.	From	To
	<b>SPECIFICATIONS:</b>	
<b>A.1.b.ii</b>	With front and rear handles	With <del>front and rear handles</del> <b>small footprint</b>
<b>A.1.d</b>	With built-in battery/ standby battery	With built-in battery / standby battery <b>∟ battery pack</b>
<b>A.1.f</b>	Console with backlit alphanumeric keyboard, trackball and user-defined hard keys	Console with backlit alphanumeric keyboard, <b>or interactive keyboard light</b> , trackball and user-defined hard keys
<b>A.2.a</b>	Must have a high-resolution LED screen of not less than 19 (nineteen) inches in diagonal length	Must have a high-resolution LED <b>or LCD</b> screen of not less than nineteen (19) inches in diagonal length
<b>4.a.i</b>	Boot from shut-down less than 85 secs	<del>Boot from shut-down less than 85 secs</del>
<b>4.a.ii</b>	Boot from standby less than 20 secs	<del>Boot from standby less than 20 secs</del>
<b>4.a.iii</b>	Shut down less than 30 secs	<del>Shut down less than 30 secs</del>
<b>4.e.viii</b>	Panoramic imaging	Panoramic imaging <b>or 360 degrees image view capability when using the endoanal / endorectal transducer</b>
	<b>TERMS AND CONDITIONS</b>	
<b>C.8</b>	Certificate of Performance Evaluation from the Single Largest Contract	Certificate of Performance Evaluation from the Single Largest Contract <b>or Certificate of Acceptance from other government hospitals.</b>

Breakdown of the unit cost for the Components of the System:

Item	Unit Cost
A. Endorectal/ endoanal Ultrasound Machine	PHP5,364,048.00
A.2 Monitor	PHP144,000.00
B. Anorectal 3D Ultrasound Transducer – for Transrectal Scanning	PHP1,897,000.00
C. Colorectal Software Application Pack	PHP1,175,000.00
D. Accessories	
D.1 Must include compatible water standoff collar or equivalent with Luer lock valve and plastic extension tubing	PHP58,636.00
D.2 Must include reusable and sterilizable metal rectosigmoidoscope	PHP43,000.00
E. Consumables	
E.1 O-rings for use with compatible Water Standoff system (Packs of 10)	PHP6,430.00
E.2 Disposable transducer covers compatible with all three transducers (Packs of 24)	PHP17,535.00

This shall form an integral part of the Bid Documents.

For the information and guidance of all concerned.

*~Original Signed~*

**Dean CHARLOTTE M. CHIONG, MD, PhD**  
*Chairperson, Bids and Awards Committee 1*

Received by the Bidder:

\_\_\_\_\_  
*Signature over Printed Name*

\_\_\_\_\_  
*Name of Company*

\_\_\_\_\_  
*Date*