



SUPPLEMENTAL / BID BULLETIN
UNIVERSITY OF THE PHILIPPINES MANILA
The Health Sciences Center
Bids and Awards Committee 1
Taft Avenue, Manila
Trunk Line No. 8554-8400 Local 3014/3015



BID BULLETIN NO. 2023-89

**for the Supply, Delivery and Testing of One (1) Unit Portable
Operating Room Light
14 December 2023**

Pursuant to Section 22.5.1 of the 2016 Revised Implementing Rules and Regulations of Republic Act No. 9184, the Bids and Awards Committee 1 is issuing this bid bulletin to modify or amend the following items in the Bid Documents in response to / to address the request / clarification of the prospective bidder/s who attended the pre-bid conference held on 7 December 2023:

The following should be modified Technical Specifications as:

Item No.	From	To
A.7	Total irradiance at maximum: approx. 469 W/m ² max. without refocusing) L1/L2: 370 /630mm	Total irradiance at maximum: approx. 469 W/m ² max. without refocusing) L1/L2: 370 /630mm <u>or at least 3.31 W/m²/Klux</u>
A.12	Light head dimensions: 685 x 630mm	Light head dimensions: 685 x 630mm <u>or manufacturer's standard</u>
B.1.a	Homogenous light field	Homogenous light field <u>or a technology that prevent eye strain to the surgeons/end-user</u>
B.2	Must have backlite mode for illumination of the operating room during endoscopic interventions	Must have backlite mode <u>or comfort halo</u> for illumination of the operating room during endoscopic interventions

This shall form an integral part of the Bid Documents.

For the information and guidance of all concerned.

Dean CHARLOTTE M. CHIONG, MD, PhD
Chairperson, Bids and Awards Committee 1

Received by the Bidder:

Signature over Printed Name

Name of Company

Date