

SUPPLEMENTAL/ BID BULLETIN

UNIVERSITY OF THE PHILIPPINES
MANILA
The Health Sciences Center
BIDS AND AWARDS COMMITTEE 3

Bid Bulletin No. 1

06 May 2022

Project Title: Supply of Digital Whole Slide Image Scanner with Server and Slide Label Printer for the Creation of Digital Library of Histology and Histopathology.

(PR No. 16335)

Notice is hereby given to all interested supplier/prospective bidder of the following amendment/modification in the bidding documents for the above cited project:

Particular	From	To
Technical Specification	<ul style="list-style-type: none">- Objective (optical magnifications): 1.25x, 5x, 20x, 40x	<ul style="list-style-type: none">- Objective (optical magnifications): to include at least 5x, 20x, and 40x
Terms and Conditions Item No. 8	<ul style="list-style-type: none">- A Notarized certification from manufacturer that bidder is authorized distributor of equipment.	<ul style="list-style-type: none">- A notarized certification from the bidder upon submission of bids. The notarized/authenticated certification from the manufacturer should be submitted as postqualification documents.
Item No. 9	<ul style="list-style-type: none">- Notarized certification from manufacturer and local distributor that in the event of change of local distributor, preventive maintenance, warranty and services agreed hereupon will be honored by principal manufacturer.	<ul style="list-style-type: none">- Notarized certification from the bidder that in the event of change of local distributor, preventive maintenance, warranty and services agreed hereupon will be honored by principal manufacturer. Same notarized/authenticated (Apostle) certification from the manufacturer and local distributor should be submitted as postqualification documents.
Item No. 10	<ul style="list-style-type: none">- Guarantee letter or certificate from manufacturer and local distributor that availability of parts, supplies and accessories will not be discontinued for the next five (5) years.	<ul style="list-style-type: none">- Guarantee letter or certificate from the bidder that availability of parts, supplies and accessories will not be discontinued for the next five (5) years.

<p>Additional Item in the Terms and Condition</p> <p>Bid Data Sheet - ITB Clause 5.3.a.</p> <p>Delivery Period</p>	<p>- Supply of ICT (Information/Communications Technology) Software Systems, Hardware and Peripherals</p> <p>- Thirty (30) calendar days from receipt of approved purchase order and notice to proceed.</p>	<p>- That the supplier shall conduct an actual demo using a demo unit or an actual installation, whichever is applicable, during the post-qualification evaluation. Venue and date of demo should be arranged by supplier.</p> <p>- Supply of ICT (Information/Communications Technology) Software Systems, Hardware and Peripherals, or Microscope Systems, or Clinical Laboratory equipment for Anatomic and Clinicalpathology.</p> <p>- Thirty (30) to ninety (90) calendar days from receipt of approved purchase order and notice to proceed.</p>
---	---	--

CLARIFICATION OF THE STATEMENT UNDER TECHNICAL SPECIFICATIONS AS FOLLOWS:

“Sharing: Able to support 16 simultaneous live views for multiple regions from said slide:

It is required that the system to have an intuitive interface which allows users to display multiple, up to 16, simultaneous live views of their slide. This multiple simultaneous viewing capability is a very useful feature as it would allow users to see the same tissue area at multiple magnifications, or different regions of interest side-by-side.

2. All the other specifications, terms and conditions remain the same

3. This is posted at PhilGEPS www.philgeps.gov.ph

For guidance and information of all concerned.



TRISTAN NATHANIEL C. RAMOS, DDM, MPH
Chair ✓

Received by the bidder:

Company/Bidder:

Date: _____