

Proj. Ref. No.: **BAC1-2023-10-0076**
End-User: **DEPARTMENT OF PEDIATRICS**
Project: **SUPPLY, DELIVERY AND TESTING OF TWO (2)
PHASED ARRAY PROBE COMPATIBLE WITH EXISTING
2D ECHOCARDIOGRAPHY MACHINE**
Contract: **SINGLE BID**

Opening of Bids: **21 December 2023**
Total ABC: **Php1,005,097.00**

Item No.	Qty.	UOM	Item Description	Unit Cost	Quotations (all taxes included)	
					In Figures	In Words
			ARRAY PROBE COMPATIBLE WITH EXISTING 2D ECHOCARDIOGRAPHY MACHINE			
1	1	Pc	Phased Array Probe 1. Field of View: 90 Degrees 2. Dept of Field: 16cm 3. Frequency: 2.7 – 8.0Mhz 4. Compatibility: Must be compatible with 2D Echo Machine with Serial Number 60265932WXO	430,756.00		
2	1	Pc	Phased Array Probe 1. Field of View: 90 Degrees 2. Dept of Field: 30cm 3. Frequency: 1.3 – 4.0Mhz 4. Compatibility: Must be compatible with 2D Echo Machine with Serial Number 6026599WXO	574,341.00		
TOTAL APPROVED BUDGET FOR THE CONTRACT:				Php1,005,097.00		

TERMS AND CONDITIONS:

A. Requirement/s if awarded the contract:

1. Delivery Period: Within Ninety (90) calendar days after receipt of Notice to Proceed (NTP).
2. Delivery Place: Equipment Section, Property & Supply Division, Philippine General Hospital, Taft Avenue, Manila
3. Warranty Period / Coverage of Warranty: One (1) year on parts and two (2) years on service. Free quarterly preventive maintenance.
4. Acceptance Procedures and Parameters: Visual and Functional testing.
5. Certificate of Exclusive Distributorship.

Approved by:

Dean CHARLOTTE M. CHIONG, MD., PhD.
Chairperson

(Signature over Printed Name of President / Gen. Manager)

(Name & Address of Company)