

Proj. Ref. No.: **BAC1-2023-10-0079**  
End-User: **DEPARTMENT OF SURGERY / DIVISION OF SURGICAL ONCOLOGY**  
Project: **SUPPLY, DELIVERY, AND TESTING OF ONE (1)  
VACUUM-ASSISTED BREAST BIOPSY DEVICE**  
Contract: **SINGLE BID**

Opening of Bids: **01 December 2023**  
Total ABC: **Php4,500,000.00**

Item No.	Qty.	UOM	Item Description	Unit Cost	Quotations (all taxes included)	
					In Figures	In Words
1	1	Unit	<b>VACUUM-ASSISTED BREAST BIOPSY DEVICE</b>	4,500,000.00		
			<b>Components:</b> A. System Control Module B. Needle handpiece or needle holster C. System Mobile Cart D. Accessories and Consumables			
			<b>Specifications:</b> A. System Control Module 1. Must have a multicolor TFT LCD display with touchscreen user interface for adjustment and monitoring of biopsy system 2. Must contain software to run the vacuum assisted biopsy device a. Provides real-time, visual confirmation of needle orientation b. Able to configure for full or half or pre-programmed sampling modes c. clear and easily readable visual cues for real-time status of needle probe 3. Must be able to record and store patient information and procedure parameters and findings 4. Able to securely attach needle holster or handpiece			

Approved by:

**Dean CHARLOTTE M. CHIONG, MD., PhD.**  
Chairperson

\_\_\_\_\_  
(Signature over Printed Name of President / Gen. Manager)

\_\_\_\_\_  
(Name & Address of Company)

Proj. Ref. No.: **BAC1-2023-10-0079**  
End-User: **DEPARTMENT OF SURGERY/ DIVISION OF SURGICAL ONCOLOGY**  
Project: **SUPPLY, DELIVERY, AND TESTING OF ONE (1)  
VACUUM-ASSISTED BREAST BIOPSY DEVICE**  
Contract: **SINGLE BID**

Opening of Bids: **01 December 2023**  
Total ABC: **Php4,500,000.00**

Item No.	Qty.	UOM	Item Description	Unit Cost	Quotations (all taxes included)	
					In Figures	In Words
			5. Able to securely attach an optional remote foot pedal 6. System must contain vacuum pump with replaceable vacuum canister 7. Must contain built in power supply unit (220-240VAC; 50/60Hz) with power switch and exhaust fan for optimal working temperature of system module 8. Control module should be rated shock, dust and water resistant 9. Can be cleaned by isopropyl alcohol or sodium hypochlorite solution B. Needle handpiece or holster 1. Must have a single power and control cable 2. Sample aperture can be manually rotated to position probe and holster to desired location 3. Must have fingertip-operable buttons for vacuum and sample modes 4. Must be capable of single-hand operation 5. Can be cleaned by isopropyl alcohol or sodium hypochlorite solution C. System Mobile Cart 1. Must be of the same brand as system control module 2. Working load should support system control module,			

Approved by:

**Dean CHARLOTTE M. CHIONG, MD., PhD.**  
Chairperson

\_\_\_\_\_  
(Signature over Printed Name of President / Gen. Manager)

\_\_\_\_\_  
(Name & Address of Company)

Proj. Ref. No.: **BAC1-2023-10-0079**  
End-User: **DEPARTMENT OF SURGERY / DIVISION OF SURGICAL ONCOLOGY**  
Project: **SUPPLY, DELIVERY, AND TESTING OF ONE (1)  
VACUUM-ASSISTED BREAST BIOPSY DEVICE**  
Contract: **SINGLE BID**

Opening of Bids: **01 December 2023**  
Total ABC: **Php4,500,000.00**

Item No.	Qty.	UOM	Item Description	Unit Cost	Quotations (all taxes included)	
					In Figures	In Words
			keyboard, optional foot switch, ultrasound holster or handpiece) 3. Must have front and rear handle 4. Must have at least 4 caster wheels with locking tabs 5. Must be equipped with holster or handpiece docking station 6. Must have a top mount base for system control module 7. Must have mid shelf or storage bin 8. Must have central cable routing channel 9. Must have side cable wrap holders 10. Must be IV solution pole D. Accessories and Consumables 1. Twenty (20) 7G or 8G Ultrasound Probe 2. Five (5) 9G or 10G Ultrasound Probe 3. Five (5) at least 800cc replaceable vacuum canister			
<b>TOTAL APPROVED BUDGET FOR THE CONTRACT:</b>				<b>Ph4,500,000.00</b>		

#### TERMS AND CONDITIONS:

##### A. Requirement/s if declared as Lowest/Single Calculated Bids:

1. Presentation of Technical data sheet and/or presentation of a prototype equipment within seven (7) calendar days after receipt of Notice of Lowest / Single Calculated Bid.

Approved by:

**Dean CHARLOTTE M. CHIONG, MD., PhD.**  
Chairperson

\_\_\_\_\_  
(Signature over Printed Name of President / Gen. Manager)

\_\_\_\_\_  
(Name & Address of Company)

Proj. Ref. No.: **BAC1-2023-10-0079**  
End-User: **DEPARTMENT OF SURGERY / DIVISION OF SURGICAL ONCOLOGY**  
Project: **SUPPLY, DELIVERY, AND TESTING OF ONE (1)  
VACUUM-ASSISTED BREAST BIOPSY DEVICE**  
Contract: **SINGLE BID**

Opening of Bids: **01 December 2023**  
Total ABC: **Php4,500,000.00**

**B. Requirement/s if awarded the contract:**

1. Delivery Period: Within Ninety (90) calendar days after receipt of Notice to Proceed (NTP).
2. Delivery Place: Equipment Section, Property & Supply Division, Philippine General Hospital, Taft Avenue, Manila
3. Warranty Period / Coverage of Warranty: One (1) year on parts and Two (2) on service. Free biannual preventive maintenance during the warranty period. Warranty Period shall commence from the date of acceptance by the end user after installation, testing and commissioning.
4. Signed Service Level agreement with the Philippine General Hospital.
5. Original hard copy (not photocopy) or soft copy of operators and service manuals in English Language.
6. Product orientation for end users and troubleshooting training for end-users by at least two (2) biomedical engineers.
7. Upgradable system firmware and upgradable operating system during the period covered by the service warranty
8. Quotation of the Annual Preventive Maintenance Cost after the warranty period expires.
9. Acceptance Procedures and Parameters: Visual and functional testing.

**C. Documents Required of the bidder to be submitted during Post-Qualification:**

1. Brochures/Technical data Sheet.
2. SEC registration to prove that the supplier is in the business of importing and supplying medical equipment.
3. Certified true copy of the Certificate of Distributorship for the last five (5) years. The principal and the local distributor must have been in business partnership for at least five (5) years.
4. The model of the equipment must be the latest model available and distributed in Asia-Pacific region
5. Certificate that the Brand has been in the local market for at least five (5) years. Proof required: Invoices or Purchase Orders.
6. The brand must have been installed in at least Ten (10) government and/or private hospitals. A list of the hospital and contact number must be submitted.
7. Certification by the supplier that at least one service engineer is available locally to provide quick on-site support
8. Certification that components of the system are of the same brand except when otherwise stated as "third party" in the specification.
9. List of local Service Center/s
10. Certificate of Performance Evaluation from the Single Largest Contract.

Approved by:

---

**Dean CHARLOTTE M. CHIONG, MD., PhD.**  
Chairperson

---

(Signature over Printed Name of President / Gen. Manager)

---

(Name & Address of Company)

Proj. Ref. No.: **BAC1-2023-10-0079**  
End-User: **DEPARTMENT OF SURGERY/ DIVISION OF SURGICAL ONCOLOGY**  
Project: **SUPPLY, DELIVERY, AND TESTING OF ONE (1)  
VACUUM-ASSISTED BREAST BIOPSY DEVICE**  
Contract: **SINGLE BID**

Opening of Bids: **01 December 2023**  
Total ABC: **Php4,500,000.00**

11. License to Operate (LTO) from the Philippine FDA.

**D. Documents required of the principal to be submitted during Post-Qualification**

1. Certification that the manufacturer has been in the business of manufacturing hospital equipment for at least 10 years.
2. Guarantee letter from the manufacturer to ensure availability of supplies, parts and accessories for at least five (5) years after expiration of the warranty period.
3. Certification by the principal that service engineers are factory trained on service and repair.
4. ISO compliance certificate of the manufacturer.
5. List of the manufacturer's office and contact details in the following territories: Western Europe, US/Canada and Japan.

Approved by:

---

**Dean CHARLOTTE M. CHIONG, MD., PhD.**  
Chairperson

---

(Signature over Printed Name of President / Gen. Manager)

---

(Name & Address of Company)