

SUPPLEMENTAL/ BID BULLETIN
UNIVERSITY OF THE PHILIPPINES MANILA
The Health Sciences Center
BIDS AND AWARDS COMMITTEE 3

Bid Bulletin No. 1
19 March 2024

Project Title: Supply of Water Purification System (Type I and Type II) for Newborn Screening Center

(PR No. 24266)

Notice is hereby given to all interested supplier/prospective bidder of the following amendment/modification in the bidding documents for the above cited project:

Particular	From	To
Technical Specification <ul style="list-style-type: none">Includes	<ul style="list-style-type: none">- Reservoir 35 L with Sanitization Module & Reservoir	<ul style="list-style-type: none">- Reservoir at least 30 L with Sanitization Module & Reservoir
	<ul style="list-style-type: none">- Conductivity	<ul style="list-style-type: none">- Conductivity for Type 1 and Type 2
<ul style="list-style-type: none">1-year warranty with Consumable set good for 1 year of operation	<ul style="list-style-type: none">- 1. Flow rate: at least 5L/hour	<ul style="list-style-type: none">- 1. Flow rate: at least 5L/hour (Type 1)

2. All the other specifications, terms and conditions remain the same

3. This is posted at PhilGEPS www.philgeps.gov.ph

For guidance and information of all concerned.


f **Dir. IMELDA G. PEÑA, RPh, DrPH**
Chair, BAC3 

Received by the bidder: _____

Company/Bidder: _____

Date: _____