

Proj. Ref. No.: **PUR21-11-1070**
 End-User: **Dept. of Otolaryngology-Head and Neck Surgery**
 Project: **SUPPLY, DELIVERY, TESTING AND COMMISSIONING OF
 ONE (1) BRAND NEW FLEXIBLE PORTABLE FIBEROPTIC
 NASOPHARYNGOLARYNGOSCOPE**
 Contract: **Single Bid**

Opening of Bids: **December 14, 2021**
 ABC: **Php1,500,000.00**

Item No.	Qty.	UOM	Item Description	Unit Price	Quotations (all taxes included)	
					in figures	in words
1	1	Unit	BRAND NEW FLEXIBLE PORTABLE FIBEROPTIC NASOPHARYNGOLARYNGOSCOPE	1,500,000.00		
			<i>Nasopharyngolaryngoscope with the following features:</i>			
			1. Main Unit			
			a. It should have an optical system with a field of view of 85 degrees b. It should have a minimum range depth of field of 5 to 50 mm c. The minimum angulation range of the tip should be from 0 to 130 degrees up and 0 to 130 degrees down d. The total length of the unit should be at least 550 mm e. The outer diameter of the distal end should not be more than 3.1 mm f. Its working length (insertion section) should be between 300mm to 400mm g. It should have an eyepiece that allows user to visualize the desired field without the need to attach the unit to a video system h. It should be compatible to present existing imaging tower through a camera head			
			2. Rechargeable LED miniature light source that can be attached to the flexible fiber-optic scope			
			a. Dimension should not be larger than 70 mm in width, 90 mm in height and 35 mm in depth b. The weight of the rechargeable light source should not be more than 120 grams c. The light source can be used for 4 hours straight if fully charged.			
			3. Light guide cable			
			4. Light guide adaptor			

Approved by:
 ~Signed~

Dean LEONARDO R. ESTACIO, JR., PhD
 Chairperson

 (Signature over Printed Name of President / Gen. Manager)

 (Name & Address of Company)

Proj. Ref. No.: **PUR21-11-1070**
 End-User: **Dept. of Otolaryngology-Head and Neck Surgery**
 Project: **SUPPLY, DELIVERY, TESTING AND COMMISSIONING OF
 ONE (1) BRAND NEW FLEXIBLE PORTABLE FIBEROPTIC
 NASOPHARYNGOLARYNGOSCOPE**
 Contract: **Single Bid**

Opening of Bids: **December 14, 2021**
 ABC: **Php1,500,000.00**

			5. Compatible hard case with protective mold and handle		
Approved Budget for the Contract (ABC) =>				PHP1,500,000.00	

TERMS & CONDITIONS:

A. Requirement/s if declared as Lowest/Single Calculated Bids:

1. Presentation of brochure/catalogue and submission of prototype sample for evaluation of a duly constituted Technical Working Group within seven (7) calendar days after receipt of Notice of Lowest/ Single Calculated Bid.

B. Requirement/s if awarded the contract:

1. Delivery Period: Within One hundred twenty (120) calendar days after receipt of Notice to Proceed (NTP)
2. Delivery Place: Property Section, Philippine General Hospital, Taft Avenue, Manila
3. Notarized certificate that product is a) brand new including its parts and accessories, b) no reconditioned
4. Warranty Period / Coverage of Warranty:
 - a. One (1) year on parts and service for nasopharyngolaryngoscope
 - b. Notarized certification from the manufacturer or principal that it will honor any warranty obligation regardless of the change in distributor
5. Training: Undertaking to provide product orientation for end users and troubleshooting training for at least two (2) biomedical engineers of OETS for one (1) day.
6. Quotation of the Annual Preventive Maintenance Cost after the warranty period expires

C. Requirements to be submitted by the bidder for bid opening:

1. Brochures/Technical data Sheet
2. SEC registration to prove that the supplier is in the business of importing and supplying medical equipment for the past ten (10) years
3. Certification that the manufacturer has been in the business of manufacturing medical equipment and instruments for at least Ten (10) years.
4. Certified true copy of the Certificate of Distributorship for the last Five (5) years. The principal and the local distributor must have been in business partnership for at least Five (5) years.
5. The brand must have been in the local market for the past five (5) years. Proof required: Invoices or Purchase Orders.
6. Guarantee letter from the manufacturer and local distributor to ensure availability of supplies, parts and accessories for at least Five (5) years after expiration of the warranty period
7. Certification by the principal that service engineers are factory trained on service and repair.

Approved by:
 ~Signed~

Dean LEONARDO R. ESTACIO, JR., PhD
 Chairperson

 (Signature over Printed Name of President / Gen. Manager)

 (Name & Address of Company)

Proj. Ref. No.: **PUR21-11-1070**
End-User: **Dept. of Otolaryngology-Head and Neck Surgery**
Project: **SUPPLY, DELIVERY, TESTING AND COMMISSIONING OF
ONE (1) BRAND NEW FLEXIBLE PORTABLE FIBEROPTIC
NASOPHARYNGOLARYNGOSCOPE**
Contract: **Single Bid**

Opening of Bids: **December 14, 2021**
ABC: **Php1,500,000.00**

8. Certification from at least three (3) government and/or private hospital within Metro Manila or Philippines that they have supplied the same brand being offered
9. Certification by the supplier that at least one service engineer or technician is available locally to provide within twenty four (24) hours support.
10. Certification that components of the system are of the same brand except when otherwise stated as "third party" in the specification.
11. List of Service Center/s
12. Certificate of Performance Evaluation.
13. ISO/IEC compliance document of the manufacturer.
14. Proof that the manufacturer has presence in any country in Western Europe, US/Canada and Japan.

Approved by:
~Signed~

Dean LEONARDO R. ESTACIO, JR., PhD
Chairperson

(Signature over Printed Name of President / Gen. Manager)

(Name & Address of Company)