

Proj. Ref. No.: **PUR22-08-0689**

Opening of Rebid: **27 December 2022**
 Original Opening of Bids: **13 December 2022**

End-User: **DEPARTMENT OF PEDIATRICS /DIVISION OF NEONATOLOGY**
 Project: **SUPPLY, DELIVERY AND TESTING OF TWO (2) UNITS**
BILIRUBINOMETER (REBID)
 Contract: **SINGLE BID**

Total ABC: **Php960,000.00**

Item No.	Qty.	UOM	Item Description	Unit Cost (PHP)	Quotations (all taxes included)	
					in figures	in figures
1	2	Unit	BILIRUBINOMETER	480,000.00		
			Components: 1. Bilirubinometer with Reusable Probe 2. Docking station 3. Power Supply			
			Specifications:			
			A. General Features: 1. The unit offered should provide a visual digital measurement that has been shown to correlate with serum bilirubin in newborn infants as young as 24 weeks. 2. Measures bilirubin level using by calculating the difference in the optical densities for light in the blue (450nm) and green (550 nm) wavelength regions. 3. The machine should be compact and lightweight with weight of not more than 203 grams 4. With a backlit LCD user interface. 5. No calibration is needed for new patients. 6. With a barcode to enter nurse and patient ID.			

Approved by:

Dean CHARLOTTE M. CHIONG, MD., PhD.
 Chairperson

 (Signature over Printed Name of President / Gen. Manager)

 (Name & Address of Company)

Proj. Ref. No.: **PUR22-08-0689**

Opening of Rebid: **27 December 2022**
 Original Opening of Bids: **13 December 2022**

End-User: **DEPARTMENT OF PEDIATRICS /DIVISION OF NEONATOLOGY**
 Project: **SUPPLY, DELIVERY AND TESTING OF TWO (2) UNITS**
BILIRUBINOMETER (REBID)
 Contract: **SINGLE BID**

Total ABC: **Php960,000.00**

Item No.	Qty.	UOM	Item Description	Unit Cost (PHP)	Quotations (all taxes included)	
					in figures	in figures
			7. With a reusable measuring probe. 8. With capability of data storage and transfer via HL7. 9. With patient flagging capability to mark a patient report for special attention. 10. With internal memory that holds at least 100 data records. 11. With Touchscreen display 12. Can display results in mg/dL or $\mu\text{mol/L}$ 13. With an integrated light checker in the docking station. 14. Light source: Pulse xenon arc lamp 15. Light source life: At least 150,000 measurements 16. Sensors: Silicon photodiodes 17. Number of measurements when fully charged: at least 250 18. Performance Specifications: a. Measurement range: 0.0mg/dL to 20.0 mg/dL ($0\mu\text{mol/L}$ to $340\mu\text{mol/L}$) b. Clinical Data Standard Error of Estimate (SEE): ± 1.5 mg/dL or			

Approved by:

Dean CHARLOTTE M. CHIONG, MD., PhD.
 Chairperson

 (Signature over Printed Name of President / Gen. Manager)

 (Name & Address of Company)

Proj. Ref. No.: PUR22-08-0689

Opening of Rebid: 27 December 2022
Original Opening of Bids: 13 December 2022

End-User: DEPARTMENT OF PEDIATRICS /DIVISION OF NEONATOLOGY
Project: SUPPLY, DELIVERY AND TESTING OF TWO (2) UNITS
BILIRUBINOMETER (REBID)
Contract: SINGLE BID

Total ABC: Php960,000.00

Item No.	Qty.	UOM	Item Description	Unit Cost (PHP)	Quotations (all taxes included)	
					in figures	in figures
			±25.5 µmol/L (> 35 weeks gestation) 19. Power supply: a. Battery: Internal NiMH b. AC adapter: Input 9 VDC, 500 mA; Output 100 V - to 240 V -, 50/60 Hz, 11 VA to 18 VA 20. Ingress Protection: IPX0 or better			
TOTAL APPROVED BUDGET FOR THE CONTRACT:				Php960,000.00		

TERMS AND CONDITIONS:

A. Requirement/s if declared as Lowest/Single Calculated Bids:

1. Presentation of Technical data sheet and/or presentation of a prototype equipment within seven (7) calendar days after receipt of Notice of Lowest / Single Calculated Bid.

B. Requirement/s if awarded the contract:

1. Delivery Period: Within One Hundred Twenty (120) calendar days after receipt of Notice to Proceed (NTP).
2. Delivery Place: Equipment Section, Property & Supply Division, Philippine General Hospital, Taft Avenue, Manila
3. Warranty Period / Coverage of Warranty: One (1) year on parts and service. Free quarterly preventive maintenance during the warranty period. Warranty Period shall commence from the date of acceptance by the end user after installation, testing and commissioning.
4. Signed service level agreement with the Philippine General Hospital.
5. Original hard copy (not photocopy) or soft copy of operators and service manuals in English Language
6. Product orientation for end users and troubleshooting training for at least two (2) biomedical engineers

Approved by:

Dean CHARLOTTE M. CHIONG, MD., PhD.
Chairperson

(Signature over Printed Name of President / Gen. Manager)

(Name & Address of Company)

Proj. Ref. No.: **PUR22-08-0689**

Opening of Rebid: **27 December 2022**
Original Opening of Bids: **13 December 2022**

End-User: **DEPARTMENT OF PEDIATRICS /DIVISION OF NEONATOLOGY**
Project: **SUPPLY, DELIVERY AND TESTING OF TWO (2) UNITS**
BILIRUBINOMETER (REBID)
Contract: **SINGLE BID**

Total ABC: **Php960,000.00**

7. Quotation of the Annual Preventive Maintenance Cost after the warranty period expires.
8. Acceptance Procedures and Parameters: Visual and functional testing.

C. Documents Required of the bidder to be submitted during Post-Qualification:

1. Brochures/Technical data Sheet.
2. SEC registration to prove that the supplier is in the business of importing and supplying medical equipment.
3. Certified true copy of the Certificate of Distributorship for last three (3) years. The principal and the local distributor must have been in business partnership for at least five (5) years.
4. Certificate that the Brand has been in the local market for at least five (5) years. Proof required: Invoices or Purchase Orders.
5. The brand must have been installed in at least ten (10) government and/or private hospitals. A list of the hospital and contact number must be submitted.
6. Certification by the supplier that at least one service engineer is available locally to provide quick on-site support
7. List of local Service Center/s
8. Certificate of Performance Evaluation from the Single Largest Contract.
9. License to Operate (LTO) from the Philippine FDA.

D. Documents required of the principal to be submitted during Post-Qualification

1. Certification that the manufacturer has been in the business of manufacturing hospital equipment for at least 10 years.
2. Guarantee letter from the manufacturer to ensure availability of supplies, parts and accessories for at least five (5) years after expiration of the warranty period
3. Certification by the principal that service engineers are factory trained on service and repair.
4. IEC compliance certificate of the manufacturer.
5. List of the manufacturer's office and contact details in the following territories: Western Europe, US/Canada and Japan.

Approved by:

Dean CHARLOTTE M. CHIONG, MD., PhD.
Chairperson

(Signature over Printed Name of President / Gen. Manager)

(Name & Address of Company)