



Bid Notice Abstract

Request for Proposal (RFP)

Reference Number 9184015
Procuring Entity UNIVERSITY OF THE PHILIPPINES - MANILA PGH
Title SVP UR22-10 - 0917 Supply and Delivery of Medical Oxygen Flowmeter with Oxequip OETS
Area of Delivery Metro Manila

Solicitation Number:	No.:PUR22-10 - 0917	Status	Active
Trade Agreement:	Implementing Rules and Regulations		
Procurement Mode:	Negotiated Procurement - Small Value Procurement (Sec. 53.9)	Associated Components	1
Classification:	Goods	Bid Supplements	1
Category:	Medical Supplies and Laboratory Instrument		
Approved Budget for the Contract:	PHP 412,750.00	Document Request List	5
Delivery Period:	7 Day/s		
Client Agency:		Date Published	04/11/2022
Contact Person:	Cherryl Diaz Administrative Assistant II Taft Avenue Manila Metro Manila Philippines 1000 63-2-5548400 Ext.3022 63-2-5548420 crdiaz@up.edu.ph	Last Updated / Time	20/12/2022 12:19 PM
		Closing Date / Time	27/12/2022 15:00 PM

Description

1 MEDICAL OXYGEN FLOWMETER with Oxequip/Medstar 20 Piece 7,750.00
 Specifications:
 * Quick-connect adaptor
 * Brass body
 * Chrome plated
 * Working/Calibrated Pressure: 50PSI, Flowrate: 0-15LPM complete with reusable humidifier bottle
 * Compatible with the existing Oxequip/Medstar wall- type oxygen outlets of the hospital
 2 MEDICAL OXYGEN FLOWMETER with Chemetron/Ohmeda 6 Piece 6,250.00
 Specifications:
 * Quick-connect adaptor
 * Brass body
 * Chrome plated
 * Working/Calibrated Pressure: 50PSI, Flowrate: 0-15LPM complete with reusable humidifier bottle
 * Compatible with the existing Chemetron/Ohmeda wall- type oxygen outlets of the hospital
 3 MEDICAL OXYGEN REGULATOR 19 Piece 5,750.00
 Specifications:
 * Piston - type with reusable humidifier bottle
 * Flowrate: 0.15 LPM
 * with single -2" (approx.) PSI UL pressure gauge or equivalent
 * Atleast 3000 psi inlet pressure
 * Chrome plated
 * Brass body
 * Outlet pressure: Fixed at 50 psi.
 * Compatible with oxygen tank/Cylinder used by the hospital
 4 DUAL GAUGE MEDICAL OXYGEN REGULATOR for TRANSFER VENTILATOR/ANESTHESIA MACHINE 3 Piece 7,400.00

Specifications:

- * Diaphragm type
- * Brass body
- * Chrome plated
- * Inlet/Outlet Pressure: 0-4000/0-200 PSI
- * Outlet Pressure: Preset/calibrated to 50 PSI
- * Compatible with existing oxygen tank used by the hospital

5 DUAL GAUGE COMPRESSED AIR REGULATOR for TRANSFER VENTILATOR/ANESTHESIA MACHINE 12 Piece 7,400.00

Specifications:

- * Diaphragm type
- * Brass body
- * Chrome plated
- * Inlet/Outlet Pressure: 0-4000/0-200 PSI
- * Outlet Pressure: Preset/calibrated to 50 PSI
- * Compatible with existing oxygen tank used by the hospital

(Please see attch brochure or photo for reference)

- * PAGE TWO (2)

Approved Budget for the Contract 412,750.00

Remarks

re post

Created by Cheryl Diaz

Date Created 03/11/2022

The PhilGEPS team is not responsible for any typographical errors or misinformation presented in the system. PhilGEPS only displays information provided for by its clients, and any queries regarding the postings should be directed to the contact person/s of the concerned party.