

University of the Philippines Manila  
The Health Sciences Center  
SPECIAL BIDS & AWARDS COMMITTEE (SBAC)

Bid No. : PUR15-05-0664  
End-User : PHILHEALTH WARD 12, PGH  
Project : SUPPLY & DELIVERY OF HOSPITAL BED WITH FOAM MATTRESS

Opening of Bid: 27 July 2015

ITEM NO.	QTY	UNIT	ITEM DESCRIPTION	APPROVED PER UNIT (PhP)	QUOTATIONS (in figures / in words (all taxes included))
1	28	unit	<p align="center"><b>SUPPLY &amp; DELIVERY OF HOSPITAL BED WITH FOAM MATTRESS</b></p> <p align="center"><b>Brand new, not reconditioned/refurbished</b></p> <p align="center"><b>Including its Parts &amp; Accessories</b></p> <p><b>Specifications</b></p> <ol style="list-style-type: none"> <li>1. Manual caregiver control should be located at the foot end of the bed</li> <li>2. Bed frame should be corrosion resistant, powder coated and with double coating.</li> <li>3. With at least 85" - 90" standard length</li> <li>4. Sleep deck width: at least 35" - 37"</li> <li>5. Minimum width under bed clearance: 5" to 5.5"</li> <li>6. With zero-transfer gap for the siderails</li> <li>7. With collapsible aluminum locking siderails with minimum of four columns</li> <li>8. With complete bed articulation: <ol style="list-style-type: none"> <li>a. Head raise: 0 to 75 degrees</li> <li>b. Knee raise: 0 to 45 degrees</li> </ol> </li> <li>9. With articulation indicator</li> <li>10. With 4 position IV pole mounts (head and foot ends) and 1 piece stainless steel retractable IV pole with at least 4 collapsible hooks.</li> <li>11. Detachable, stable and heavy duty head and foot board with the following features: <ol style="list-style-type: none"> <li>a. Should be light weight and made with hard polymer material as substitute for metal</li> <li>b. Height of headboard and footboard: at least 42" to 44"</li> <li>c. Width of headboard and footboard: at least 35" to 37"</li> <li>d. With single step removal (push handle)</li> </ol> </li> <li>12. With four (4) caster brake and steer caster (mobile casters - 5" to 6")</li> <li>13. With Emergency CPR feature</li> <li>14. With Trendelenburg and Reverse Trendelenburg feature</li> <li>15. With three (3) crank mechanism</li> <li>16. With drainage bag holders on both side of the bed</li> <li>17. With provision for patient restraint straps</li> <li>18. Mattress: 76" (L) x 36" (W) with at least 5 - 6 inches thickness, original manufacturer foam</li> <li>19. With original manufacturer mattress cover made of soft nylon top cover, resistant to blood, urine and chemical stains.</li> <li>20. Sealed seal to prevent fluid leakage - welded wrap around zipper; allows for easy top cover</li> <li>21. Removal and replacements: with handle for ease in placing on and removing bed frame</li> <li>22. Can withstand at least 450lbs workload</li> </ol>	95,000.00	
<b>TOTAL APPROVED BUDGET FOR THE CONTRACT</b>				<b>2,660,000.00</b>	

Approved by:

**ORIGINAL SIGNED**  
Dean VICENTE O. MEDINA III, DDM, PhD  
Chairman */s/*

\_\_\_\_\_  
Signature over Printed Name of President/Gen. Manager)

\_\_\_\_\_  
Name & Address of Company)

Bid No. : PUR15-05-0664  
End-User : PHILHEALTH WARD 12, PGH  
Project : SUPPLY & DELIVERY OF HOSPITAL BED WITH FOAM MATTRESS

Opening of Bid: 27 July 2015

**TERMS & CONDITIONS:**

1. Indicate brand, model and country of origin.
2. Notarized certificate that the product is brand new, not reconditioned including parts and accessories.
3. Delivery should be done within ninety (90) calendar days commencing on the 3rd working day of notification through confirmed fax that the approved Purchase Order / Contract is already available for pick-up.
4. The bidder must be in the business of supplying medical equipment for the past ten (10) years.
5. The bidder must be in partnership with the principal over the past five (5) years.
6. Three (3) years warranty on service and at least two (2) years on parts.
7. Warranty period shall commence from the date of acceptance by the end-user after installation, testing and commissioning.
8. Automatic provision of functioning service unit (same or higher capacity) within three (3) days in case of repair / pull-out of the purchased unit during the warranty period.
9. The bed offered must have been demonstrated and tried by the end-user/s. Certification must be presented.
10. The bidder must have a notarized letter of undertaking that within five (5) years under normal use, the bed should be corrosion/rust free or else the winning bidder will provide a brand new unit/s with the same functionalities as a replacement.
11. Notarized certificates of good track records within five to ten (5-10) years from government and/or private hospitals within Metro Manila.
12. The winning bidder must provide original and hard copy of operators and service manuals upon delivery.
13. The bidder guarantees that the price offered is the lowest in the market.
14. **The bidder is required to submit single bid for the item. Two or more bids offer is automatically disqualified.**
15. **Compliance with the provisions of Republic Act No. 9184 and other applicable laws.**

Approved by:

**ORIGINAL SIGNED**  
Dean VICENTE O. MEDINA III, DDM, PhD  
Chairman *jr*

\_\_\_\_\_  
(Signature over Printed Name of President/Gen. Manager)

\_\_\_\_\_  
(Name & Address of Company)