

University of the Philippines
The Health Sciences Center
SPECIAL BIDS & AWARDS COMMITTEE (SBAC III)

Bid No. : PUR 15-06-0899
End User : DEPARTMENT OF PEDIATRICS
Project : SUPPLY, DELIVERY, TESTING AND COMMISSIONING OF BRAND
NEW INTENSIVE INFANT WARMERS

Opening of Bid: 23 September 2015

ITEM NO.	QTY	UNIT	ITEM DESCRIPTION	APPROVED PER UNIT (PhP)	QUOTATIONS (In figures/In words)(all taxes included)
1	15	unit	Supply, Delivery, Testing and Commissioning of Brand New Intensive Infant Warmers A. Specifications <ol style="list-style-type: none"> Warmer: Operating Temp. Range: 18°C to 30°C; Operating Humidity Range: 5% to 75%; Storage Temp. Range: -25°C to +60°C; Storage Humidity Range: 0% to 85% Resuscitation Unit: Operating Temp. Range: 18°C to 41°C; Operating Humidity Range: 0% to 90%; Storage Temp. Range: -25°C to +60°C; Storage Humidity Range: 0% to 95%; Storage Pressure Range: 50 kPa to 106 kPa Electrical Power Requirements: 110-210 volts AC; 50/60Hz models; Nominal power consumption; 520 watts at max. heater power and max lighting or better; Complies with IEC 60601-1 for electrical safety. System Control Characteristics: Full Color Display: 16.5 cm (6.5 in); Observation Lights: 2 dimmable halogen lights (35 W each), 0-1500 lux average; Procedure Light: 50 W, 2000 lux average; estimated life 3000 hrs based on manufacturer's rating; Warmup Mode: automatically started at power up; heater operates at 100% power for 7 minutes, then reverts to 50% power; check patient alarms eliminated during warmup mode; Microprocessor Based Control System: Self-test function performed at power on and continuously during operation; Heater Warmup Time at 100% Power: less than 3 minutes; Heater Element: 360 watts; Heater Output: adjustable from 0-100% in 5% increments; Patient Control (Baby) Mode: 34°C to 37.5°C in increments of 0.1°C (95°F to 99.5°F); Manual Mode: Indicates manual mode heat selection range from 0 to 100% in 5% increments; Temperature Measurement Accuracy: $\pm 0.3^{\circ}\text{C}$; Resolution: $\pm 0.1^{\circ}\text{C}$; Probe Accuracy: $\pm 0.1^{\circ}\text{C}$; Range: 30-42°C (71.6-107.6°F) 	375,000.00	
			TOTAL APPROVED BUDGET FOR THE CONTRACT	5,625,000.00	

Approved by:

(Signature over Printed Name of President/Gen. Manager)

Dean AGNES D. MEJIA, MD
Chairperson

(Name and Address of Company)

University of the Philippines
The Health Sciences Center
SPECIAL BIDS & AWARDS COMMITTEE (SBAC III)

Bid No. : PUR 15-06-0899

Opening of Bid: 23 September 2015

End User : DEPARTMENT OF PEDIATRICS

**Project : SUPPLY, DELIVERY, TESTING AND COMMISSIONING OF BRAND
NEW INTENSIVE INFANT WARMERS**

Terms and Condition:

1. Indicate brand, model and country of origin.
2. Notarized certificate that the product is brand new, not reconditioned/refurbished including parts and accessories.
3. Delivery and testing should be done within ninety (90) calendar days commencing on the 3rd working day of notification through confirmed fax that the approved Purchase Order/Contract is already available for pick-up.
4. Warranty with preventive maintenance should be five (5) years and five (5) years on parts.
5. Warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.
6. Provide regular quarterly preventive maintenance during the warranty service.
7. Provide a certification issued by the Pediatric Chairman after the item/unit has been demonstrated and/or site visited/inspected and evaluated by the Task force Head
8. The site visit/demo should be the exact machine model to be offered and the end-user's evaluation must be done prior to the opening of the bids.
9. A notarized certification from the manufacturer that the bidder is an authorized distributor of the equipment
10. Certification from the manufacturer and local distributor that in the event of a change of local distributor, preventive maintenance, warranty and services agreed here upon will be honored by the principal manufacturer.
11. Certification that the local distributor and the manufacturer have been in business relationship of providing the said ventilators for at least 5 years.
12. Notarized certification with proof that the company is in the business of supplying the unit/item for the last three (3) years.
13. Notarized certificate of SEC Registration that the company is in the business of supplying medical equipment for the last ten (10) years.
14. Guarantee letter or certificate from the manufacturer and local distributor to ensure the availability of supplies, spare parts, and accessories for the next five (5) years.
15. Provide separate quotation for maintenance, repair and replacement of parts after warranty.
16. Bidder must provide costing of all accessories that are consumed on a per patient basis
17. Notarized certificates of good track record for at least ten (10) years from government and/or other private hospitals with Metro Manila (or Philippines if applicable).
18. Must provide a soft copy (compatible with UPM-PGH Operating System) in CD and hard copy of the technical specifications.
19. Provide Operator's and Service's Manual upon delivery (in English language).
20. The winning bidder must provide training for end users (PGH nurses and PGH physicians) on site every 6 months for the duration of the warranty period. A notarized certificate must be provided.
21. The bidder must submit certification that they have the capability for corrective and preventive maintenance of the unit.
22. Must guarantee that diagnostic technician to be able to assess the unit within 24 hours of notification.
23. Automatic provision of a functioning service unit (same model) in case repair of more than 3 days/ pull out of the purchased unit during the warranty period.

(Signature over Printed Name of President/Gen. Manager)

Approved by:


Dean AGNES D. MEJIA, MD
Chairperson

(Name and Address of Company)

University of the Philippines
The Health Sciences Center
SPECIAL BIDS & AWARDS COMMITTEE (SBAC III)

Bid No. : PUR 15-06-0899

Opening of Bid: 23 September 2015

End User : DEPARTMENT OF PEDIATRICS

**Project : SUPPLY, DELIVERY, TESTING AND COMMISSIONING OF BRAND
NEW INTENSIVE INFANT WARMERS**

24. The winning bidder will be responsible of all the fees for the performance testing, delivery, installation, and commissioning of the units.
25. All certification required in the terms and condition shall be included/submitted in the envelope 1 (Technical Component).
26. It is understood that all taxes and other impositions are included in the bid offer. However, the bidder concerned shall be liable for additional or other taxes and duties which may be imposed by the government during the effectivity of the contract.
27. The bidder is required to submit single bid for the item. Two or more bids offer is automatically disqualified.
28. Compliance with Republic Act. No. 9184 and other applicable laws.

(Signature over Printed Name of President/Gen. Manager)

Approved by:


Dean AGNES D. MEJIA, MD
Chairperson

(Name and Address of Company)