

University of the Philippines
The Health Sciences Center
SPECIAL BIDS & AWARDS COMMITTEE (SBAC III)

Bid No. : PUR 15-06-0901

Opening of Bid: 14 October 2015


End User : DEPARTMENT OF PEDIATRICS

Project : SUPPLY, DELIVERY, TESTING AND COMMISSIONING OF BRAND NEW PULSE OXIMETER WITH BP APPARATUS

ITEM NO.	QTY	UNIT	ITEM DESCRIPTION	APPROVED PER UNIT (PhP)	QUOTATIONS (In figures/In words)(all taxes included)
1	20	unit	<p>Supply, Delivery, Testing and Commissioning of Brand New Pulse Oximeter with BP Apparatus</p> <p>A. Specifications</p> <ol style="list-style-type: none"> 1. NIBP: Technique: Oscillometric; Patient Range: Neonate to Children; Measurement Range: Systolic: 30-255 mmHg; Diastolic: 15-220 mmHg; MAP: 20-235 mmHg; Pulse Rate Range: 30-240 BPM; Accuracy: Blood Pressure: ± 5 mmHg standard deviation; no greater than 8 mmHg; Pulse Rate: $\pm 2\%$ or ± 2 BPM, whichever is greater 2. Provision of 25 Neonatal BP cuffs per unit, sizes:#1 (5 pieces), #2 (5 pieces), #3 (10 pieces), #4 (5 pieces) 3. Oximetry: SpO2 Range: 0-100%; SpO2 Accuracy: 70-100%, ± 3 digits (1 S.D.) Neonate; Pulse Rate Range: 25-240 BPM; Pulse Rate Accuracy: ± 3 BPM 4. Temperature: Temperature Range: 30° C to 43° C (85° F to 110° F); Accuracy: $\pm 0.1^\circ$ C ($\pm 0.2^\circ$ F); Meets or exceeds ASTM Standards 5. Display: LED display of measurement results, instructions, troubleshooting messages, and signal strength bar; Parameters Displayed: Systolic Pressure, Diastolic Pressure, Mean Arterial Pressure (MAP), Pulse Rate, %SpO2, Temperature ($^\circ$C or $^\circ$F) 6. Patient Alarms: Adjustable high and low alarms for Pulse Rate, SpO2, Systolic Pressure, Diastolic Pressure and Mean Arterial Pressure 7. Operating Modes: History: Review of previous measurements; NIBP: Manual, STAT or Automatic(at preset intervals); %SpO2: Continuous Monitoring; Temperature: Predictive or Direct 	68,000.00	

(Signature over Printed Name of President/Gen. Manager)

Approved by:


Dean AGNES B. MEJIA, MD
Chairperson



(Name and Address of Company)

University of the Philippines
The Health Sciences Center
SPECIAL BIDS & AWARDS COMMITTEE (SBAC III)

Bid No. : PUR 15-06-0901 **Opening of Bid: 14 October 2015**
End User : DEPARTMENT OF PEDIATRICS
Project : SUPPLY, DELIVERY, TESTING AND COMMISSIONING OF BRAND NEW PULSE OXIMETER WITH BP APPARATUS

ITEM NO.	QTY	UNIT	ITEM DESCRIPTION	APPROVED PER UNIT (PhP)	QUOTATIONS (In figures/In words)(all taxes included)
			<p>8. Power: Source: External line or internal battery; AC Power: 100-240 VAC, 50-60 Hz, 0.5A; T1.25A L250V fuse rating; DC Power (EMS Option): 12-30 VDC; Battery: NiMH battery pack (removable); Charge Time: 4 hours; Operating Time: Minimum 4 hours of continuous operation; Leakage Current: 100 μA (maximum); with one extra rechargeable battery</p> <p>9. Automatic Cuff Deflation: If cuff pressure exceeds 145 mmHg (Neonate); If measurement time exceeds 90 seconds (Neonate); If safety timer detects microprocessor failure</p> <p>10. Operating Environment: Operating Temperature: 0° C to 50° C (32° F to 122° F); Storage Temperature: -20° C to 65° C (-4° F to 149° F); Humidity: 15 - 95%, non-condensing; Altitude: Up to 3048 meters (10,000 feet); Monitors may not meet performance specifications if stored or used outside temperature and humidity ranges</p> <p>11. Accessories: RS-232 interface; Swiveled hard mount (for ambulance applications); Roll stand and basket; Protective carrying case.</p> <p>12. Must provide the following each unit/item:</p> <ul style="list-style-type: none"> a. Two (2) reusable pedia finger sensors b. Two (2) reusable multi sensors c. Ten (10) disposable sensor for pediatric d. Forty (40) disposable sensor for neonatal e. Thirty (30) disposable blood pressure cuffs for neonate f. Ten (10) disposable blood pressure cuffs for infant g. Ten (10) disposable blood pressure cuffs for child h. Must submit price list of all consumables 		
			TOTAL APPROVED BUDGET FOR THE CONTRACT	1,360,000.00	

 (Signature over Printed Name of President/Gen. Manager)

Approved by: 
 Dean AGNES D. MEJIA, MD
 Chairperson 

 (Name and Address of Company)

University of the Philippines
The Health Sciences Center
SPECIAL BIDS & AWARDS COMMITTEE (SBAC III)

Bid No. : PUR 15-06-0901 **Opening of Bid: 14 October 2015**
End User : DEPARTMENT OF PEDIATRICS
Project : SUPPLY, DELIVERY, TESTING AND COMMISSIONING OF BRAND NEW PULSE OXIMETER WITH BP APPARATUS

Terms and Condition:

1. Indicate brand, model and country of origin.
2. Notarized certificate that the product is brand new, not reconditioned/refurbished including parts and accessories.
3. Delivery and testing should be done within thirty (30) calendar days commencing on the 3rd working day of notification through confirmed fax that the approved Purchase Order/Contract is already available for pick-up.
4. Warranty period of 2 years parts and 2 years services.
5. Warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.
6. Provide regular quarterly preventive maintenance during the warranty service.
7. Provide a certification issued by the Pediatric Chairman after the item/unit has been demonstrated and/or site visited/inspected and evaluated by the Task force Head.
8. A notarized certification of authorized distributorship from the principal manufacturer mentioning the name of the Philippine representative.
9. Notarized certification with proof that the company is in the business of supplying the unit/item for the last three (3) years.
10. Notarized certificate of SEC Registration that the company is in the business of supplying medical equipment for the last ten (10) years.
11. Guarantee letter or certificate from the manufacturer and local distributor to ensure the availability of supplies, spare parts, and accessories for the next five years.
12. Provide separate quotation for maintenance, repair and replacement of parts after expiration of warranty.
13. Notarized certificates of good track record for at least ten (10) years from government and/or other private hospitals with Metro Manila (or Philippines if applicable).
14. Must provide a soft copy (compatible with UPM-PGH Operating System) in CD and hard copy of the technical specifications.
15. Provide Operator's and Service's Manual upon delivery (in English language).
16. The winning bidder must provide training for users and maintenance personnel from the PGH-OETS. A notarized certificate must be submitted.
17. The bidder must submit certification that they have the capability for corrective and preventive maintenance of the unit.
18. Availability of trained service engineers 24/7 with guaranteed uptime within 48 hours; must provide engineers training certificate.
19. Must guarantee that diagnostic technician to be able to assess the unit within 24 hours of notification.
20. All certification required in the terms and condition shall be included/submitted in the envelope 1 (Technical Component).
21. It is understood that all taxes and other impositions are included in the bid offer. However, the bidder concerned shall be liable for additional or other taxes and duties which may be imposed by the government during the effectivity of the contract.
22. The bidder is required to submit single bid for the item. Two or more bids offer is automatically disqualified.
23. Compliance with Republic Act. No. 9184 and other applicable laws.

(Signature over Printed Name of President/Gen. Manager)

Approved by:


Dean AGNES D. MEJIA, MD
Chairperson

(Name and Address of Company)