




University of the Philippines Manila  
 The Health Sciences Center  
**SPECIAL BIDS & AWARDS COMMITTEE (DOH FUND)**  
 BAC Office, PGH Compound  
 Taft Avenue, Manila  
 Tel. Nos.: 554-8400 loc. 3014 to 3015



**NOTICE OF AWARD**

9 September 2016

**WELLNESS PRO, INC.**  
 Unit 1006 Richmond Plaza  
 San Miguel Avenue corner Lourdes Street  
 Ortigas Center, Pasig City  
 Tel. No. 636-3580

Attention: **Mr. OCTOBER L. SUMULONG**  
 Area Health Liaison Specialist

Dear Sir:

We are happy to notify you that your bid dated 14 July 2016 but awarded on 5 September 2016 for the **SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING OF DIAGNOSTIC DISTORTION PRODUCT OTOACOUSTIC EMISSION & TRANSIENT EVOKED OTOACOUSTIC EMISSION MACHINE - SINGLE BID** under Project Reference No. PUR16-06-0691 for the Contract Price amounting to **FOUR HUNDRED NINETY FIVE THOUSAND PESOS & 00/100 (PHP495,000.00) ONLY** as corrected and modified in accordance with Instructions to Bidders is hereby accepted.

You are hereby to provide within **TEN (10) DAYS** the performance security in the form of: (a) Cash, Certified Check, Cashier's Check, Manager's Check, Bank Draft or Irrevocable Letter of Credit confirmed by a reputable local bank at 5% of the total amount of award which amounts to **TWENTY FOUR THOUSAND SEVEN HUNDRED FIFTY PESOS & 00/100 (PHP24,750.00) ONLY**; or (b) 30% of the total amount of award in the form of Security Bond callable upon demand issued by any reputable surety or insurance company in the amount of **ONE HUNDRED FORTY EIGHT THOUSAND FIVE HUNDRED PESOS & 00/100 (PHP148,500.00) ONLY** to be submitted to the BAC Office. Failure to provide the performance security shall constitute sufficient ground for the cancellation of the award and forfeiture of the bid security.

**EAR UNIT, DEPARTMENT OF OTORHINOLARYNGOLOGY, PGH**

Item No.	Qty.	Unit	Item Description	Unit Cost (PhP)	Total Cost (PhP)
1	1	unit	<b>DIAGNOSTIC DISTORTION PRODUCT            OTOACOUSTIC EMISSION &amp; TRANSIENT EVOKED            OTOACOUSTIC EMISSION MACHINE</b> <i>Brand new, not reconditioned            including its parts &amp; accessories</i> Brand / Model: <b>Eroscan Plus</b> Country of Origin: <b>Poland</b> Note: Please refer to the attached Bid Form (3 pages)	495,000.00	495,000.00
<b>TOTAL AMOUNT OF AWARD</b>				<b>495,000.00</b>	

*October L. Sumulong*  
 9-21-16

*S*



**BID FORM  
TECHNICAL SPECIFICATION**

**BID NO:** PUR16-06-0691  
**DATE:** JULY 14 2016  
**SUPPLY, DELIVERY, INSTALLATION TESTING AND COMMISSIONING OF DIAGNOSTIC EVOKED DISTORTION PRODUCT OTOACOUSTIC EMISSION (OAE) AND TRANSIENT EVOKED OTOACOUSTIC EMISSION MACHINE (OAE) - (SINGLE**

*Handwritten marks and signatures at the top left of the page.*

**WILL BE A TRUE COPY**  
**WILL BE A TRUE COPY**  
**WILL BE A TRUE COPY**

UNIT:	AGENCY SPECIFICATION	OFFERED SPECIFICATION	REFERENCES	EVALUATED BY:
1	Power supply: Lithium ion rechargeable	Power supply: Lithium ion rechargeable /	BROCHURE	Pls see attached
2	Battery life: Minimum of 300 test	Battery life: Minimum of 300 test	BROCHURE	Pls see attached
3	User Interference Display : LCD or OLED	OLED	BROCHURE	Pls see attached
4	Automatic date and time setting	Automatic date and time setting	BROCHURE	Pls see attached
5	Must be capable of ff measurements: a. Distortion Product Otoacoustic Emission ( DPOAE) b. Transient Otoacoustic Emission (TEOAE)	a. Distortion Product Otoacoustic Emission ( DPOAE) b. Transient Otoacoustic Emission (TEOAE)	BROCHURE	Pls see attached
6	Must be able to deliver ff frequency range a. DPOAE 1.5 to 12 KHz b. TPOAE 0.7 to 4 KHz	a. DPOAE 1.5 to 12 KHz b. TPOAE 0.7 to 4 KHz	BROCHURE	Pls see attached
7	Stimulus Intensity should be: a. DPOAE- DPOAE: 40-70 dB SPL b. TEOAE 83 dB SPL	a. DPOAE- DPOAE: 40-70 dB SPL b. TEOAE 83 dB SPL	BROCHURE	Pls see attached
8	Averaging time needed ; a. DPOE: 0.5-4 seconds b. TEOAE: 8-64 seconds	a. DPOE: 0.5-4 seconds b. TEOAE: 8-64 seconds	BROCHURE	Pls see attached
9	Must have passing criteria comparing to SNR ( Signal to Noise Ratio) a. DPOAE 6 dB (at least ) b. TPOAE 4 dB ( atleast)	a. DPOAE 6 dB (at least ) b. TPOAE 4 dB ( atleast)	BROCHURE	Pls see attached

*Handwritten signature and date: 9-21-16*

BY: OCTOBER 15, 2010  
 ARE HEALTHILIAS SPECIALIST  
 CERTIFIED TRUE COPY

10	Number of Passing Frequencies for Overall Test. "Pass": 3 frequencies out of 4	COMPLY	Number of Passing Frequencies for Overall Test. "Pass": 3 frequencies out of 4	BROCHURE	Pls see attached
<b>Must be delivered with ff. accessories</b>					
1	With handheld carrying bag/ case	comply	With handheld carrying bag	BROCHURE	Pls see attached
2	With Cradle Complete	COMPLY	With Cradle Complete	BROCHURE	Pls see attached
3	With Database Software for permanently storing test result, viewing result and then generating a professional report on standard paper	COMPLY	With Database Software for permanently storing test result, viewing result and then generating a professional report on standard paper	BROCHURE	Pls see attached
4	Alkaline or Built in battery	COMPLY	Alkaline or Built in battery	BROCHURE	Pls see attached
5	With Thermal Printer	COMPLY	With Thermal Printer	BROCHURE	Pls see attached
6	With UPS ( good for 1 hour)	COMPLY	With UPS ( good for 1 hour)	BROCHURE	Pls see attached
8	with extra extension probe cable (1) piece	COMPLY	with extra extension probe cable (1) piece	BROCHURE	Pls see attached
9	With Extra battery	COMPLY	With Extra battery	BROCHURE	Pls see attached
<b>With Complete computer system ( Notebook PC)</b>					
1	- Operating System : Windows 7 Professional 64-bit	COMPLY	- Operating System : Windows 7 Professional 64-bit	BROCHURE	Pls see attached
2	- Restore media: DRDVD Windows 7	COMPLY	- Restore media: DRDVD Windows 7	BROCHURE	Pls see attached
3	- Intel Core i353210 M Dc ( 2.5 Ghz, 3 MB L3 cache, 2 cores) ; Webcam intergrated HD 720 p	COMPLY	- Intel Core i353210 M Dc ( 2.5 Ghz, 3 MB L3 cache, 2 cores) ; Webcam intergrated HD 720 p	BROCHURE	Pls see attached
4	-Display: at least 14-inch High definition SV anti glare	COMPLY	-Display: at least 14-inch High definition SV anti glare	BROCHURE	Pls see attached
5	- Memory 4GB 1333 DDR3 (1x4GB)	COMPLY	- Memory 4GB 1333 DDR3 (1x4GB)	BROCHURE	Pls see attached
6	-Primary Hard Drive 3: 1 TB 7200 RPM HDD or if the internal HD is less tha 1 TB, will provide 1TB external HD	COMPLY	-Primary Hard Drive 3: 1 TB 7200 RPM HDD or if the internal HD is less tha 1 TB, will provide 1TB external HD	BROCHURE	Pls see attached
7	Fix Optical device: DVD RW Super Multi	COMPLY	Fix Optical device: DVD RW Super Multi	BROCHURE	Pls see attached

9-21-10  
 1-9-10

WELLNESS PRO INC.  
 CERTIFIED TECHNOLOGY SPECIALIST

8	Keyboard: 101/102; Finger print Reader with Security; 10/100/1000 Ethernet Controlller; Wlan A 802.11 bgn + Blietooth; HDMI Port; RJ 45 (1); USB 2.0 Battery ; Cell 47 Whr Lithium Ion Battery	COMPLY	Keyboard: 101/102; Finger print Reader with Security; 10/100/1000 Ethernet Controlller; Wlan A 802.11 bgn + Blietooth; HDMI Port; RJ 45 (1); USB 2.0 Battery ; Cell 47 Whr Lithium Ion Battery	BROCHURE
9	Power Adapter : External 90W Amart AC Adapter; Carry case	COMPLY	Power Adapter : External 90W Amart AC Adapter; Carry case	BROCHURE
<b>PRINTER</b>				
1	Multi - Function Laser color	COMPLY	Multi - Function Laser color	BROCHURE
2	Duty Cycle : 40000 pg/month	COMPLY	Duty Cycle : 40000 pg/month	BROCHURE
3	Speed Mono Print: 20 ppm	COMPLY	Speed Mono Print: 20 ppm	BROCHURE
4	Speed Color Print : 20 ppm	COMPLY	Speed Color Print : 20 ppm	BROCHURE
5	Speed Mono Copier : 20 ppm	COMPLY	Speed Mono Copier : 20 ppm	BROCHURE
6	Speed color Copier: 20 ppm	COMPLY	Speed color Copier: 20 ppm	BROCHURE
7	Resolution print ; 600x600 dpi	COMPLY	Resolution print ; 600x600 dpi	BROCHURE
8	Resolution copier; 300x420 dpi	COMPLY	Resolution copier; 300x420 dpi	BROCHURE
9	Resolution Scanner : 1200x1200 dpi	COMPLY	Resolution Scanner : 1200x1200 dpi	BROCHURE
10	Includes Installation : one year Warranty	COMPLY	Includes Installation : one year Warranty	BROCHURE

I Hereby certify that the statement of compliance to the foregoing technical spacification are true and correct, otherwise, if found if found to be false either during bid evaluation or post qualification, the same shall give rise to automatic disqualification of our bid.

JULY 7 2016  
Date

Signature Over Printed Name of Authorized Representative

*October E. Sumulong*  
October E. Sumulong

Wellness Pro Inc.  
Name of Company / Bidder

Name of Company / Bidder

*October E. Sumulong*  
9-21-16