



SUPPLEMENTAL / BID BULLETIN
UNIVERSITY OF THE PHILIPPINES MANILA
The Health Sciences Center
Special Bids and Awards Committee 2
Taft Avenue, Manila
Tel. #554-84-00 loc. 3026/3020; 526-4359



3 December 2015

ADDENDUM NO. SBAC-VOM2015-57

This Addendum No. **SBAC-VOM2015-57** is issued to modify or amend the following in connection with the **Supply, Delivery, Testing and Commissioning of Wrist Arthroscopy System - Single Bid** with PR/Bid No. **PUR15-09-1435** intended for the Department of Orthopedics which is scheduled on **10 December 2015**. This shall form an integral part of the Bid Documents.

1. *The following should be deleted on the Specifications as:*

- > **1.c. 2.7mm outer diameter**
- > **2.d. Diameter 2.9mm**

2. *The following should be corrected on the Specifications as:*

From	To
1.a.i. One (1) short scope (1 piece)	1.a.i. One (1) short scope, <u>2.4mm - 2.7mm diameter</u> (1 piece)
1.a.ii. One (1) long scope (1 piece)	1.a.i. One (1) long scope, <u>2.7mm diameter</u> (1 piece)
2.a. High flow cannula for wrist scope	2.a. <u>Low to</u> high flow cannula for wrist scope
2.a.i. 2.9mm for 2.7mm short scope (1 piece)	2.a.i. <u>2.9mm - 3.5mm</u> for <u>2.4 - 2.7mm</u> short scope (1 piece)
2.a.ii. 2.9mm for 2.7mm long scope (1 piece)	2.a.ii. <u>2.9mm - 4.0mm</u> for 2.7mm long scope (1 piece)
2.b. Dual port	2.b. <u>Single</u> or dual port
2.f. Must be supplied with blunt obturator, conical tip (2 pieces)	2.f. Must be supplied with <u>compatible</u> blunt obturator, conical tip
3.d. 2.5x optical zoom	3.d. 2.5x optical zoom <u>or equivalent zoom</u>
5.a. 150W LED or Xenon light source	5.a. 150W LED or Xenon light source <u>or better</u>
9.c. Compatible with sagittal saw attachment with sterile disposable blades (1 piece)	9.c. Compatible sagittal saw attachment with sterile disposable blades (1 piece) <u>with or without depth stop</u>
9.c.i. Sagittal saw blade, 5.8mm x 18mm x 0.4mm with straight fine teeth, etched marking and depth stop (3 boxes)	9.c.i. Sagittal saw blade, <u>5.8mm - 6.0mm x 18mm - 25mm x 0.4mm</u> with straight fine teeth, (3 boxes)

From	To
9.c.ii. Sagittal saw blade, 5.8mm x 12mm x 0.4mm with straight fine teeth, no markings and depth stop (3 boxes)	9.c.ii. Sagittal saw blade, <u>5.8mm – 8.0mm x 12mm – 25mm x 0.4mm</u> with straight fine teeth, (3 boxes)
9.c.iii. Sagittal saw blade, 9.4mm x 25.5mm x 0.4mm with straight fine teeth, with etched markings, no depth stop (3 boxes)	9.c.iii. Sagittal saw blade, <u>9.4mm – 10.0mm x 25.0mm – 25.5mm x 0.4mm</u> with straight fine teeth, with etched markings, no depth stop (3 boxes)
10.a.i. Tip grasper 2mm x 90mm (1 piece)	10.a.i. <u>Micrograsper 2.0mm – 2.75mm x 85mm – 120mm</u> (1 piece)
10.a.ii. Grasper 2.75mm x 90mm (1 piece)	10.a.ii. <u>Micrograsper 2.75mm – 3.0mm x 85mm – 120.0mm</u> (1 piece)
10.a.iii. Mosquito tip grasper 2mm x 90mm (1 piece)	10.a.iii. <u>Suction punch straight, tip profile 2.5mm or with sheath diameter 3.0 – 4.8mm x 85.0 – 130mm</u> (1 piece)
10.a.iv. Mosquito tip grasper 2.75mm x 90mm (1 piece)	10.a.iv. <u>Micropunch straight cutting width 1.0 – 2.0mm x 85 – 120mm</u> (1 piece)
10.a.v. 45-degree right tip biter 2mm x 90mm (1 piece)	10.a.v. <u>Micropunch curved right cutting width 2.0 – 2.7mm x 85.0 – 120mm</u> (1 piece)
10.a.vi. 45-degree left tip biter 2mm x 90mm (1 piece)	10.a.vi. <u>Micropunch curved left cutting width 2.0 – 2.7mm x 85.0 – 120mm</u> (1 piece)
10.a.vii. Blunt probe, straight 2mm x 90mm (1 piece)	10.a.vii. <u>Blunt probe 90 degrees graduated 2.0 – 3.0mm hook 85.0 – 120mm (2 pieces)</u>
10.a.viii. Oval punch straight 2mm x 90mm (1 piece)	10.a.viii. <u>Oval punch biter 2.0 – 2.5mm x 85.0 – 120mm</u> (1 piece)
10.b.i. 2.0mm cutter x 70mm (5 pieces)	10.b.i. <u>2.0-3.5mm cutter x 70 – 85mm</u> (5 pieces)
10.b.ii. 2.0mm full radius x 70mm (5 pieces)	10.b.ii. <u>2.0-3.5mm full radius x 70 – 85mm</u> (5 pieces)
10.b.iii. 3.5mm full radius x 70mm (5 pieces)	10.b.iii. <u>3.5mm full radius x 70 – 85mm</u> (5 pieces)
11.a.iv. At least 835 x 1580 x 750mm dimensions	11.a.iv. At least 835 x 1580 x 750mm dimensions <u>or compatible with equipment</u>
11.a.viii. With four (4) hook IV stand with adjustable height.	11.a.viii. With <u>two to four (2-4) hooks</u> IV stand with adjustable height.

3. The following should be added on the Specifications as:

- 10.a.x. Microfracture pick/chisel 30-40 degrees x 85-120mm (1 piece)
- 10.a.xi. Microfracture pick/chisel 65-70 degrees x 85-120mm (1 piece)

4. The following should be corrected on the Terms and Conditions as:

From	To
7. Delivery should be done within sixty (60) calendar days commencing on the 3 rd working day of notification through	7. Delivery should be done within <u>90 – 120 calendar days</u> commencing on the 3 rd working day of notification through

confirmed fax that the approved Purchase Order / Contract is already available for pick-up.	confirmed fax that the approved Purchase Order / Contract is already available for pick-up.
10. Automatic provision of a functioning service unit during the warranty period (same model) in case of repair of the purchased unit within seventy two (72) hours of notification.	10. Automatic provision of a functioning service unit during the warranty period (same model <i>or compatible</i>) in case of repair of the purchased unit within seventy two (72) hours of notification.
20. Certificate of the unit that it conforms to the International Electrotechnical Commission (IEC) Standards or its equivalent.	20. Certificate <i>that manufacturer</i> conforms to the International Electrotechnical Commission (IEC) Standards / <i>ISO Standards</i> or its equivalent.
22. Include technical specifications (brochure / catalogue).	22. <i>Must provide soft copy in CD (in Excel and Word format) and hard copy (brochure / catalogue) of the technical specifications.</i>

5. Signatories for Envelope 1 and 2 should be the same person – President or General Manager.

6. Eligible bidders are required to submit the Bid Security either in the form of:

a. Cash/Cashier's/Manager's Check payable to the University of the Philippines Manila/Philippine General Hospital at two percent (2%) of the total Approved Budget of the Contract (ABC). Said amount should be paid at the UPM-PGH Cash Services Division and a photocopy of the Official Receipt should be attached to Envelope) 1.

b. A combination of Cash/Cashier's/Manager's Check at a minimum of P50,000.00 and the balance at 2% Bank draft/guarantee or Irrevocable Letter of Credit issued by Universal or Commercial Bank; Provided, however, that it shall be confirmed or authenticated by a Universal or Commercial Bank, if issued by a foreign bank. The Php50,000.00 should be paid at the UPM-PGH Cash Services Division and a photocopy of the Official Receipt should be attached to Envelope 1.

NOTE: Amount of Bid Security (Equal to percentage of the ABC) – First Php50,000.00 and the rest 2% bank draft/guarantee or irrevocable letter of credit.

c. A combination of Cash/Cashier's/Manager's Check at a minimum of Php50,000.00 and the balance at 5% Surety Bond callable upon demand, it shall be accompanied by a certification by the Insurance Commission that the surety or insurance company is authorized to issue such instruments. The Php50,000.00 should be paid at the UPM-PGH Cash Services Division and a photocopy of the Official Receipt and original copy of the Surety Bond should be attached to Envelope 1.


NOTE: Amount of Bid Security (Equal to Percentage of the ABC) – First Php50,000.00 and the rest 5% Surety Bond.

d. Bid Securing Declaration

7. Bid Bulletin, Financial bid form – which includes bid prices, bill of quantities and applicable price schedules, and any other document required in the BIDS will form part of Envelope No. 2 and will be submitted together with Envelope No. 1, **on or before 9:00AM, 10 December 2015 (Thursday)** at the Office of the Internal Audit, UP Manila, 8th Floor, Central Block Building, PGH Complex, Taft Avenue, Manila.

8. Please refer to the Instruction to Bidders for other information.

For the information and guidance of all concerned.

For: Leonor L. Cabral-Rivera
Dean VICENTE O. MEDINA III, DDM, PhD
Chairperson, Special Bids and Awards Committee 2 

Received by the Bidder:

Signature over Printed Name

Name of Company

Date