



University of the Philippines Manila
The Health Sciences Center
SPECIAL BIDS & AWARDS COMMITTEE 2
 BAC Office, PGH Compound
 Taft Avenue, Manila
 Tel. Nos.: 554-8400 loc. 3014 to 3016



NOTICE OF AWARD

28 December 2015

MEDTRONIX MEDICAL SUPPLIES & EQUIPMENT

72-A Scout Limbaga Street
 Diliman, Quezon City
 Tel. No. 415-2528

Attention: **Mr. FELIX M. DULAY, JR.**
Sales / Service Manager

Dear Sir:

We are happy to notify you that your bid dated 7 December 2015 but awarded on 17 December 2015 for the SUPPLY, DELIVERY, TESTING & COMMISSIONING OF DESK-BASED BLOOD ANALYSIS SYSTEM (SINGLE BID) under Bid No. PUR15-09-1395 for the Contract Price amounting to SEVEN HUNDRED THIRTY EIGHT THOUSAND PESOS & 00/100 (PHP738,000.00) ONLY as corrected and modified in accordance with Instructions to Bidders is hereby accepted.

You are hereby to provide within TEN (10) DAYS the performance security in the form of: (a) Cash, Certified Check, Cashier's Check, Manager's Check, Bank Draft or Irrevocable Letter of Credit confirmed by a reputable local bank at 5% of the total amount of award which amounts to THIRTY SIX THOUSAND NINE HUNDRED PESOS & 00/100 (PHP36,900.00) ONLY; or (b) 30% of the total amount of award in the form of Security Bond callable upon demand issued by any reputable surety or insurance company in the amount of TWO HUNDRED TWENTY ONE THOUSAND FOUR HUNDRED PESOS & 00/100 (PHP221,400.00) ONLY to be submitted to the BAC Office. Failure to provide the performance security shall constitute sufficient ground for the cancellation of the award and forfeiture of the bid security.

DEPARTMENT OF SURGERY

Item No.	Qty.	Unit	Item Description	Unit Cost (PhP)	Total Cost (PhP)
1	1	unit	DESK-BASED BLOOD ANALYSIS SYSTEM Brand: ITC Model: IRMA TRUPOINT Country of Origin: U.S.A. Brand new, not reconditioned including its parts & accessories Note: Please refer to the attached document for the complete specifications	738,000.00	738,000.00
TOTAL AMOUNT OF AWARD				738,000.00	


Bids and Awards Committee I
 CROSS CHECKED FROM ORIGINAL
 BY: Evelyn DATE: 1/20/16

[Signature]
CARLOS F. CALANCA JR.
 1-15-2016


TERMS AND CONDITIONS

1. Delivery should be done within one hundred twenty (120) calendar days commencing on the 3rd working day of notification through confirmed fax that the approved Purchase Order / Contract is already available for pick-up.
2. The system must be delivered ready to use. Cables, connectors, adapters, and accessories not specified in the specifications but are essential for the unit to function must be provided free.
3. The machine must be the latest model.
4. Provide product orientation and operations training for end-user and troubleshooting training for Biomedical Engineers.
5. Two (2) years warranty on service and on parts.
6. Warranty period shall commence from the date of acceptance by the end-user after installation, testing and commissioning.
7. Free quarterly preventive maintenance during the warranty period.
8. Service engineers are certified factory trained on service and repair. One (1) service engineer must be available locally to provide quick on-site support.
9. Automatic provision of functioning service unit (same or higher capacity) within three (3) days in case of repair/pull-out of the purchased unit during the warranty period.
10. All components must be of the same brand from the same manufacturer unless otherwise specified as "Third Party."
11. Must guarantee that diagnostic technician will be able to assess the unit within 24 hours upon notification.
12. Provide Operator's & Service's Manual (in English language) upon delivery.
13. Compliance with Republic Act No. 9184 and other applicable laws.


Very truly yours,


JOSE C. GONZALES, M.D. DEC 29 2015
Director, Philippine General Hospital

CONFORME:


CARLOS F. CHANKIAN JR.
(Signature over Printed Name)
SERVICE ENGINEER.
(Designation)
JAN 15, 2016
(Date)

Side and Awards Committee
CROSS CHECKED FROM ORIGINAL
BY: Quik DATE: 1/20/16


CARLOS F. CHANKIAN JR.
1-15-2016

BID FORM
(Technical Specification)

RY, TESTING & COMMISSIONING OF A BRAND NEW BLOOD ANALYSIS SYSTEM - (SINGLE-BID)

TECHNICAL SPECIFICATIONS	Offered Specifications	Reference (include Supporting Documents)	Comments/Remarks (clarity, inclusion/exclusion)	EVALUATED BY: NAME OF END-USER/REPRESENTATIVE
BLOOD ANALYSIS SYSTEM not reconditioned parts and Accessories	Brand new not reconditioned including its parts and Accessories			
desk based	Point of care, desk based	Brochure/Manual		
edge	Single-use cartridge	Brochure/Manual		
interactive touch screen user in thermal printer	Interactive touch screen user interface with built- in thermal printer	Brochure/Manual	Compliant	
ure cartridge storage, no need preferred. Shelf life of three (3) to four (4) months at	Room temperature cartridge storage (15 - 30 degree Celsius), no need for refrigeration. Shelf life of cartridges between three (3) to four (4) months at room temperature.	Brochure/Manual	Compliant	
efore running test	Automatically Self calibrates before running test	Brochure/Manual		
atic electronic quality control	On-board automatic (EQC) electronic quality control test	Brochure/Manual		
ard calibrant. No external required.	Prepacked onboard calibrant. No external reagents or gases required.	Brochure/Manual		
ntroduction system	Closed sample introduction system (luer lock)	Brochure/Manual		
2ml of blood sample from a	Minimum of 0.2ml of blood sample from a syringe.	Brochure/Manual		
le in 2 minutes or less.	Results available in 2 minutes or less.	Brochure/Manual		
e sodium, potassium, Calcium 2, pO2, Urea, Glucose, tocrit, TCO2, HCO3, and SO2	Can analyse the sodium, potassium, Calcium Chloride, pH, pCO2, pO2, Urea, Glucose, Hemoglobin, Hematocrit, TCO2, HCO3, and SO2.	Brochure/Manual		

ment of compliance to the foregoing technical specifications are true and correct, otherwise, if found to be false either
qualification, the same shall give rise to automatic disqualification of our bid.

PLIES & EQUIPMENT _____ 1-Dec-15 Date
Bidder _____
Signature Over Printed Name of Authorized Representative

Felix M. Dulay Jr.
Signature Over Printed Name of Authorized Representative

1800 18th Avenue Committee 1
CROSS CHECKED FROM ORIGINAL
BY: *Edith* DATE: *1/20/16*

Carlos F. Calambanan Jr.
1-15-2016