

Bid No. : PUR 15-07-1105

Opening of Bid: 11 September 2015

End User : CCU - MAT

**Project : SUPPLY, TESTING, AND COMMISSIONING OF NETWORKABLE
MULTI-PARAMETER MECHANICAL VENTILATOR FOR ADULT AND
PEDIATRIC PATIENTS FOR THE UP-PGH CCU-MAT**

ITEM NO.	QTY	UNIT	ITEM DESCRIPTION	APPROVED PER UNIT (PhP)	QUOTATIONS (In figures/In words)(all taxes included)
1	30	unit	<p>Supply, Delivery, Testing and Commissioning of Networkable Multi-Parameter Mechanical Ventilators for Adult and Pediatric Patients for the UP-PGH CCU-MAT</p> <p>A. Specifications</p> <ol style="list-style-type: none"> 1. Brand new mechanical ventilators for adult and pediatric patients. 2. Capable of handling patients at least 5 kgs. Ventilators capable for handling patients below 5 kgs are acceptable. 3. Must have the following invasive ventilation modes: <ol style="list-style-type: none"> a. Controlled Ventilation <ul style="list-style-type: none"> - Volume Controlled Ventilation (VCV) - Pressure Controlled Ventilation (PCV) b. Supported Ventilation <ul style="list-style-type: none"> - Pressure Support Ventilation or Continuous Positive Airway Pressure (CPAP) c. Combined Ventilation <ul style="list-style-type: none"> - Synchronized intermittent mandatory Ventilation (VC) + PS - Synchronized intermittent mandatory Ventilation (PC) + PS d. Newer Modes of Ventilation <ul style="list-style-type: none"> - Airway Pressure Release Ventilation - Proportional Assist Ventilation or Adaptive Support Ventilation 4. Must have Non-invasive ventilation 5. Must be able to alter flow patterns into a square or decelerating pattern 6. Breathing circuitry must be a dual limb ventilation 7. Range of respiratory rate at least between 1 to 100 breaths per minute. Rates above 100 pbm may be included. 	2,300,000.00	

(Signature over Printed Name of President/Gen. Manager)

Approved by:


Dean AGNES D. MEJIA, MD
Chairperson *18 6*

(Name and Address of Company)

Bid No. : PUR 15-07-1105

Opening of Bid: 11 September 2015


End User : CCU - MAT

Project : SUPPLY, TESTING, AND COMMISSIONING OF NETWORKABLE
 MULTI-PARAMETER MECHANICAL VENTILATOR FOR ADULT AND
 PEDIATRIC PATIENTS FOR THE UP-PGH CCU-MAT

ITEM NO.	QTY	UNIT	ITEM DESCRIPTION	APPROVED PER UNIT (PhP)	QUOTATIONS (In figures/In words)(all taxes included)
			8. Range of Tidal Volume between at least 20 ml to 2000 ml 9. Must be able to provide both pressure and flow triggering and sensitivity with the following ranges: a. Flow triggering - 0.2 to 20 LPM b. Pressure triggering - 0.5 to 20 cmH ₂ O 10. Must be able to provide the following range of pressures a. PEEP/CPAP/EPAP: 0 - 35 cmH ₂ O (minimum range) b. IPAP/ Pressure Control/ Pressure Support: 0 - 40 cm H ₂ O (minimum range) 11. FiO ₂ delivered must be between 21 - 100% 12. Must be able to alter any of the following flow parameters a. Inspiratory time with ranges from 0.1 to 12 second b. Inspiratory flow rates between 6 to 150 LPM c. IE ratio between 1:00 to 3:1 13. Must be able to generate waveforms that include all of the following: airway pressure, flow, volume. 14. Must be able to provide loops for monitoring: PV, V-Flow, P-flow loops 15. Provide continuous capnography or ETCO ₂ measurements 16. Must be able to provide the following numeric monitoring variables: respiratory frequency, inhaled and exhaled tidal volume; minute ventilation; PEEP and auto-PEEP; Mean airway, peak airway and Plateau airway Pressures; RSBI (Rapid Shallow Breathing Index), dynamic and static compliance		

 (Signature over Printed Name of President/Gen. Manager)

Approved by:


 Dean AGNES B. MEJIA, MD
 Chairperson

 (Name and Address of Company)

Bid No. : PUR 15-07-1105

Opening of Bid: 11 September 2015

End User : CCU - MAT

**Project : SUPPLY, TESTING, AND COMMISSIONING OF NETWORKABLE
MULTI-PARAMETER MECHANICAL VENTILATOR FOR ADULT AND
PEDIATRIC PATIENTS FOR THE UP-PGH CCU-MAT**

ITEM NO.	QTY	UNIT	ITEM DESCRIPTION	APPROVED PER UNIT (PhP)	QUOTATIONS (In figures/In words)(all taxes included)
			<p>17. Able to perform the following maneuvers for monitoring:</p> <ul style="list-style-type: none"> a. Inspiratory pause or Inspiratory Hold Maneuver b. Expiratory pause or Expiratory Hold Maneuver or an assessment of negative inspiratory force or Negative inspiratory pressure or a P0.1 max <p>18. Monitoring and trending of events recorded up to 1000 events at least.</p> <p>19. Graphic trends recorded for at least 36 hours.</p> <p>20. Must have leak compensation and tube compensation capability.</p> <p>21. Must have apnea back up ventilation.</p> <p>22. Must have flow sensing.</p> <p>23. Built in nebulizer or nebulizing system.</p> <p>24. Must be able to adjust volume of alarm settings.</p> <p>25. Internal back up battery of at least 3 hours.</p> <p>26. Built in or back up Compressor.</p> <p>27. Able to provide ventilation support during suctioning.</p> <p>28. Must be able to use universal breathing circuitry or tubing.</p> <p>29. Provision of humidifier and chamber</p> <p>30. Display must be LED of at least 9 inches</p> <p>31. Must be capable of displaying at least 3 waveforms at a time.</p> <p>32. Must provide consumables of at least 3 per ventilator per patient type.</p> <p>33. Must be usable on a 2-prong convenience output.</p> <p>34. Must be able to provide operating hour reading of Mechanical Ventilator usage.</p>		

(Signature over Printed Name of President/Gen. Manager)

Approved by:


Dean AGNES D. MEJIA, MD
Chairperson *K 6*

(Name and Address of Company)

Bid No. : PUR 15-07-1105

Opening of Bid: 11 September 2015

End User : CCU - MAT

**Project : SUPPLY, TESTING, AND COMMISSIONING OF NETWORKABLE
MULTI-PARAMETER MECHANICAL VENTILATOR FOR ADULT AND
PEDIATRIC PATIENTS FOR THE UP-PGH CCU-MAT**

ITEM NO.	QTY	UNIT	ITEM DESCRIPTION	APPROVED PER UNIT (Php)	QUOTATIONS (In figures/In words)(all taxes included)
			<p>35. And all equipment must be turn key. Bidder must provide the following consumables (whenever applicable) of at least 3 per ventilator per patient type:</p> <ul style="list-style-type: none"> a. humidifier base w/chamber b. flow sensor oxygen cell inspiratory bacteria filter c. HEPA filter expiratory filter d. exhalation valve assembly e. expiratory flow sensor CO2 sampling line f. test bag 2.0L g. test hose h. test lung i. oxygen hose assembly j. air hose assembly k. other needed accessories for each operation 		
			TOTAL APPROVED BUDGET FOR THE CONTRACT	69,000,000.00	

(Signature over Printed Name of President/Gen. Manager)

Approved by:


Dean AGNES D. MEJIA, MD
Chairperson 

(Name and Address of Company)

Bid No. : PUR 15-07-1105

Opening of Bid: 11 September 2015

End User : CCU - MAT

Project : **SUPPLY, TESTING, AND COMMISSIONING OF NETWORKABLE
MULTI-PARAMETER MECHANICAL VENTILATOR FOR ADULT AND
PEDIATRIC PATIENTS FOR THE UP-PGH CCU-MAT**

Terms and Condition:

1. Indicate brand, model and country of origin.
2. Notarized certificate that the product is brand new and not reconditioned/refurbished including parts and accessories.
3. Delivery and installation should be done within ninety (90) calendar days commencing on the 3rd working day of notification through confirmed fax that the approved Purchase Order / Contract is already available for pick up.
4. Warranty with preventive maintenance should be five (5) years and five (5) years on parts.
5. Warranty period shall commence from the date of acceptance by the end-user.
6. Provide regular free quarterly preventive maintenance and calibration during the warranty service.
7. Provide a certification issued by CCU - MAT Head after the item/unit has been demonstrated and/or site visited/inspected and evaluated by specific end-user heads.
8. The site visit/demo should be exact machine model to be offered and the end-user's evaluation must be done prior to the opening of the bids.
9. A notarized certification from the manufacturer that the bidder is an authorized distributor of the equipment.
10. Certification from the manufacturer and local distributor that in the event of a change of local distributor, preventive maintenance, warranty and services agreed here upon will be honored by the principal manufacturer.
11. Certification that the local distributor and the manufacturer have been in business relationship of providing the said ventilators for at least 5 years.
12. Notarized certification accompanied with proof that the principal manufacturer is in the business of supplying the unit/item for more than ten (10) years.
13. Guarantee letter or certificate from manufacturer and local distributor to ensure the availability of parts, supplies and accessories and will not be discontinued for the next 10 years.
14. Provide separate quotation for maintenance, repair and replacement of parts after warranty.
15. Bidder must provide costing of all accessories that are consumed on a per patient basis.
16. Notarized certification of good track records at least ten (10) years from government and/or private hospitals within Metro Manila or the Philippines.
17. Must provide a soft copy (compatible with UPM-PGH Operating system) in CD and Hard copy of the technical specifications.
18. Provide operator's & service's manual upon delivery in the English Language.
19. The winning bidder must provide training for end users (PGH nurses and PGH physicians) on site every 6 months for the duration of the warranty period. A notarized certificate must be provided.
20. The bidder must submit certification that they have a capability for corrective and preventive maintenance of the unit.
21. Availability of trained service engineers 24/7 with guaranteed uptime within 48 hours; must provide engineers training certificate.
22. Must guarantee that diagnostic technician to be able to assess the unit within 24 hours of notification.

(Signature over Printed Name of President/Gen. Manager)

Approved by:


Dean AGNES D. MEJIA, MD
Chairperson 

(Name and Address of Company)

University of the Philippines
The Health Sciences Center
SPECIAL BIDS & AWARDS COMMITTEE (SBAC III)

Bid No. : PUR 15-07-1105

Opening of Bid: 11 September 2015

End User : CCU - MAT

**Project : SUPPLY, TESTING, AND COMMISSIONING OF NETWORKABLE
MULTI-PARAMETER MECHANICAL VENTILATOR FOR ADULT AND
PEDIATRIC PATIENTS FOR THE UP-PGH CCU-MAT**

23. Automatic provision of a functioning service unit (same model) in case repair of more than 3 days/ pull out of the purchased unit during the warranty period.
24. The winning bidder will be responsible of all the fees for the performance testing, delivery, installation, and commissioning of the units.
25. Installation and provision of fifteen (15) free adaptors that is compatible with three service areas.
26. All certification required in the terms and condition shall be included/submitted in the envelope 1 (Technical Component).
27. It is understood that all taxes and other impositions are included in the bid offer. However, the bidder concerned shall be liable for additional or other taxes and duties which may be imposed by the government during the effectivity of the contract.
28. The bidder is required to submit single bid for the item. Two or more bids offer is automatically disqualified.
29. Compliance with Republic Act. No. 9184 and other applicable laws.

(Signature over Printed Name of President/Gen. Manager)

Approved by:


Dean AGNES D. MEJIA, MD
Chairperson

(Name and Address of Company)